



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q2552

Continuing Calib Date: 08/08/2025

Initial Calibration Date(s): 08/08/2025 08/08/2025

Continuing Calib Time: 14:32

Initial Calibration Time(s): 10:48 11:41

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DBCP	5.96	5.96	5.86	6.06	0.00
EDB	2.61	2.61	2.51	2.71	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q2552

Continuing Calib Date: 08/08/2025

Initial Calibration Date(s): 08/08/202508/08/2025

Continuing Calib Time: 14:32

Initial Calibration Time(s): 10:4811:41

GC Column: ZB-MR2ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DBCP	4.99	4.99	4.89	5.09	0.00
EDB	1.97	1.97	1.87	2.07	0.00