

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2552
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/08/2025 08/08/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/08/2025
Lab Sample No.: M8011.504.1 0.1 PF **Data File :** PQ070752.D **Time Analyzed:** 14:32

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
DBCP	5.956	5.857	6.057	0.100	0.100	0.0
EDB	2.607	2.506	2.706	0.110	0.100	10.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2552
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/08/2025 08/08/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/08/2025
Lab Sample No.: M8011.504.1 0.1 PF **Data File :** PQ070752.D **Time Analyzed:** 14:32

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
DBCP	4.987	4.887	5.087	0.110	0.100	10.0
EDB	1.974	1.874	2.074	0.110	0.100	10.0