



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2552</u>	Analytical Method:	<u>8011</u>
Client:	<u>Alliance Technical Group, LLC - Newark</u>	Datafile :	<u>PQ070748.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB169164BS (Column 1)	DBCP	0.25	0.22	ug/L	88				70	130	
	EDB	0.25	0.22	ug/L	88				70	130	
PB169164BS (Column 2)	DBCP	0.25	0.25	ug/L	100				70	130	
	EDB	0.25	0.23	ug/L	92				70	130	