

Hit Summary Sheet
SW-846

SDG No.: Q2552
Client: Alliance Technical Group, LLC - Newark

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	WS0725-PT-RVOA-WS							
Q2552-07	WS0725-PT-RVOA Water		Vinyl Chloride	22.8	E	0.12	0.50	ug/L
Q2552-07	WS0725-PT-RVOA Water		1,1-Dichloroethene	8.60		0.13	0.50	ug/L
Q2552-07	WS0725-PT-RVOA Water		Methylene Chloride	13.3		0.44	0.50	ug/L
Q2552-07	WS0725-PT-RVOA Water		trans-1,2-Dichloroethene	18.5	E	0.14	0.50	ug/L
Q2552-07	WS0725-PT-RVOA Water		Carbon Tetrachloride	2.30		0.13	0.50	ug/L
Q2552-07	WS0725-PT-RVOA Water		cis-1,2-Dichloroethene	5.10		0.13	0.50	ug/L
Q2552-07	WS0725-PT-RVOA Water		1,1,1-Trichloroethane	17.9	E	0.12	0.50	ug/L
Q2552-07	WS0725-PT-RVOA Water		Benzene	15.0	E	0.11	0.50	ug/L
Q2552-07	WS0725-PT-RVOA Water		1,2-Dichloroethane	13.2		0.14	0.50	ug/L
Q2552-07	WS0725-PT-RVOA Water		Trichloroethene	11.9		0.13	0.50	ug/L
Q2552-07	WS0725-PT-RVOA Water		1,2-Dichloropropane	14.4		0.13	0.50	ug/L
Q2552-07	WS0725-PT-RVOA Water		Toluene	2.30		0.11	0.50	ug/L
Q2552-07	WS0725-PT-RVOA Water		1,1,2-Trichloroethane	14.3		0.13	0.50	ug/L
Q2552-07	WS0725-PT-RVOA Water		Tetrachloroethene	13.1		0.14	0.50	ug/L
Q2552-07	WS0725-PT-RVOA Water		Chlorobenzene	17.8	E	0.11	0.50	ug/L
Q2552-07	WS0725-PT-RVOA Water		Ethyl Benzene	16.4	E	0.41	0.50	ug/L
Q2552-07	WS0725-PT-RVOA Water		Total Xylenes	26.6		1.09	1.50	ug/L
Q2552-07	WS0725-PT-RVOA Water		m/p-Xylenes	13.4		0.73	1.00	ug/L
Q2552-07	WS0725-PT-RVOA Water		o-Xylene	13.2		0.36	0.50	ug/L
Q2552-07	WS0725-PT-RVOA Water		Styrene	2.60		0.38	0.50	ug/L
Q2552-07	WS0725-PT-RVOA Water		1,4-Dichlorobenzene	6.10		0.13	0.50	ug/L
Q2552-07	WS0725-PT-RVOA Water		1,2-Dichlorobenzene	5.60		0.14	0.50	ug/L
Q2552-07	WS0725-PT-RVOA Water		1,2,4-Trichlorobenzene	16.9	E	0.21	0.50	ug/L
Total Voc :				265				
Total Concentration:				265				
Client ID:	WS0725-PT-RVOA-WSDL							
Q2552-07DL	WS0725-PT-RVOA Water		Vinyl Chloride	20.0	D	0.48	2.00	ug/L
Q2552-07DL	WS0725-PT-RVOA Water		1,1-Dichloroethene	7.40	D	0.52	2.00	ug/L
Q2552-07DL	WS0725-PT-RVOA Water		Methylene Chloride	14.4	D	1.80	2.00	ug/L
Q2552-07DL	WS0725-PT-RVOA Water		trans-1,2-Dichloroethene	16.8	D	0.56	2.00	ug/L
Q2552-07DL	WS0725-PT-RVOA Water		Carbon Tetrachloride	2.00	D	0.52	2.00	ug/L
Q2552-07DL	WS0725-PT-RVOA Water		cis-1,2-Dichloroethene	4.90	D	0.52	2.00	ug/L
Q2552-07DL	WS0725-PT-RVOA Water		1,1,1-Trichloroethane	15.3	D	0.48	2.00	ug/L
Q2552-07DL	WS0725-PT-RVOA Water		Benzene	12.9	D	0.44	2.00	ug/L
Q2552-07DL	WS0725-PT-RVOA Water		1,2-Dichloroethane	12.1	D	0.56	2.00	ug/L
Q2552-07DL	WS0725-PT-RVOA Water		Trichloroethene	9.80	D	0.52	2.00	ug/L

Hit Summary Sheet SW-846

SDG No.: Q2552

Client: Alliance Technical Group, LLC - Newark

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q2552-07DL	WS0725-PT-RVOA	Water	1,2-Dichloropropane	13.4	D	0.52	2.00	ug/L
Q2552-07DL	WS0725-PT-RVOA	Water	Toluene	1.90	JD	0.44	2.00	ug/L
Q2552-07DL	WS0725-PT-RVOA	Water	1,1,2-Trichloroethane	13.4	D	0.52	2.00	ug/L
Q2552-07DL	WS0725-PT-RVOA	Water	Tetrachloroethene	11.3	D	0.56	2.00	ug/L
Q2552-07DL	WS0725-PT-RVOA	Water	Chlorobenzene	15.5	D	0.44	2.00	ug/L
Q2552-07DL	WS0725-PT-RVOA	Water	Ethyl Benzene	13.9	D	1.60	2.00	ug/L
Q2552-07DL	WS0725-PT-RVOA	Water	Total Xylenes	21.8	D	4.30	6.00	ug/L
Q2552-07DL	WS0725-PT-RVOA	Water	m/p-Xylenes	10.9	D	2.90	4.00	ug/L
Q2552-07DL	WS0725-PT-RVOA	Water	o-Xylene	10.9	D	1.40	2.00	ug/L
Q2552-07DL	WS0725-PT-RVOA	Water	1,4-Dichlorobenzene	5.10	D	0.52	2.00	ug/L
Q2552-07DL	WS0725-PT-RVOA	Water	1,2-Dichlorobenzene	4.80	D	0.56	2.00	ug/L
Q2552-07DL	WS0725-PT-RVOA	Water	1,2,4-Trichlorobenzene	13.4	D	0.84	2.00	ug/L
Total Voc :				230				
Total Concentration:				230				