

Surrogate Summary

SDG No.: Q2552

Client: Alliance Technical Group, LLC - Newark

Analytical Method: SW524.2

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2552-07	WS0725-PT-RVOA-WS	1,2-Dichlorobenzene-d4	1	1.01	101		70	130
		4-Bromofluorobenzene	1	1.00	100		70	130
Q2552-07DL	WS0725-PT-RVOA-WSDL	1,2-Dichlorobenzene-d4	1	1.02	102		70	130
		4-Bromofluorobenzene	1	1.03	103		70	130
VU0717WBL01	VU0717WBL01	1,2-Dichlorobenzene-d4	1	0.94	94		70	130
		4-Bromofluorobenzene	1	1.07	107		70	130
VU0717WBS01	VU0717WBS01	1,2-Dichlorobenzene-d4	1	0.94	94		70	130
		4-Bromofluorobenzene	1	1.00	100		70	130
VU0717WBSD0	VU0717WBSD01	1,2-Dichlorobenzene-d4	1	1.10	110		70	130
		4-Bromofluorobenzene	1	1.03	103		70	130
VU0718WBL01	VU0718WBL01	1,2-Dichlorobenzene-d4	1	0.72	72		70	130
		4-Bromofluorobenzene	1	0.79	79		70	130
VU0718WBS02	VU0718WBS02	1,2-Dichlorobenzene-d4	1	1.05	105		70	130
		4-Bromofluorobenzene	1	1.03	103		70	130