

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ALLI03  
 Lab Code: ACE SDG NO.: Q2552  
 Lab File ID: VU063532.D Date Analyzed: 07/18/2025  
 Instrument ID: MSVOA\_U Time Analyzed: 09:11  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                     | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT # | IS3<br>AREA # | RT # |
|---------------------|---------------|-------|---------------|------|---------------|------|
| 12 HOUR STD         | 22137         | 6.10  | 0             | 0.00 | 0             | 0.00 |
| UPPER LIMIT         | 28778.1       | 6.597 | 0             |      | 0             |      |
| LOWER LIMIT         | 15495.9       | 5.597 | 0             |      | 0             |      |
| EPA SAMPLE NO.      |               |       |               |      |               |      |
| WS0725-PT-RVOA-WSDL | 16405         | 6.10  | 0             | 0.00 | 0             | 0.00 |
| VU0718WBL01         | 25256         | 6.10  | 0             | 0.00 | 0             | 0.00 |
| VU0718WBS02         | 20233         | 6.10  | 0             | 0.00 | 0             | 0.00 |

IS1 = Fluorobenzene

IS2 =

IS3 =

AREA UPPER LIMIT = +30% of internal standard area

AREA LOWER LIMIT = -30% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.