

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ALLI03
 Lab Code: ACE SDG NO.: Q2552
 Lab File ID: VU063532.D Date Analyzed: 07/18/2025
 Instrument ID: MSVOA_U Time Analyzed: 09:11
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	0	0				
UPPER LIMIT	0					
LOWER LIMIT	0					
EPA SAMPLE NO.						
WS0725-PT-RVOA-WSDL	0	0.00				
VU0718WBL01	0	0.00				
VU0718WBS02	0	0.00				

IS4 =

AREA UPPER LIMIT = +30% of internal standard area
 AREA LOWER LIMIT = -30% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.