

Instrument ID: MSVOA_U

Daily Analysis Runlog For Sequence/QC Batch ID # VU071625

Review By	Maresh Dadoda	Review On	7/17/2025 8:19:30 AM
Supervise By	Semsettin Yesilyurt	Supervise On	7/17/2025 8:26:51 AM
SubDirectory	VU071625	HP Acquire Method	MSVOA_U
		HP Processing Method	524u071625dw.m

STD. NAME	STD REF.#
Tune/Reschk	VP134776
Initial Calibration Stds	VP134778,VP134779,VP134780,VP134781,VP134782,VP134783
CCC	
Internal Standard/PEM	VP134722
ICV/I.BLK	VP134784
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VU063510.D	16 Jul 2025 08:13		MD/SY	Ok
2	VSTDICC0.5	VSTDICC0.5	VU063511.D	16 Jul 2025 09:24		MD/SY	Ok,M
3	VSTDICC001	VSTDICC001	VU063512.D	16 Jul 2025 09:54	LR - 10 And QR - 80,86	MD/SY	Ok,M
4	VSTDICC002	VSTDICC002	VU063513.D	16 Jul 2025 10:27		MD/SY	Ok,M
5	VSTDICC005	VSTDICC005	VU063514.D	16 Jul 2025 11:00		MD/SY	Ok,M
6	VSTDICCC010	VSTDICCC010	VU063515.D	16 Jul 2025 11:37		MD/SY	Ok,M
7	VSTDICC015	VSTDICC015	VU063516.D	16 Jul 2025 12:11		MD/SY	Ok,M
8	VSTDICV010	ICVVU071625	VU063517.D	16 Jul 2025 14:19	ICV failed for com.#13,69,80	MD/SY	Ok,M

M : Manual Integration

Instrument ID: MSVOA_U

Daily Analysis Runlog For Sequence/QC Batch ID # VU071825

Review By	Mahesh Dadoda	Review On	7/21/2025 1:56:39 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	7/21/2025 9:34:13 PM		
SubDirectory	VU071825	HP Acquire Method	MSVOA_U	HP Processing Method	524u071625dw.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP134835			
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard		VP134836,VP134837,RL-VP134838,VP134839			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VU063531.D	18 Jul 2025 08:07		MD/SY	Ok
2	VSTDCCC010	VSTDCCC010	VU063532.D	18 Jul 2025 09:11	524.2 PT study	MD/SY	Ok,M
3	VU0718WBL01	VU0718WBL01	VU063533.D	18 Jul 2025 10:49		MD/SY	Ok
4	VU0718WBS01	VU0718WBS01	VU063534.D	18 Jul 2025 11:20	Recovery failed	MD/SY	Not Ok
5	VU0718WBS02	VU0718WBS02	VU063535.D	18 Jul 2025 11:50		MD/SY	Ok,M
6	VU0718WBSD02	VU0718WBSD02	VU063536.D	18 Jul 2025 13:28		MD/SY	Ok,M
7	Q2552-11	WS0725-PT-EDBCP-W	VU063537.D	18 Jul 2025 15:24	prep error	MD/SY	Not Ok
8	Q2552-08	WS0725-PT-UNRVOA-	VU063538.D	18 Jul 2025 15:54	Need 4X(UNRVOA)	MD/SY	Dilution
9	Q2552-08DL	WS0725-PT-UNRVOA-	VU063539.D	18 Jul 2025 16:24		MD/SY	Ok
10	VIBLK	VIBLK	VU063540.D	18 Jul 2025 16:55		MD/SY	Ok
11	Q2552-07DL	WS0725-PT-RVOA-WS	VU063541.D	18 Jul 2025 17:25	RVOA	MD/SY	Ok
12	RL-CHECK	RL-CHECK	VU063542.D	18 Jul 2025 17:56	RL-0.5 ppb	MD/SY	Ok,M
13	RL-CHECK	RL-CHECK	VU063543.D	18 Jul 2025 18:26	RL-1.0 ppb	MD/SY	Ok,M
14	VSTDCCC010	VSTDCCC010EC	VU063544.D	18 Jul 2025 18:56		MD/SY	Ok,M

M : Manual Integration