

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2552</u>	Analytical Method:	<u>SW524.2</u>
Client:	<u>Alliance Technical Group, LLC - New</u>	Datafile :	<u>VU063522.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VU0717WBSD01	Chloroform	2	2.20	ug/L	110	5		70	130	20
	Bromodichloromethane	2	1.80	ug/L	90	0		70	130	20
	Dibromochloromethane	2	1.80	ug/L	90	0		70	130	20
	Bromoform	2	1.60	ug/L	80	0		70	130	20