

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>ALLI03</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2552</u>
Instrument ID:	<u>MSVOA_U</u>	Calibration Date/Time:	<u>07/21/2025 09:13</u>
Lab File ID:	<u>VU063546.D</u>	Init. Calib. Date(s):	<u>07/16/2025 07/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:24 12:11</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
1,2-Dibromoethane	0.172	0.164		-4.65	30
1,2,3-Trichloropropane	0.190	0.175		-7.89	30
1,2-Dibromo-3-Chloropropane	0.033	0.036		9.09	30
1,2-Dichlorobenzene-d4	0.350	0.349		-0.29	30
4-Bromofluorobenzene	0.379	0.428		12.93	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.