

284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900,

Fax: 908 789 8922

## **Report of Analysis**

Client: Alliance Technical Group, LLC - Newark Date Collected:

Project: NJ Drinking Water PT Date Received:

Client Sample ID: VU0721WBSD01 SDG No.: Q2552
Lab Sample ID: VU0721WBSD01 Matrix: Water

Analytical Method: E524.2 % Solid: 0

Sample Wt/Vol: 25 Units: mL Final Vol: 25000 uL

Soil Aliquot Vol: uL Test: VOCMS Group5

GC Column: DB-624UI ID: 0.18 Level: LOW

Prep Method:

File ID/Qc Batch: Dilution: Date Analyzed Prep Batch ID VU063549.D 1 07/21/25 14:07 VU072125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
106-93-4	1,2-Dibromoethane	1.80		0.13	0.50	ug/L
96-18-4	1,2,3-Trichloropropane	1.80		0.19	0.50	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	1.80		0.24	0.50	ug/L
SURROGATES						
2199-69-1	1,2-Dichlorobenzene-d4	0.88		70 - 130	88%	SPK: 1
460-00-4	4-Bromofluorobenzene	0.83		70 - 130	83%	SPK: 1
INTERNAL STA	ANDARDS					
462-06-6	Fluorobenzene	22400	6.1			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products