

Instrument ID: MSVOA\_U

**Daily Analysis Runlog For Sequence/QC Batch ID # VU071625**

|              |                     |                      |                      |
|--------------|---------------------|----------------------|----------------------|
| Review By    | Maresh Dadoda       | Review On            | 7/17/2025 8:19:30 AM |
| Supervise By | Semsettin Yesilyurt | Supervise On         | 7/17/2025 8:26:51 AM |
| SubDirectory | VU071625            | HP Acquire Method    | MSVOA_U              |
|              |                     | HP Processing Method | 524u071625dw.m       |

| STD. NAME                | STD REF.#                                             |
|--------------------------|-------------------------------------------------------|
| Tune/Reschk              | VP134776                                              |
| Initial Calibration Stds | VP134778,VP134779,VP134780,VP134781,VP134782,VP134783 |
| CCC                      |                                                       |
| Internal Standard/PEM    | VP134722                                              |
| ICV/I.BLK                | VP134784                                              |
| Surrogate Standard       |                                                       |
| MS/MSD Standard          |                                                       |
| LCS Standard             |                                                       |

| Sr# | SampleID    | ClientID    | Data File Name | Date-Time         | Comment                      | Operator | Status |
|-----|-------------|-------------|----------------|-------------------|------------------------------|----------|--------|
| 1   | BFB         | BFB         | VU063510.D     | 16 Jul 2025 08:13 |                              | MD/SY    | Ok     |
| 2   | VSTDICC0.5  | VSTDICC0.5  | VU063511.D     | 16 Jul 2025 09:24 |                              | MD/SY    | Ok,M   |
| 3   | VSTDICC001  | VSTDICC001  | VU063512.D     | 16 Jul 2025 09:54 | LR - 10 And QR - 80,86       | MD/SY    | Ok,M   |
| 4   | VSTDICC002  | VSTDICC002  | VU063513.D     | 16 Jul 2025 10:27 |                              | MD/SY    | Ok,M   |
| 5   | VSTDICC005  | VSTDICC005  | VU063514.D     | 16 Jul 2025 11:00 |                              | MD/SY    | Ok,M   |
| 6   | VSTDICCC010 | VSTDICCC010 | VU063515.D     | 16 Jul 2025 11:37 |                              | MD/SY    | Ok,M   |
| 7   | VSTDICC015  | VSTDICC015  | VU063516.D     | 16 Jul 2025 12:11 |                              | MD/SY    | Ok,M   |
| 8   | VSTDICV010  | ICVVU071625 | VU063517.D     | 16 Jul 2025 14:19 | ICV failed for com.#13,69,80 | MD/SY    | Ok,M   |

M : Manual Integration

Instrument ID: MSVOA\_U

**Daily Analysis Runlog For Sequence/QC Batch ID # VU072125**

|              |                     |                      |                      |
|--------------|---------------------|----------------------|----------------------|
| Review By    | Maresh Dadoda       | Review On            | 7/22/2025 1:42:44 PM |
| Supervise By | Semsettin Yesilyurt | Supervise On         | 7/22/2025 1:55:20 PM |
| SubDirectory | VU072125            | HP Acquire Method    | MSVOA_U              |
|              |                     | HP Processing Method | 524u071625dw.m       |

| STD. NAME                                                                                          | STD REF.#                               |
|----------------------------------------------------------------------------------------------------|-----------------------------------------|
| Tune/Reschk<br>Initial Calibration Stds                                                            | VP134867                                |
| CCC<br>Internal Standard/PEM<br>ICV/I.BLK<br>Surrogate Standard<br>MS/MSD Standard<br>LCS Standard | VP134871,VP134872,MDL-VP134879,VP134880 |

| Sr# | SampleID     | ClientID          | Data File Name | Date-Time         | Comment        | Operator | Status |
|-----|--------------|-------------------|----------------|-------------------|----------------|----------|--------|
| 1   | BFB          | BFB               | VU063545.D     | 21 Jul 2025 08:14 |                | MD/SY    | Ok     |
| 2   | VSTDCCC010   | VSTDCCC010        | VU063546.D     | 21 Jul 2025 09:13 | 524.2 PT study | MD/SY    | Ok,M   |
| 3   | VU0721WBL01  | VU0721WBL01       | VU063547.D     | 21 Jul 2025 11:13 |                | MD/SY    | Ok     |
| 4   | VU0721WBS01  | VU0721WBS01       | VU063548.D     | 21 Jul 2025 12:52 |                | MD/SY    | Ok,M   |
| 5   | VU0721WBSD01 | VU0721WBSD01      | VU063549.D     | 21 Jul 2025 14:07 |                | MD/SY    | Ok,M   |
| 6   | Q2552-11     | WS0725-PT-EDBCP-W | VU063550.D     | 21 Jul 2025 15:38 | PT-EDBCP       | MD/SY    | Ok     |
| 7   | RL-CHECK     | RL-CHECK          | VU063551.D     | 21 Jul 2025 16:42 |                | MD/SY    | Ok,M   |
| 8   | RL-CHECK     | RL-CHECK          | VU063552.D     | 21 Jul 2025 17:12 |                | MD/SY    | Ok,M   |
| 9   | VSTDCCC010   | VSTDCCC010EC      | VU063553.D     | 21 Jul 2025 19:14 |                | MD/SY    | Ok,M   |

M : Manual Integration