

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2552

Analytical Method: SW524.2

Client: Alliance Technical Group, LLC - New

Datafile : VU063548.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VU0721WBS01	1,2-Dibromoethane	2	1.90	ug/L	95			70	130	
	1,2,3-Trichloropropane	2	2.00	ug/L	100			70	130	
	1,2-Dibromo-3-Chloropropane	2	1.90	ug/L	95			70	130	