

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ALLI03
 Lab Code: ACE SDG NO.: Q2552
 Lab File ID: VU063546.D Date Analyzed: 07/21/2025
 Instrument ID: MSVOA_U Time Analyzed: 09:13
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	25321	6.10	0	0.00	0	0.00
UPPER LIMIT	32917.3	6.6	0		0	
LOWER LIMIT	17724.7	5.6	0		0	
EPA SAMPLE NO.						
WS0725-PT-EDBCP-WS	21983	6.10	0	0.00	0	0.00
VU0721WBL01	22148	6.10	0	0.00	0	0.00
VU0721WBS01	21614	6.10	0	0.00	0	0.00
VU0721WBSD01	22358	6.10	0	0.00	0	0.00

IS1 = Fluorobenzene

IS2 =

IS3 =

AREA UPPER LIMIT = +30% of internal standard area

AREA LOWER LIMIT = -30% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.