

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2553

Client ID : SSTDCCC040

Date Analyzed: 07/14/2025

Lab File ID: BM050420.D

Time Analyzed: 10:50

Instrument ID: BNA_M

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	419216	7.863	1606680	10.66	1065310	14.49
UPPER LIMIT	838432	8.363	3213360	11.157	2130620	14.986
LOWER LIMIT	209608	7.363	803340	10.157	532655	13.986
EPA SAMPLE NO.						
01 PB168816BL	402471	7.86	1401210	10.66	832017	14.49

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.