



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
 Fax : 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2553

Client ID : SSTDCCC040

Date Analyzed: 07/14/2025

Lab File ID: BP025121.D

Time Analyzed: 15:36

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	502754	7.437	2056760	10.19	1334830	14.08
UPPER LIMIT	1005510	7.937	4113520	10.69	2669660	14.584
LOWER LIMIT	251377	6.937	1028380	9.69	667415	13.584
EPA SAMPLE NO.						
01 AOC-201	556276	7.43	2052330	10.18	1263040	14.08
02 AOC-202	459808	7.44	1740090	10.18	1019900	14.08
03 AOC-203	356736	7.44	1359330	10.18	848645	14.08
04 AOC-205	458642	7.44	1720260	10.18	992285	14.08

IS1 (DCB) = 1,4-Dichlorobenzene-d4
 IS2 (NPT) = Naphthalene-d8
 IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.