

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX070225

Review By	John Carlone	Review On	7/3/2025 8:44:30 AM
Supervise By	Mahesh Dadoda	Supervise On	7/3/2025 3:09:59 PM
SubDirectory	VX070225	HP Acquire Method	HP Processing Method 82X070225W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134642 VP134633,VP134634,VP134635,VP134638,VP134639,VP134640 VP134641

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046859.D	02 Jul 2025 11:12		JC/MD	Ok
2	VSTDICC001	VSTDICC001	VX046860.D	02 Jul 2025 12:11		JC/MD	Ok,M
3	VSTDICC005	VSTDICC005	VX046861.D	02 Jul 2025 12:37	%D Failed for com.#13 in 5 ppb	JC/MD	Ok,M
4	VSTDICC020	VSTDICC020	VX046862.D	02 Jul 2025 13:18		JC/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VX046863.D	02 Jul 2025 13:39	LR-13	JC/MD	Ok,M
6	VSTDICC100	VSTDICC100	VX046864.D	02 Jul 2025 14:10		JC/MD	Ok,M
7	VSTDICC150	VSTDICC150	VX046865.D	02 Jul 2025 14:31		JC/MD	Ok,M
8	VIBLK	VIBLK	VX046866.D	02 Jul 2025 14:53		JC/MD	Ok
9	VSTDICV050	ICVVX070225	VX046867.D	02 Jul 2025 15:14		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX071025

Review By	John Carlone	Review On	7/11/2025 8:53:41 AM
Supervise By	Maresh Dadoda	Supervise On	7/14/2025 7:47:20 AM
SubDirectory	VX071025	HP Acquire Method	HP Processing Method 82X070225W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134710
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134714,VP134715

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046932.D	10 Jul 2025 08:43		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX046933.D	10 Jul 2025 09:54		JC/MD	Ok,M
3	VX0710MBL01	VX0710MBL01	VX046934.D	10 Jul 2025 10:22		JC/MD	Ok
4	VX0710WBL01	VX0710WBL01	VX046935.D	10 Jul 2025 11:44		JC/MD	Ok
5	VX0710WBS01	VX0710WBS01	VX046936.D	10 Jul 2025 12:10		JC/MD	Ok,M
6	VX0710WBSD01	VX0710WBSD01	VX046937.D	10 Jul 2025 12:35		JC/MD	Ok,M
7	IBLK	IBLK	VX046938.D	10 Jul 2025 12:57		JC/MD	Ok
8	Q2553-06	TB	VX046939.D	10 Jul 2025 13:18	vial A pH<2 TB	JC/MD	Ok
9	Q2553-05	FB	VX046940.D	10 Jul 2025 13:39	vial A pH<2 FB;Hit of com.#17	JC/MD	ReRun
10	Q2554-02	250709059-10 Trip blan	VX046941.D	10 Jul 2025 14:00	vial A pH#6.0 TB	JC/MD	Ok
11	Q2554-04	250709059-11 Trip blan	VX046942.D	10 Jul 2025 14:21	vial A pH#6.0 TB	JC/MD	Ok
12	Q2553-01	AOC-201	VX046943.D	10 Jul 2025 14:43	vial A pH<2	JC/MD	Ok,M
13	Q2553-02	AOC-202	VX046944.D	10 Jul 2025 15:04	vial A pH<2	JC/MD	Ok,M
14	Q2553-03	AOC-203	VX046945.D	10 Jul 2025 15:26	vial A pH<2	JC/MD	Ok,M
15	Q2553-04	AOC-205	VX046946.D	10 Jul 2025 15:47	vial A pH<2	JC/MD	Ok,M
16	PB168762TB	PB168762TB	VX046947.D	10 Jul 2025 16:08		JC/MD	Ok,M
17	Q2517-04	TP-14	VX046948.D	10 Jul 2025 16:30	vial A pH#5.0	JC/MD	Ok
18	Q2519-04	TP-15	VX046949.D	10 Jul 2025 16:51	vial A pH#5.0	JC/MD	Ok,M

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX071025

Review By	John Carlone	Review On	7/11/2025 8:53:41 AM
Supervise By	Maresh Dadoda	Supervise On	7/14/2025 7:47:20 AM
SubDirectory	VX071025	HP Acquire Method	HP Processing Method 82X070225W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134710		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134714,VP134715		

19	Q2554-03	250709104-01 VOA	VX046950.D	10 Jul 2025 17:12	vial A pH#6.0	JC/MD	Ok
20	Q2554-01	250709071-01 VOA	VX046951.D	10 Jul 2025 17:34	vial A pH#6.0 Need 5X. No more vials.	JC/MD	Ok,M
21	IBLK	IBLK	VX046952.D	10 Jul 2025 17:55		JC/MD	Ok
22	VX0710MBS01	VX0710MBS01	VX046953.D	10 Jul 2025 18:16	BS Failed Low for com.#05	JC/MD	Ok,M
23	VX0710MBSD01	VX0710MBSD01	VX046954.D	10 Jul 2025 18:37	BSD Failed Low for com.#05	JC/MD	Ok,M
24	Q2480-05	GPX5	VX046955.D	10 Jul 2025 18:59		JC/MD	Ok,M
25	Q2480-06ME	GPX6ME	VX046956.D	10 Jul 2025 19:20		JC/MD	Ok,M
26	Q2480-07ME	GPX7ME	VX046957.D	10 Jul 2025 19:41		JC/MD	Ok,M
27	Q2480-08	GPX8	VX046958.D	10 Jul 2025 20:02		JC/MD	Ok,M
28	VSTDCCC050	VSTDCCC050EC	VX046959.D	10 Jul 2025 20:23		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX071425

Review By	Mahesh Dadoda	Review On	7/15/2025 4:54:26 AM
Supervise By	Semsettin Yesilyurt	Supervise On	7/15/2025 4:55:39 AM
SubDirectory	VX071425	HP Acquire Method	HP Processing Method 82X070225W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134751
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134752,VP134753

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046960.D	14 Jul 2025 08:10		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX046961.D	14 Jul 2025 09:16	pH#Lot#V12668	JC/MD	Ok,M
3	VX0714MBL01	VX0714MBL01	VX046962.D	14 Jul 2025 10:54		JC/MD	Ok
4	VX0714WBL01	VX0714WBL01	VX046963.D	14 Jul 2025 11:18		JC/MD	Ok
5	VX0714MBS01	VX0714MBS01	VX046964.D	14 Jul 2025 11:39		JC/MD	Ok,M
6	VX0714MBSD01	VX0714MBSD01	VX046965.D	14 Jul 2025 12:06		JC/MD	Ok,M
7	Q2480-04ME	GPX4ME	VX046966.D	14 Jul 2025 12:27		JC/MD	Ok
8	Q2480-02ME	GPX2ME	VX046967.D	14 Jul 2025 12:49		JC/MD	Ok,M
9	VX0714WBS01	VX0714WBS01	VX046968.D	14 Jul 2025 13:10		JC/MD	Ok,M
10	VX0714WBSD01	VX0714WBSD01	VX046969.D	14 Jul 2025 13:37		JC/MD	Ok,M
11	Q2523-01	MOO-25-0190	VX046970.D	14 Jul 2025 13:58	Need Soil Run	JC/MD	Not Ok
12	Q2526-02	LAW-25-0101	VX046971.D	14 Jul 2025 14:19	wipe sample	JC/MD	Ok
13	Q2565-03	MOO-25-0191	VX046972.D	14 Jul 2025 14:41	wipe sample	JC/MD	Ok
14	Q2565-04	MOO-25-0196	VX046973.D	14 Jul 2025 15:02	wipe sample	JC/MD	Ok,M
15	Q2565-05	MOO-25-0180	VX046974.D	14 Jul 2025 15:23	Need Soil Run	JC/MD	Not Ok
16	IBLK	IBLK	VX046975.D	14 Jul 2025 15:45		JC/MD	Ok
17	Q2578-01	WC-A5-01-G	VX046976.D	14 Jul 2025 16:06	Surrogate fail	JC/MD	ReRun
18	Q2578-05	WC-A2-09-G	VX046977.D	14 Jul 2025 16:28	Surrogate fail	JC/MD	ReRun

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX071425

Review By	Maresh Dadoda	Review On	7/15/2025 4:54:26 AM
Supervise By	Semsettin Yesilyurt	Supervise On	7/15/2025 4:55:39 AM
SubDirectory	VX071425	HP Acquire Method	HP Processing Method 82X070225W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134751		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134752,VP134753		

19	Q2578-09	WC-A2-10-G	VX046978.D	14 Jul 2025 16:49	Surrogate fail	JC/MD	ReRun
20	Q2578-13	WC-A2-11-G	VX046979.D	14 Jul 2025 17:10	Surrogate fail	JC/MD	ReRun
21	Q2578-17	WC-A2-12-G	VX046980.D	14 Jul 2025 17:31	Surrogate fail	JC/MD	ReRun
22	Q2579-01	WC-A2-13-G	VX046981.D	14 Jul 2025 17:53	Surrogate fail	JC/MD	ReRun
23	Q2579-05	WC-A2-14-G	VX046982.D	14 Jul 2025 18:14	Surrogate fail	JC/MD	ReRun
24	IBLK	IBLK	VX046983.D	14 Jul 2025 18:35		JC/MD	Ok
25	Q2553-05	FB	VX046984.D	14 Jul 2025 18:56	vial B pH<2 FB	JC/MD	Ok
26	Q2595-01	CHRT-24849	VX046985.D	14 Jul 2025 19:17	jelly like sample	JC/MD	Ok
27	VSTDCCC050	VSTDCCC050EC	VX046986.D	14 Jul 2025 19:38		JC/MD	Ok,M

M : Manual Integration