

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: FIRS02  
 Lab Code: ACE SDG NO.: Q2553  
 Lab File ID: VX046933.D Date Analyzed: 07/10/2025  
 Instrument ID: MSVOA\_X Time Analyzed: 09:54  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|----------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD    | 309661        | 5.56  | 506196        | 6.76  | 449744        | 10.05  |
| UPPER LIMIT    | 619322        | 6.056 | 1012390       | 7.263 | 899488        | 10.549 |
| LOWER LIMIT    | 154831        | 5.056 | 253098        | 6.263 | 224872        | 9.549  |
| EPA SAMPLE NO. |               |       |               |       |               |        |
| AOC-201        | 320576        | 5.56  | 555832        | 6.77  | 518959        | 10.06  |
| AOC-202        | 457019        | 5.57  | 802756        | 6.77  | 753351        | 10.06  |
| AOC-203        | 394931        | 5.56  | 673452        | 6.77  | 623779        | 10.06  |
| AOC-205        | 431287        | 5.56  | 760941        | 6.77  | 704204        | 10.06  |
| TB             | 354436        | 5.56  | 642674        | 6.77  | 599480        | 10.06  |
| VX0710WBL01    | 311309        | 5.56  | 568946        | 6.76  | 543608        | 10.06  |
| VX0710WBS01    | 342120        | 5.56  | 567066        | 6.76  | 521192        | 10.06  |
| VX0710WBSD01   | 319533        | 5.56  | 525170        | 6.76  | 482068        | 10.06  |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.