

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>FIRS02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2553</u>
Lab File ID:	<u>VX046933.D</u>	Date Analyzed:	<u>07/10/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>09:54</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	231018	12.018				
UPPER LIMIT	462036	12.518				
LOWER LIMIT	115509	11.518				
EPA SAMPLE NO.						
AOC-201	227982	12.02				
AOC-202	386856	12.02				
AOC-203	297586	12.02				
AOC-205	362481	12.02				
TB	308124	12.02				
VX0710WBL01	271904	12.02				
VX0710WBS01	270401	12.02				
VX0710WBSD01	249133	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.