

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>FIRS02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2553</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>07/10/2025 09:54</u>
Lab File ID:	<u>VX046933.D</u>	Init. Calib. Date(s):	<u>07/02/2025 07/02/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>12:11 14:31</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.478	0.491		2.72	20
Chloromethane	0.522	0.530	0.1	1.53	20
Vinyl Chloride	0.569	0.575		1.05	20
Bromomethane	0.366	0.384		4.92	20
Chloroethane	0.376	0.378		0.53	20
Trichlorofluoromethane	0.880	0.925		5.11	20
1,1,2-Trichlorotrifluoroethane	0.561	0.602		7.31	20
1,1-Dichloroethene	0.542	0.561		3.51	20
Acetone	0.205	0.237		15.61	20
Carbon Disulfide	1.422	1.331		-6.4	20
Methyl tert-butyl Ether	1.531	1.789		16.85	20
Methyl Acetate	0.541	0.697		28.83	20
Methylene Chloride	0.607	0.664		9.39	20
trans-1,2-Dichloroethene	0.555	0.582		4.86	20
1,1-Dichloroethane	1.064	1.156	0.1	8.65	20
Cyclohexane	0.944	0.949		0.53	20
2-Butanone	0.274	0.323		17.88	20
Carbon Tetrachloride	0.484	0.514		6.2	20
cis-1,2-Dichloroethene	0.677	0.724		6.94	20
Bromochloromethane	0.525	0.563		7.24	20
Chloroform	1.087	1.184		8.92	20
1,1,1-Trichloroethane	0.908	0.966		6.39	20
Methylcyclohexane	0.573	0.572		-0.17	20
Benzene	1.385	1.459		5.34	20
1,2-Dichloroethane	0.470	0.518		10.21	20
Trichloroethene	0.361	0.366		1.38	20
1,2-Dichloropropane	0.350	0.379		8.29	20
Bromodichloromethane	0.518	0.576		11.2	20
4-Methyl-2-Pentanone	0.353	0.415		17.56	20
Toluene	0.858	0.906		5.59	20
t-1,3-Dichloropropene	0.462	0.530		14.72	20
cis-1,3-Dichloropropene	0.527	0.594		12.71	20
1,1,2-Trichloroethane	0.320	0.359		12.19	20
2-Hexanone	0.234	0.282		20.51	20
Dibromochloromethane	0.383	0.429		12.01	20
1,2-Dibromoethane	0.321	0.361		12.46	20
Tetrachloroethene	0.333	0.343		3	20
Chlorobenzene	1.081	1.156	0.3	6.94	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>FIRS02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2553</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>07/10/2025 09:54</u>
Lab File ID:	<u>VX046933.D</u>	Init. Calib. Date(s):	<u>07/02/2025 07/02/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>12:11 14:31</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.851	1.991		7.56	20
m/p-Xylenes	0.698	0.748		7.16	20
o-Xylene	0.674	0.731		8.46	20
Styrene	1.153	1.280		11.02	20
Bromoform	0.270	0.310	0.1	14.81	20
Isopropylbenzene	3.515	3.800		8.11	20
1,1,2,2-Tetrachloroethane	0.966	1.071	0.3	10.87	20
1,3-Dichlorobenzene	1.637	1.724		5.32	20
1,4-Dichlorobenzene	1.704	1.741		2.17	20
1,2-Dichlorobenzene	1.587	1.671		5.29	20
1,2-Dibromo-3-Chloropropane	0.169	0.191		13.02	20
1,2,4-Trichlorobenzene	1.078	1.137		5.47	20
1,2,3-Trichlorobenzene	1.019	1.098		7.75	20
1,2-Dichloroethane-d4	0.661	0.672		1.66	20
Dibromofluoromethane	0.339	0.344		1.48	20
Toluene-d8	1.197	1.173		-2.01	20
4-Bromofluorobenzene	0.455	0.452		-0.66	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.