

## **DATA PACKAGE**

VOLATILE ORGANICS

**PROJECT NAME : WASTE WATER 2025**

**GARDEN STATE LABORATORIES, INC.**

**410 Hillside Avenue**

**Hillside, NJ - 07205**

**Phone No: 800-273-8901**

**ORDER ID : Q2554**

**ATTENTION : Sharon Ercoliani**



**Laboratory Certification ID # 20012**



1) Signature Page	3
2) Case Narrative	4
2.1) VOCMS Group1- Case Narrative	4
2.2) VOCMS Group2- Case Narrative	6
3) Qualifier Page	8
4) QA Checklist	9
5) VOCMS Group1 Data	10
6) VOCMS Group2 Data	15
7) Shipping Document	24
7.1) CHAIN OF CUSTODY	25
7.2) Lab Certificate	26
7.3) Internal COC	27

1

2

3

4

5

6

7

## Cover Page

**Order ID :** Q2554

**Project ID :** Waste Water 2025

**Client :** Garden State Laboratories, Inc.

**Lab Sample Number**

Q2554-03  
Q2554-04

**Client Sample Number**

250709104-01 VOA  
250709059-11 Trip blank

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : \_\_\_\_\_

**APPROVED**

*By Nimisha Pandya, QA/QC Supervisor at 8:23 am, Jul 29, 2025*

Date: 7/18/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

## **CASE NARRATIVE**

**Garden State Laboratories, Inc.**  
**Project Name: Waste Water 2025**  
**Project # N/A**  
**Order ID # Q2554**  
**Test Name: VOCMS Group1**

### **A. Number of Samples and Date of Receipt:**

2 Water samples were received on 07/10/2025.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: VOCMS Group1 and VOCMS Group2. This data package contains results for VOCMS Group1.

### **C. Analytical Techniques:**

The analysis performed on instrument MSVOA\_N were done using GC column Rxi-624SIL MS 30m, 0.25mm, 1.4 um, Cat. #138 68. The analysis of VOCMS Group1 was based on method 624.1.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable requirements.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria.

The Blank Spike met requirements for all samples.

The Blank Spike Duplicate met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.

### **E. Additional Comments:**

“As per method, MS/MSD is required to be performed with the sample analysis. However, Lab did not receive sufficient volume to perform the MS/MSD therefore MS/MSD were not performed for this project. However, Lab has performed LCS/LCSD instead.”

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <35% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 35% for the Initial Calibration curve for SW-846 analysis.

**F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

**APPROVED**

*By Nimisha Pandya, QA/QC Supervisor at 8:23 am, Jul 29, 2025*

Signature \_\_\_\_\_

## **CASE NARRATIVE**

**Garden State Laboratories, Inc.**  
**Project Name: Waste Water 2025**  
**Project # N/A**  
**Order ID # Q2554**  
**Test Name: VOCMS Group2**

### **A. Number of Samples and Date of Receipt:**

2 Water samples were received on 07/10/2025.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: VOCMS Group1 and VOCMS Group2. This data package contains results for VOCMS Group2.

### **C. Analytical Techniques:**

The analysis performed on instrument MSVOA\_X were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI The analysis of VOCMS Group2 was based on method 8260D.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria.

The Blank Spike met requirements for all samples.

The Blank Spike Duplicate for { VX0710WBSD01 } with File ID: VX046937.D met requirements for all samples except for 4-Methyl-2-Pentanone[120%] is failing high but no positive hit in associate sample therefore no corrective action taken.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration File ID VX046933.D met the requirements except for 2-Hexanone and Methyl Acetate are failing high but no positive hit in associate sample therefore no corrective action taken.

The Tuning criteria met requirements.

**E. Additional Comments:**

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

**F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_

**APPROVED**

*By Nimisha Pandya, QA/QC Supervisor at 8:23 am, Jul 29, 2025*

## DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
<b>U</b>	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
<b>ND</b>	Indicates the analyte was analyzed for, but not detected
<b>J</b>	Indicates an estimated value. This flag is used: <ol style="list-style-type: none"> <li>(1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.)</li> <li>(2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.</li> </ol>
<b>B</b>	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
<b>E</b>	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
<b>D</b>	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
<b>P</b>	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
<b>N</b>	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
<b>A</b>	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
<b>Q</b>	Indicates the LCS did not meet the control limits requirements



## APPENDIX A

### QA REVIEW GENERAL DOCUMENTATION

Project #: Q2554

Completed

For thorough review, the report must have the following:

#### GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

#### COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

#### CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

#### ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 07/18/2025

**Hit Summary Sheet**  
**624.1**

**SDG No.:** Q2554

**Client:** Garden State Laboratories, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

**Client ID:**

0

**Total Voc :**

**Total Concentration:**

A

B

C

D



# SAMPLE DATA

## Report of Analysis

Client:	Garden State Laboratories, Inc.		Date Collected:	07/09/25	
Project:	Waste Water 2025		Date Received:	07/10/25	
Client Sample ID:	250709104-01 VOA		SDG No.:	Q2554	
Lab Sample ID:	Q2554-03		Matrix:	Water	
Analytical Method:	E624.1		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087322.D	1	07/10/25 13:48	VN071025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
<b>TARGETS</b>						
107-02-8	Acrolein	6.60	U	6.60	25.0	ug/L
107-13-1	Acrylonitrile	2.80	U	2.80	25.0	ug/L
<b>SURROGATES</b>						
17060-07-0	1,2-Dichloroethane-d4	29.1		91 - 110	97%	SPK: 30
2037-26-5	Toluene-d8	28.0		91 - 112	93%	SPK: 30
460-00-4	4-Bromofluorobenzene	28.7		63 - 112	96%	SPK: 30
<b>INTERNAL STANDARDS</b>						
74-97-5	Bromochloromethane	39000	7.824			
540-36-3	1,4-Difluorobenzene	188000	9.106			
3114-55-4	Chlorobenzene-d5	179000	11.865			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

Client:	Garden State Laboratories, Inc.		Date Collected:	07/09/25	
Project:	Waste Water 2025		Date Received:	07/10/25	
Client Sample ID:	250709059-11 Trip blank		SDG No.:	Q2554	
Lab Sample ID:	Q2554-04		Matrix:	Water	
Analytical Method:	E624.1		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087321.D	1	07/10/25 13:27	VN071025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
<b>TARGETS</b>						
107-02-8	Acrolein	6.60	U	6.60	25.0	ug/L
107-13-1	Acrylonitrile	2.80	U	2.80	25.0	ug/L
<b>SURROGATES</b>						
17060-07-0	1,2-Dichloroethane-d4	29.3		91 - 110	98%	SPK: 30
2037-26-5	Toluene-d8	28.5		91 - 112	95%	SPK: 30
460-00-4	4-Bromofluorobenzene	26.5		63 - 112	88%	SPK: 30
<b>INTERNAL STANDARDS</b>						
74-97-5	Bromochloromethane	38000	7.824			
540-36-3	1,4-Difluorobenzene	189000	9.106			
3114-55-4	Chlorobenzene-d5	172000	11.865			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## LAB CHRONICLE

<b>OrderID:</b>	Q2554	<b>OrderDate:</b>	7/10/2025 9:19:00 AM
<b>Client:</b>	Garden State Laboratories, Inc.	<b>Project:</b>	Waste Water 2025
<b>Contact:</b>	Sharon Ercoliani	<b>Location:</b>	VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
<b>Q2554-03</b>	<b>250709104-01 VOA</b>	<b>Water</b>			<b>07/09/25</b>			<b>07/10/25</b>
			VOCMS Group1	624.1			07/10/25	
			VOCMS Group2	8260-Low			07/10/25	
<b>Q2554-04</b>	<b>250709059-11 Trip blank</b>	<b>Water</b>			<b>07/09/25</b>			<b>07/10/25</b>
			VOCMS Group1	624.1			07/10/25	
			VOCMS Group2	8260-Low			07/10/25	

**Hit Summary Sheet**  
**8260-Low**

**SDG No.:** Q2554  
**Client:** Garden State Laboratories, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
<b>Client ID: 250709104-01 VOA</b>								
Q2554-03	250709104-01 VOA Water		Chloroethane	0.74	J	0.47	1.00	ug/L
Q2554-03	250709104-01 VOA Water		Acetone	6.50		1.50	5.00	ug/L
Q2554-03	250709104-01 VOA Water		Carbon Disulfide	0.34	J	0.21	1.00	ug/L
Q2554-03	250709104-01 VOA Water		Methyl tert-butyl Ether	0.54	J	0.16	1.00	ug/L
Q2554-03	250709104-01 VOA Water		Benzene	1.60		0.15	1.00	ug/L
Q2554-03	250709104-01 VOA Water		Toluene	0.71	J	0.14	1.00	ug/L
Q2554-03	250709104-01 VOA Water		Chlorobenzene	15.5		0.12	1.00	ug/L
Q2554-03	250709104-01 VOA Water		Ethyl Benzene	2.70		0.13	1.00	ug/L
Q2554-03	250709104-01 VOA Water		m/p-Xylenes	3.20		0.24	2.00	ug/L
Q2554-03	250709104-01 VOA Water		o-Xylene	2.50		0.12	1.00	ug/L
Q2554-03	250709104-01 VOA Water		Isopropylbenzene	1.30		0.12	1.00	ug/L
Q2554-03	250709104-01 VOA Water		1,4-Dichlorobenzene	7.70		0.19	1.00	ug/L
Q2554-03	250709104-01 VOA Water		1,2-Dichlorobenzene	0.56	J	0.16	1.00	ug/L
<b>Total Voc :</b>				<b>43.9</b>				
Q2554-03	250709104-01 VOA Water		Tetrahydrofuran	* 150	J	0.99	5.00	ug/L
Q2554-03	250709104-01 VOA Water		Tert butyl alcohol	* 220	J	5.50	25.0	ug/L
Q2554-03	250709104-01 VOA Water		Diethyl Ether	* 12.5	J	0.31	1.00	ug/L
Q2554-03	250709104-01 VOA Water		n-propylbenzene	* 0.70	J	0.13	1.00	ug/L
Q2554-03	250709104-01 VOA Water		2-Chlorotoluene	* 0.62	J	0.14	1.00	ug/L
Q2554-03	250709104-01 VOA Water		1,3,5-Trimethylbenzene	* 0.30	J	0.15	1.00	ug/L
Q2554-03	250709104-01 VOA Water		1,2,4-Trimethylbenzene	* 1.80	J	0.14	1.00	ug/L
Q2554-03	250709104-01 VOA Water		Naphthalene	* 11.4	J	0.20	1.00	ug/L
Q2554-03	250709104-01 VOA Water		1,4-Dioxane	* 190	J	6.90	100	ug/L
<b>Total Tics :</b>				<b>587</b>				
<b>Total Concentration:</b>				<b>631</b>				



# SAMPLE DATA



## Report of Analysis

Client:	Garden State Laboratories, Inc.		Date Collected:	07/09/25	
Project:	Waste Water 2025		Date Received:	07/10/25	
Client Sample ID:	250709104-01 VOA		SDG No.:	Q2554	
Lab Sample ID:	Q2554-03		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group2	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX046950.D	1	07/10/25 17:12	VX071025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
<b>TARGETS</b>						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.74	J	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	6.50		1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.34	J	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.54	J	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	1.60		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	UQ	0.68	5.00	ug/L
108-88-3	Toluene	0.71	J	0.14	1.00	ug/L

## Report of Analysis

Client:	Garden State Laboratories, Inc.		Date Collected:	07/09/25	
Project:	Waste Water 2025		Date Received:	07/10/25	
Client Sample ID:	250709104-01 VOA		SDG No.:	Q2554	
Lab Sample ID:	Q2554-03		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group2	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX046950.D	1	07/10/25 17:12	VX071025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.17	U	0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.16	U	0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	0.89	U	0.89	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.15	U	0.15	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
108-90-7	Chlorobenzene	15.5		0.12	1.00	ug/L
100-41-4	Ethyl Benzene	2.70		0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	3.20		0.24	2.00	ug/L
95-47-6	o-Xylene	2.50		0.12	1.00	ug/L
100-42-5	Styrene	0.15	U	0.15	1.00	ug/L
75-25-2	Bromoform	0.19	U	0.19	1.00	ug/L
98-82-8	Isopropylbenzene	1.30		0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	7.70		0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.56	J	0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
<b>SURROGATES</b>						
17060-07-0	1,2-Dichloroethane-d4	51.1		74 - 125	102%	SPK: 50
1868-53-7	Dibromofluoromethane	47.5		75 - 124	95%	SPK: 50
2037-26-5	Toluene-d8	49.5		86 - 113	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.4		77 - 121	101%	SPK: 50
<b>INTERNAL STANDARDS</b>						
363-72-4	Pentafluorobenzene	455000	5.568			
540-36-3	1,4-Difluorobenzene	801000	6.769			
3114-55-4	Chlorobenzene-d5	734000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	378000	12.018			
<b>TENTATIVE IDENTIFIED COMPOUNDS</b>						

## Report of Analysis

Client:	Garden State Laboratories, Inc.		Date Collected:	07/09/25	
Project:	Waste Water 2025		Date Received:	07/10/25	
Client Sample ID:	250709104-01 VOA		SDG No.:	Q2554	
Lab Sample ID:	Q2554-03		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group2	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX046950.D	1	07/10/25 17:12	VX071025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
60-29-7	Diethyl Ether	12.5	J		2.15	ug/L
75-65-0	Tert butyl alcohol	220	J		2.95	ug/L
109-99-9	Tetrahydrofuran	150	J		5.01	ug/L
123-91-1	1,4-Dioxane	190	J		7.67	ug/L
103-65-1	n-propylbenzene	0.70	J		11.3	ug/L
95-49-8	2-Chlorotoluene	0.62	J		11.4	ug/L
108-67-8	1,3,5-Trimethylbenzene	0.30	J		11.5	ug/L
95-63-6	1,2,4-Trimethylbenzene	1.80	J		11.8	ug/L
91-20-3	Naphthalene	11.4	J		13.8	ug/L

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

## Report of Analysis

Client:	Garden State Laboratories, Inc.		Date Collected:	07/09/25	
Project:	Waste Water 2025		Date Received:	07/10/25	
Client Sample ID:	250709059-11 Trip blank		SDG No.:	Q2554	
Lab Sample ID:	Q2554-04		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group2	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX046942.D	1	07/10/25 14:21	VX071025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
<b>TARGETS</b>						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	1.50	U	1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	UQ	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L

## Report of Analysis

Client:	Garden State Laboratories, Inc.		Date Collected:	07/09/25	
Project:	Waste Water 2025		Date Received:	07/10/25	
Client Sample ID:	250709059-11 Trip blank		SDG No.:	Q2554	
Lab Sample ID:	Q2554-04		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group2	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX046942.D	1	07/10/25 14:21	VX071025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.17	U	0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.16	U	0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	0.89	U	0.89	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.15	U	0.15	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	1.00	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	0.24	U	0.24	2.00	ug/L
95-47-6	o-Xylene	0.12	U	0.12	1.00	ug/L
100-42-5	Styrene	0.15	U	0.15	1.00	ug/L
75-25-2	Bromoform	0.19	U	0.19	1.00	ug/L
98-82-8	Isopropylbenzene	0.12	U	0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
<b>SURROGATES</b>						
17060-07-0	1,2-Dichloroethane-d4	54.5		74 - 125	109%	SPK: 50
1868-53-7	Dibromofluoromethane	49.4		75 - 124	99%	SPK: 50
2037-26-5	Toluene-d8	50.2		86 - 113	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.4		77 - 121	101%	SPK: 50
<b>INTERNAL STANDARDS</b>						
363-72-4	Pentafluorobenzene	339000	5.562			
540-36-3	1,4-Difluorobenzene	601000	6.769			
3114-55-4	Chlorobenzene-d5	555000	10.055			
3855-82-1	1,4-Dichlorobenzene-d4	279000	12.018			

## Report of Analysis

Client:	Garden State Laboratories, Inc.		Date Collected:	07/09/25	
Project:	Waste Water 2025		Date Received:	07/10/25	
Client Sample ID:	250709059-11 Trip blank		SDG No.:	Q2554	
Lab Sample ID:	Q2554-04		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group2	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VX046942.D	1	07/10/25 14:21	VX071025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected  
 LOQ = Limit of Quantitation  
 MDL = Method Detection Limit  
 LOD = Limit of Detection  
 E = Value Exceeds Calibration Range  
 Q = indicates LCS control criteria did not meet requirements  
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value  
 B = Analyte Found in Associated Method Blank  
 N = Presumptive Evidence of a Compound  
 \* = Values outside of QC limits  
 D = Dilution  
 () = Laboratory InHouse Limit  
 A = Aldol-Condensation Reaction Products

## LAB CHRONICLE

<b>OrderID:</b>	Q2554	<b>OrderDate:</b>	7/10/2025 9:19:00 AM
<b>Client:</b>	Garden State Laboratories, Inc.	<b>Project:</b>	Waste Water 2025
<b>Contact:</b>	Sharon Ercoliani	<b>Location:</b>	VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
<b>Q2554-03</b>	<b>250709104-01 VOA</b>	<b>Water</b>			<b>07/09/25</b>			<b>07/10/25</b>
			VOCMS Group1	624.1			07/10/25	
			VOCMS Group2	8260-Low			07/10/25	
<b>Q2554-04</b>	<b>250709059-11 Trip blank</b>	<b>Water</b>			<b>07/09/25</b>			<b>07/10/25</b>
			VOCMS Group1	624.1			07/10/25	
			VOCMS Group2	8260-Low			07/10/25	



# SHIPPING DOCUMENTS



# Garden State Laboratories, Inc.

Main Lab - 410 Hillside Avenue, Hillside NJ 07205 - NJDEP Lab Cert. #20044  
Jersey Shore Lab - 54 Main Street, Waretown NJ 08758 - NJDEP Lab Cert. #15037

Tel. 800-273-8901/908-688-8900 Fax 908-688-8966 www.gslabs.com info@gslabs.com

## Office and Drop off Locations

North Jersey Office: 225 Sparta Avenue, Sparta, NJ 07871 Tel. 973-729-1827

West Jersey Office: 2050 Route 31 North, Glen Gardner, NJ 08826 Tel. 908-537-7414

## CLIENT INFORMATION (REPORT TO BE SENT TO)

Name: Garden State Laboratories, Inc.

Contact/Authorized by: Robert Szot

Mailing Address: 410 Hillside Ave.

Phone: 908-688-8900 EXT 129

City/State/Zip: Hillside, NJ. 07205

Email: rszot@gslabs.com

## SAMPLE INFORMATION

SAMPLE TYPE: WASTE WATER

SAMPLE LOCATION: ACUA SW LANDFILL LEACHATE TANKS

## FOR SAMPLE RECEIVING USE ONLY

DATE/TIME/TEMP. REC'D AT LAB:

Page \_\_\_\_\_ of \_\_\_\_\_

## GSL CLIENT #

MICRO #

CHEM. #

SAMPLE REC'D BY:

☐ GSL FIELD SAMPLER/PICK-UP

☐ PICK-UP AT DROP OFF LOCATION

☐ DELIVERED BY CLIENT

Grab	Comp	SAMPLE ID	SAMPLE COLLECTION				ANALYSIS REQUIRED (Print Legibly)		CONTAINER INFORMATION			
			Date	Time	AM	PM	<input type="checkbox"/> List attached	Total Pages	No.	Type*	Size	Pres.*
X	X	150709071-01 VOA	7/9/25	9:25			<input type="checkbox"/>		3	V	40mL	A
X		250709059-10 Trip blank					<input type="checkbox"/>		2	V	40mL	A
X		250709104-01 VOA	7/9/25	10:42			<input type="checkbox"/>		3	V	40mL	A
X		250709259-11 Trip blank					<input type="checkbox"/>		2	V	40mL	A

\*Container type: P = Plastic G = Glass A = Amber Glass I = Sterile Thio V = Vial Other/Specify: \_\_\_\_\_  
\*Preservation Code: A = Non Preserved B = Sulfuric Acid C = Sodium Hydroxide D = Nitric Acid  
E = Hydrochloric Acid F = Zinc Acetate G = Sodium Ithiosulfate H = Ascorbic Acid I = Cooled Other/Specify: \_\_\_\_\_

☒ SUBCONTRACTED WORK

TURNAROUND TIME: ☒ Standard ☐ Rush (IF RUSH REQUESTED) Rush Due by:

SEND TO: Chem Tech

REPORT FORMAT: ☒ Standard Report ☐ Other/Specify:

DATE/TIME:

☐ Standard Report + E2 PWS ID#:

METHOD OF SHIPMENT:

## PAYMENT INFORMATION

☐ Sampling/Pick-up Fee: \$

☐ Composite Fee: \$

☐ Rush Fee: \$

Amount Due: \$

Payment Method: ☐ Credit Card Type:

☐ Check #

Other: See Quote

Note:

VOA UNPRESERVED DUE TO EFFERVESCENCE - 3 DAY TAT PER JORDAN HEDVAT

**SAMPLE CUSTODY EXCHANGES MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION**

**PLEASE PRINT YOUR NAME LEGIBLY, USE FULL LEGAL SIGNATURE, DATE AND TIME**

Sampled by (PRINT):

Signature:

Date/Time:

Client/Client's Representative (PRINT):

Signature:

Date/Time:

1. Received/Relinquished by (PRINT): Megan Howardich

Signature:

Date/Time: 7/9/25 17:00

2. Received/Relinquished by (PRINT): NAGH JACKSON

Signature:

Date/Time: 7/10/25 9:14 AM

Jahmir Davis

The liability of Garden State Laboratories, Inc. for services rendered shall in no event exceed the amount of the invoice.  
Main Lab certified by NJ Dept. of Health, NJDEP-TNI, NY Dept. of Health #1580 and PADEP #68-03680

7-10-25 0914

3.4<sup>c</sup>

CHAIN OF CUSTODY RECORD - PRESS HARD AND PRINT CLEARLY - USE BALL POINT PEN  
IMPORTANT: PRINTED NAMES & SIGNATURES ARE REQUIRED

2 2554

### Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

## LOGIN REPORT/SAMPLE TRANSFER


<b>Order ID :</b> Q2554	<b>GARD04</b>	<b>Order Date :</b> 7/10/2025 9:19:00 AM	<b>Project Mgr :</b>
<b>Client Name :</b> Garden State Laboratories, I		<b>Project Name :</b> Waste Water 2025	<b>Report Type :</b> Level 1
<b>Client Contact :</b> Sharon Ercoliani		<b>Receive DateTime :</b> 7/10/2025 9:14:00 AM	<b>EDD Type :</b> EXCEL NOCLEANUP
<b>Invoice Name :</b> Garden State Laboratories, I		<b>Purchase Order :</b>	<b>Hard Copy Date :</b>
<b>Invoice Contact :</b> Sharon Ercoliani			<b>Date Signoff :</b>


LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2554-01	250709071-01 VOA	Water	07/09/2025	09:25					
					VOCMS Group1		624.1	10 Bus. Days	
					VOCMS Group2		8260-Low	10 Bus. Days	
Q2554-02	250709059-10 Trip blank	Water	07/09/2025	09:25					
					VOCMS Group1		624.1	10 Bus. Days	
					VOCMS Group2		8260-Low	10 Bus. Days	
Q2554-03	250709104-01 VOA	Water	07/09/2025	10:42					
					VOCMS Group1		624.1	10 Bus. Days	
					VOCMS Group2		8260-Low	10 Bus. Days	
Q2554-04	250709059-11 Trip blank	Water	07/09/2025	10:42					
					VOCMS Group1		624.1	10 Bus. Days	
					VOCMS Group2		8260-Low	10 Bus. Days	

## LOGIN REPORT/SAMPLE TRANSFER

<b>Order ID :</b> Q2554	<b>GARD04</b>	<b>Order Date :</b> 7/10/2025 9:19:00 AM	<b>Project Mgr :</b>
<b>Client Name :</b> Garden State Laboratories, I		<b>Project Name :</b> Waste Water 2025	<b>Report Type :</b> Level 1
<b>Client Contact :</b> Sharon Ercoliani		<b>Receive DateTime :</b> 7/10/2025 9:14:00 AM	<b>EDD Type :</b> EXCEL NOCLEANUP
<b>Invoice Name :</b> Garden State Laboratories, I		<b>Purchase Order :</b>	<b>Hard Copy Date :</b>
<b>Invoice Contact :</b> Sharon Ercoliani			<b>Date Signoff :</b>

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
--------	-----------	--------	----------------	----------------	------	------------	--------	----------	--------------

**Relinquished By :**   
**Date / Time :** 7-10-25 1000

**Received By :**   
**Date / Time :** 7/10/25 1000

**Storage Area :** VOA Refridgerator Room