

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>GARD04</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2554</u>
Instrument ID:	<u>MSVOA_N</u>	Calibration Date/Time:	<u>07/10/2025 10:41</u>
Lab File ID:	<u>VN087316.D</u>	Init. Calib. Date(s):	<u>06/25/2025 06/25/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>08:59 10:24</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Acrolein	0.438	0.481		9.82	
Acrylonitrile	1.330	0.933		-29.85	
1,2-Dichloroethane-d4	2.260	2.112	0.01	-6.55	
Toluene-d8	1.347	1.289	0.01	-4.31	
4-Bromofluorobenzene	0.463	0.467	0.2	0.86	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.