

Report of Analysis

Client:	Garden State Laboratories, Inc.		Date Collected:	07/09/25	
Project:	Waste Water 2025		Date Received:	07/10/25	
Client Sample ID:	250709059-10 Trip blank		SDG No.:	Q2554	
Lab Sample ID:	Q2554-02		Matrix:	Water	
Analytical Method:	E624.1		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087320.D	1	07/10/25 13:06	VN071025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
107-02-8	Acrolein	6.60	U	6.60	25.0	ug/L
107-13-1	Acrylonitrile	2.80	U	2.80	25.0	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	29.4		91 - 110	98%	SPK: 30
2037-26-5	Toluene-d8	29.6		91 - 112	99%	SPK: 30
460-00-4	4-Bromofluorobenzene	27.7		63 - 112	92%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	37000	7.824			
540-36-3	1,4-Difluorobenzene	183000	9.106			
3114-55-4	Chlorobenzene-d5	168000	11.865			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products