

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

|                                                 |                                                          |
|-------------------------------------------------|----------------------------------------------------------|
| Lab Name: <u>Alliance</u>                       | Contract: <u>GARD04</u>                                  |
| Lab Code: <u>ACE</u>                            | SDG No.: <u>Q2554</u>                                    |
| Instrument ID: <u>MSVOA_X</u>                   | Calibration Date(s): <u>07/02/2025</u> <u>07/02/2025</u> |
| Heated Purge: (Y/N) <u>N</u>                    | Calibration Time(s): <u>12:11</u> <u>14:31</u>           |
| GC Column: <u>DB-624UI</u> ID: <u>0.18</u> (mm) |                                                          |

|                                |        |                     |        |                     |        |                     |       |       |
|--------------------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|-------|
| LAB FILE ID:                   |        | RRF001 = VX046860.D |        | RRF005 = VX046861.D |        | RRF020 = VX046862.D |       |       |
|                                |        | RRF050 = VX046863.D |        | RRF100 = VX046864.D |        | RRF150 = VX046865.D |       |       |
| COMPOUND                       | RRF001 | RRF005              | RRF020 | RRF050              | RRF100 | RRF150              | RRF   | % RSD |
| Dichlorodifluoromethane        | 0.445  | 0.433               | 0.512  | 0.487               | 0.503  | 0.490               | 0.478 | 6.7   |
| Chloromethane                  | 0.505  | 0.507               | 0.541  | 0.508               | 0.543  | 0.530               | 0.522 | 3.4   |
| Vinyl Chloride                 | 0.568  | 0.549               | 0.591  | 0.550               | 0.589  | 0.566               | 0.569 | 3.2   |
| Bromomethane                   |        | 0.402               | 0.402  | 0.360               | 0.369  | 0.295               | 0.366 | 12    |
| Chloroethane                   | 0.443  | 0.365               | 0.383  | 0.347               | 0.362  | 0.354               | 0.376 | 9.4   |
| Trichlorofluoromethane         | 0.847  | 0.874               | 0.921  | 0.857               | 0.893  | 0.885               | 0.880 | 3     |
| 1,1,2-Trichlorotrifluoroethane | 0.527  | 0.571               | 0.583  | 0.548               | 0.572  | 0.562               | 0.561 | 3.6   |
| 1,1-Dichloroethene             | 0.543  | 0.555               | 0.546  | 0.524               | 0.546  | 0.540               | 0.542 | 1.9   |
| Acetone                        | 0.236  | 0.210               | 0.195  | 0.193               | 0.197  | 0.200               | 0.205 | 7.9   |
| Carbon Disulfide               | 1.585  | 1.425               | 1.405  | 1.342               | 1.398  | 1.376               | 1.422 | 6     |
| Methyl tert-butyl Ether        | 1.447  | 1.470               | 1.540  | 1.528               | 1.587  | 1.615               | 1.531 | 4.2   |
| Methyl Acetate                 | 0.488  | 0.474               | 0.560  | 0.553               | 0.570  | 0.599               | 0.541 | 9.1   |
| Methylene Chloride             | 0.623  | 0.625               | 0.624  | 0.585               | 0.596  | 0.588               | 0.607 | 3.1   |
| trans-1,2-Dichloroethene       | 0.552  | 0.564               | 0.575  | 0.541               | 0.551  | 0.545               | 0.555 | 2.3   |
| 1,1-Dichloroethane             | 1.080  | 1.084               | 1.074  | 1.039               | 1.055  | 1.050               | 1.064 | 1.7   |
| Cyclohexane                    |        | 0.973               | 0.977  | 0.918               | 0.926  | 0.927               | 0.944 | 3     |
| 2-Butanone                     | 0.260  | 0.270               | 0.275  | 0.275               | 0.276  | 0.286               | 0.274 | 3.1   |
| Carbon Tetrachloride           | 0.471  | 0.505               | 0.495  | 0.476               | 0.480  | 0.478               | 0.484 | 2.7   |
| cis-1,2-Dichloroethene         | 0.711  | 0.669               | 0.688  | 0.660               | 0.670  | 0.666               | 0.677 | 2.8   |
| Bromochloromethane             | 0.529  | 0.543               | 0.524  | 0.528               | 0.516  | 0.508               | 0.525 | 2.3   |
| Chloroform                     | 1.150  | 1.106               | 1.112  | 1.048               | 1.052  | 1.050               | 1.087 | 3.9   |
| 1,1,1-Trichloroethane          | 0.944  | 0.897               | 0.908  | 0.886               | 0.901  | 0.910               | 0.908 | 2.2   |
| Methylcyclohexane              | 0.539  | 0.588               | 0.589  | 0.567               | 0.578  | 0.576               | 0.573 | 3.2   |
| Benzene                        | 1.391  | 1.445               | 1.452  | 1.356               | 1.342  | 1.324               | 1.385 | 3.9   |
| 1,2-Dichloroethane             | 0.499  | 0.477               | 0.487  | 0.460               | 0.453  | 0.446               | 0.470 | 4.4   |
| Trichloroethene                | 0.386  | 0.377               | 0.365  | 0.344               | 0.347  | 0.344               | 0.361 | 5     |
| 1,2-Dichloropropane            | 0.364  | 0.340               | 0.362  | 0.346               | 0.345  | 0.342               | 0.350 | 3     |
| Bromodichloromethane           | 0.538  | 0.508               | 0.535  | 0.510               | 0.508  | 0.507               | 0.518 | 2.9   |
| 4-Methyl-2-Pentanone           | 0.330  | 0.340               | 0.368  | 0.366               | 0.353  | 0.357               | 0.353 | 4.3   |
| Toluene                        | 0.855  | 0.894               | 0.906  | 0.841               | 0.831  | 0.823               | 0.858 | 3.9   |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

|                                                 |                                                          |
|-------------------------------------------------|----------------------------------------------------------|
| Lab Name: <u>Alliance</u>                       | Contract: <u>GARD04</u>                                  |
| Lab Code: <u>ACE</u>                            | SDG No.: <u>Q2554</u>                                    |
| Instrument ID: <u>MSVOA_X</u>                   | Calibration Date(s): <u>07/02/2025</u> <u>07/02/2025</u> |
| Heated Purge: (Y/N) <u>N</u>                    | Calibration Time(s): <u>12:11</u> <u>14:31</u>           |
| GC Column: <u>DB-624UI</u> ID: <u>0.18</u> (mm) |                                                          |

|                             |        |                     |        |                     |        |                     |       |       |
|-----------------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|-------|
| LAB FILE ID:                |        | RRF001 = VX046860.D |        | RRF005 = VX046861.D |        | RRF020 = VX046862.D |       |       |
|                             |        | RRF050 = VX046863.D |        | RRF100 = VX046864.D |        | RRF150 = VX046865.D |       |       |
| COMPOUND                    | RRF001 | RRF005              | RRF020 | RRF050              | RRF100 | RRF150              | RRF   | % RSD |
| t-1,3-Dichloropropene       | 0.408  | 0.431               | 0.467  | 0.472               | 0.490  | 0.504               | 0.462 | 7.8   |
| cis-1,3-Dichloropropene     | 0.496  | 0.499               | 0.532  | 0.533               | 0.549  | 0.554               | 0.527 | 4.7   |
| 1,1,2-Trichloroethane       | 0.325  | 0.329               | 0.328  | 0.318               | 0.310  | 0.308               | 0.320 | 2.8   |
| 2-Hexanone                  | 0.205  | 0.227               | 0.245  | 0.247               | 0.238  | 0.242               | 0.234 | 6.8   |
| Dibromochloromethane        | 0.383  | 0.390               | 0.393  | 0.377               | 0.376  | 0.377               | 0.383 | 1.9   |
| 1,2-Dibromoethane           | 0.323  | 0.317               | 0.332  | 0.320               | 0.317  | 0.317               | 0.321 | 1.8   |
| Tetrachloroethene           | 0.353  | 0.346               | 0.345  | 0.315               | 0.320  | 0.318               | 0.333 | 5.1   |
| Chlorobenzene               | 1.111  | 1.117               | 1.113  | 1.049               | 1.051  | 1.045               | 1.081 | 3.3   |
| Ethyl Benzene               | 1.810  | 1.868               | 1.929  | 1.830               | 1.846  | 1.824               | 1.851 | 2.3   |
| m/p-Xylenes                 | 0.702  | 0.710               | 0.731  | 0.688               | 0.683  | 0.672               | 0.698 | 3.1   |
| o-Xylene                    | 0.676  | 0.677               | 0.703  | 0.665               | 0.663  | 0.660               | 0.674 | 2.4   |
| Styrene                     | 1.070  | 1.152               | 1.228  | 1.175               | 1.157  | 1.134               | 1.153 | 4.5   |
| Bromoform                   | 0.269  | 0.261               | 0.273  | 0.269               | 0.275  | 0.275               | 0.270 | 1.9   |
| Isopropylbenzene            | 3.267  | 3.428               | 3.611  | 3.601               | 3.592  | 3.590               | 3.515 | 4     |
| 1,1,2,2-Tetrachloroethane   | 1.000  | 0.948               | 0.981  | 0.963               | 0.939  | 0.968               | 0.966 | 2.3   |
| 1,3-Dichlorobenzene         | 1.615  | 1.664               | 1.673  | 1.627               | 1.625  | 1.619               | 1.637 | 1.5   |
| 1,4-Dichlorobenzene         | 1.852  | 1.776               | 1.711  | 1.632               | 1.622  | 1.632               | 1.704 | 5.5   |
| 1,2-Dichlorobenzene         | 1.626  | 1.592               | 1.625  | 1.568               | 1.560  | 1.550               | 1.587 | 2.1   |
| 1,2-Dibromo-3-Chloropropane | 0.158  | 0.151               | 0.164  | 0.171               | 0.178  | 0.192               | 0.169 | 8.7   |
| 1,2,4-Trichlorobenzene      | 1.016  | 1.020               | 1.078  | 1.087               | 1.121  | 1.145               | 1.078 | 4.9   |
| 1,2,3-Trichlorobenzene      | 0.900  | 0.965               | 1.033  | 1.045               | 1.069  | 1.099               | 1.019 | 7.2   |
| 1,2-Dichloroethane-d4       |        | 0.722               | 0.652  | 0.647               | 0.641  | 0.640               | 0.661 | 5.3   |
| Dibromofluoromethane        |        | 0.355               | 0.346  | 0.339               | 0.332  | 0.325               | 0.339 | 3.5   |
| Toluene-d8                  |        | 1.278               | 1.220  | 1.195               | 1.160  | 1.135               | 1.197 | 4.6   |
| 4-Bromofluorobenzene        |        | 0.493               | 0.469  | 0.455               | 0.433  | 0.426               | 0.455 | 6     |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.