

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: GARD04

Lab Code: ACE

SDG NO.: Q2554

Lab File ID: VX046932.D

BFB Injection Date: 07/10/2025

Instrument ID: MSVOA_X

BFB Injection Time: 08:43

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.2
75	30.0 - 60.0% of mass 95	52.3
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.7 (1) 1
174	50.0 - 100.0% of mass 95	75.5
175	5.0 - 9.0% of mass 174	5.7 (7.6) 1
176	95.0 - 101.0% of mass 174	71.9 (95.2) 1
177	5.0 - 9.0% of mass 176	5.1 (7.1) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX046933.D	07/10/2025	09:54
VX0710WBL01	VX0710WBL01	VX046935.D	07/10/2025	11:44
VX0710WBS01	VX0710WBS01	VX046936.D	07/10/2025	12:10
VX0710WBSD01	VX0710WBSD01	VX046937.D	07/10/2025	12:35
250709059-10 Trip blank	Q2554-02	VX046941.D	07/10/2025	14:00
250709071-01 VOA	Q2554-01	VX046951.D	07/10/2025	17:34