

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GARD04
 Lab Code: ACE SDG NO.: Q2554
 Lab File ID: VX046933.D Date Analyzed: 07/10/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:54
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	231018	12.018				
UPPER LIMIT	462036	12.518				
LOWER LIMIT	115509	11.518				
EPA SAMPLE NO.						
250709104-01 VOA	377959	12.02				
250709059-11 Trip blank	279121	12.02				
VX0710WBL01	271904	12.02				
VX0710WBS01	270401	12.02				
VX0710WBSD01	249133	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.