

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2555
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/14/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031022.D **Time Analyzed:** 16:34

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.646	9.547	9.747	783.010	712.500	9.9
2,4,5-TP(Silvex)	9.351	9.253	9.453	715.530	712.500	0.4
2,4-D	8.466	8.367	8.567	745.710	705.000	5.8
2,4-DB	10.223	10.125	10.325	830.760	712.500	16.6
2,4-DCAA	7.332	7.233	7.433	695.670	750.000	-7.2
Dalapon	2.695	2.598	2.798	613.180	682.500	-10.2
DICAMBA	7.521	7.422	7.622	650.910	705.000	-7.7
DICHLORPROP	8.234	8.135	8.335	655.790	705.000	-7.0
Dinoseb	11.440	11.339	11.539	724.700	705.000	2.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2555
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/14/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031022.D **Time Analyzed:** 16:34

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.356	10.258	10.458	620.360	712.500	-12.9
2,4,5-TP(Silvex)	9.929	9.832	10.032	618.290	712.500	-13.2
2,4-D	9.024	8.927	9.127	599.860	705.000	-14.9
2,4-DB	10.924	10.826	11.026	610.180	712.500	-14.4
2,4-DCAA	7.766	7.669	7.869	627.630	750.000	-16.3
Dalapon	2.703	2.608	2.808	580.910	682.500	-14.9
DICAMBA	7.968	7.871	8.071	603.010	705.000	-14.5
DICHLORPROP	8.688	8.591	8.791	597.240	705.000	-15.3
Dinoseb	11.308	11.210	11.410	608.040	705.000	-13.8