

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2555
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/14/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031028.D **Time Analyzed:** 21:09

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.644	9.547	9.747	779.770	712.500	9.4
2,4,5-TP(Silvex)	9.350	9.253	9.453	713.000	712.500	0.1
2,4-D	8.465	8.367	8.567	718.070	705.000	1.9
2,4-DB	10.221	10.125	10.325	828.510	712.500	16.3
2,4-DCAA	7.331	7.233	7.433	615.310	750.000	-18.0
Dalapon	2.695	2.598	2.798	592.860	682.500	-13.1
DICAMBA	7.520	7.422	7.622	637.220	705.000	-9.6
DICHLORPROP	8.232	8.135	8.335	653.380	705.000	-7.3
Dinoseb	11.436	11.339	11.539	719.910	705.000	2.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2555
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/14/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031028.D **Time Analyzed:** 21:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.358	10.258	10.458	625.920	712.500	-12.2
2,4,5-TP(Silvex)	9.932	9.832	10.032	624.030	712.500	-12.4
2,4-D	9.027	8.927	9.127	604.190	705.000	-14.3
2,4-DB	10.926	10.826	11.026	615.440	712.500	-13.6
2,4-DCAA	7.768	7.669	7.869	636.340	750.000	-15.2
Dalapon	2.706	2.608	2.808	581.140	682.500	-14.9
DICAMBA	7.971	7.871	8.071	608.800	705.000	-13.6
DICHLORPROP	8.691	8.591	8.791	602.040	705.000	-14.6
Dinoseb	11.311	11.210	11.410	610.300	705.000	-13.4