

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2555
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/18/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031144.D **Time Analyzed:** 21:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.636	9.547	9.747	850.050	712.500	19.3
2,4,5-TP(Silvex)	9.342	9.253	9.453	804.740	712.500	12.9
2,4-D	8.456	8.367	8.567	830.710	705.000	17.8
2,4-DB	10.212	10.125	10.325	846.420	712.500	18.8
2,4-DCAA	7.324	7.233	7.433	726.770	750.000	-3.1
Dalapon	2.695	2.598	2.798	587.500	682.500	-13.9
DICAMBA	7.514	7.422	7.622	633.820	705.000	-10.1
DICHLORPROP	8.225	8.135	8.335	765.780	705.000	8.6
Dinoseb	11.427	11.339	11.539	785.940	705.000	11.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2555
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/18/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031144.D **Time Analyzed:** 21:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.355	10.258	10.458	657.810	712.500	-7.7
2,4,5-TP(Silvex)	9.928	9.832	10.032	675.600	712.500	-5.2
2,4-D	9.023	8.927	9.127	699.880	705.000	-0.7
2,4-DB	10.922	10.826	11.026	628.760	712.500	-11.8
2,4-DCAA	7.766	7.669	7.869	674.350	750.000	-10.1
Dalapon	2.703	2.608	2.808	597.360	682.500	-12.5
DICAMBA	7.968	7.871	8.071	640.590	705.000	-9.1
DICHLORPROP	8.687	8.591	8.791	675.920	705.000	-4.1
Dinoseb	11.307	11.210	11.410	635.440	705.000	-9.9