

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2555
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/19/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031154.D **Time Analyzed:** 01:23

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.634	9.547	9.747	850.700	712.500	19.4
2,4,5-TP(Silvex)	9.341	9.253	9.453	799.060	712.500	12.1
2,4-D	8.456	8.367	8.567	832.130	705.000	18.0
2,4-DB	10.212	10.125	10.325	844.050	712.500	18.5
2,4-DCAA	7.324	7.233	7.433	724.370	750.000	-3.4
Dalapon	2.690	2.598	2.798	582.240	682.500	-14.7
DICAMBA	7.513	7.422	7.622	632.640	705.000	-10.3
DICHLORPROP	8.225	8.135	8.335	762.560	705.000	8.2
Dinoseb	11.425	11.339	11.539	764.380	705.000	8.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2555
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/19/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031154.D **Time Analyzed:** 01:23

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.355	10.258	10.458	653.850	712.500	-8.2
2,4,5-TP(Silvex)	9.928	9.832	10.032	673.340	712.500	-5.5
2,4-D	9.023	8.927	9.127	693.610	705.000	-1.6
2,4-DB	10.923	10.826	11.026	629.860	712.500	-11.6
2,4-DCAA	7.765	7.669	7.869	671.450	750.000	-10.5
Dalapon	2.702	2.608	2.808	594.770	682.500	-12.9
DICAMBA	7.967	7.871	8.071	639.590	705.000	-9.3
DICHLORPROP	8.687	8.591	8.791	665.820	705.000	-5.6
Dinoseb	11.306	11.210	11.410	623.520	705.000	-11.6