

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2555
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/22/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031173.D **Time Analyzed:** 10:11

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.634	9.536	9.736	734.800	712.500	3.1
2,4,5-TP(Silvex)	9.339	9.242	9.442	711.570	712.500	-0.1
2,4-D	8.455	8.356	8.556	705.490	705.000	0.1
2,4-DB	10.210	10.113	10.313	750.220	712.500	5.3
2,4-DCAA	7.323	7.225	7.425	738.150	750.000	-1.6
Dalapon	2.689	2.590	2.790	653.950	682.500	-4.2
DICAMBA	7.512	7.414	7.614	697.100	705.000	-1.1
DICHLORPROP	8.223	8.124	8.324	682.330	705.000	-3.2
Dinoseb	11.425	11.327	11.527	698.650	705.000	-0.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2555
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/22/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031173.D **Time Analyzed:** 10:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.353	10.256	10.456	696.510	712.500	-2.2
2,4,5-TP(Silvex)	9.926	9.829	10.029	697.540	712.500	-2.1
2,4-D	9.022	8.924	9.124	677.080	705.000	-4.0
2,4-DB	10.921	10.823	11.023	694.370	712.500	-2.5
2,4-DCAA	7.764	7.666	7.866	720.360	750.000	-4.0
Dalapon	2.703	2.603	2.803	643.740	682.500	-5.7
DICAMBA	7.966	7.868	8.068	690.700	705.000	-2.0
DICHLORPROP	8.685	8.588	8.788	675.290	705.000	-4.2
Dinoseb	11.305	11.208	11.408	684.930	705.000	-2.8