

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2555
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL06 **Date Analyzed:** 07/22/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031177.D **Time Analyzed:** 15:16

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.633	9.536	9.736	764.550	712.500	7.3
2,4,5-TP(Silvex)	9.339	9.242	9.442	731.510	712.500	2.7
2,4-D	8.454	8.356	8.556	738.670	705.000	4.8
2,4-DB	10.209	10.113	10.313	781.920	712.500	9.7
2,4-DCAA	7.323	7.225	7.425	756.860	750.000	0.9
Dalapon	2.687	2.590	2.790	669.820	682.500	-1.9
DICAMBA	7.511	7.414	7.614	717.290	705.000	1.7
DICHLORPROP	8.222	8.124	8.324	701.800	705.000	-0.5
Dinoseb	11.423	11.327	11.527	718.060	705.000	1.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2555
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL06 **Date Analyzed:** 07/22/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031177.D **Time Analyzed:** 15:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.353	10.256	10.456	715.750	712.500	0.5
2,4,5-TP(Silvex)	9.927	9.829	10.029	716.110	712.500	0.5
2,4-D	9.022	8.924	9.124	694.930	705.000	-1.4
2,4-DB	10.921	10.823	11.023	716.620	712.500	0.6
2,4-DCAA	7.764	7.666	7.866	738.240	750.000	-1.6
Dalapon	2.702	2.603	2.803	661.420	682.500	-3.1
DICAMBA	7.966	7.868	8.068	711.020	705.000	0.9
DICHLORPROP	8.685	8.588	8.788	689.190	705.000	-2.2
Dinoseb	11.305	11.208	11.408	704.420	705.000	-0.1