

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2555</u>	Analytical Method:	<u>8151A</u>
Client:	<u>Nobis Group</u>	Datafile :	<u>PS031024.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB168811BS (Column 1)	DICAMBA	166.6	156	ug/Kg	94				38	132	
	Dalapon	166.6	150	ug/Kg	90				70	130	
	DICHLORPROP	166.6	161	ug/Kg	97				28	155	
	2,4-D	166.6	185	ug/Kg	111				28	144	
	2,4,5-TP(Silvex)	166.6	174	ug/Kg	104				43	129	
	2,4,5-T	166.6	185	ug/Kg	111				31	138	
	2,4-DB	166.6	193	ug/Kg	116				34	142	
	Dinoseb	166.6	173	ug/Kg	104				57	152	
	DICAMBA	166.6	143	ug/Kg	86				38	132	
PB168811BS (Column 2)	Dalapon	166.6	140	ug/Kg	84				70	130	
	DICHLORPROP	166.6	143	ug/Kg	86				28	155	
	2,4-D	166.6	145	ug/Kg	87				28	144	
	2,4,5-TP(Silvex)	166.6	149	ug/Kg	89				43	129	
	2,4,5-T	166.6	149	ug/Kg	89				31	138	
	2,4-DB	166.6	146	ug/Kg	88				34	142	
	Dinoseb	166.6	144	ug/Kg	86				57	152	