

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2555

Continuing Calib Date: 07/14/2025

Initial Calibration Date(s): 07/07/2025

07/08/2025

Continuing Calib Time: 09:32

Initial Calibration Time(s): 21:03

04:24

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1016-1       | (1) | 5.64       | 5.64      | 5.54      | 5.74  | 0.00       |
| Aroclor-1016-2       | (2) | 5.66       | 5.66      | 5.56      | 5.76  | 0.00       |
| Aroclor-1016-3       | (3) | 5.72       | 5.72      | 5.62      | 5.82  | 0.00       |
| Aroclor-1016-4       | (4) | 5.82       | 5.82      | 5.72      | 5.92  | 0.00       |
| Aroclor-1016-5       | (5) | 6.11       | 6.11      | 6.01      | 6.21  | 0.00       |
| Aroclor-1260-1       | (1) | 7.23       | 7.23      | 7.13      | 7.33  | 0.00       |
| Aroclor-1260-2       | (2) | 7.48       | 7.48      | 7.38      | 7.58  | 0.00       |
| Aroclor-1260-3       | (3) | 7.84       | 7.84      | 7.74      | 7.94  | 0.00       |
| Aroclor-1260-4       | (4) | 8.06       | 8.06      | 7.96      | 8.16  | 0.00       |
| Aroclor-1260-5       | (5) | 8.38       | 8.38      | 8.28      | 8.48  | 0.00       |
| Tetrachloro-m-xylene |     | 4.49       | 4.49      | 4.39      | 4.59  | 0.00       |
| Decachlorobiphenyl   |     | 10.17      | 10.18     | 10.08     | 10.28 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2555

Continuing Calib Date: 07/14/2025

Initial Calibration Date(s): 07/07/2025

07/08/2025

Continuing Calib Time: 09:32

Initial Calibration Time(s): 21:03

04:24

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.85       | 4.86      | 4.76      | 4.96 | 0.01       |
| Aroclor-1016-2       | (2) | 4.87       | 4.88      | 4.78      | 4.98 | 0.01       |
| Aroclor-1016-3       | (3) | 5.05       | 5.05      | 4.95      | 5.15 | 0.00       |
| Aroclor-1016-4       | (4) | 5.09       | 5.09      | 4.99      | 5.19 | 0.00       |
| Aroclor-1016-5       | (5) | 5.30       | 5.31      | 5.21      | 5.41 | 0.01       |
| Aroclor-1260-1       | (1) | 6.33       | 6.34      | 6.24      | 6.44 | 0.01       |
| Aroclor-1260-2       | (2) | 6.52       | 6.53      | 6.43      | 6.63 | 0.01       |
| Aroclor-1260-3       | (3) | 6.67       | 6.68      | 6.58      | 6.78 | 0.01       |
| Aroclor-1260-4       | (4) | 7.14       | 7.15      | 7.05      | 7.25 | 0.01       |
| Aroclor-1260-5       | (5) | 7.38       | 7.39      | 7.29      | 7.49 | 0.01       |
| Tetrachloro-m-xylene |     | 3.78       | 3.78      | 3.68      | 3.88 | 0.00       |
| Decachlorobiphenyl   |     | 8.77       | 8.78      | 8.68      | 8.88 | 0.01       |