

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2555
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/11/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073726.D **Time Analyzed:** 16:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.639	5.539	5.739	506.790	500.000	1.4
Aroclor-1016-2	5.660	5.560	5.760	509.720	500.000	1.9
Aroclor-1016-3	5.722	5.622	5.822	514.970	500.000	3.0
Aroclor-1016-4	5.819	5.720	5.920	525.790	500.000	5.2
Aroclor-1016-5	6.111	6.012	6.212	540.380	500.000	8.1
Aroclor-1260-1	7.228	7.129	7.329	521.900	500.000	4.4
Aroclor-1260-2	7.482	7.383	7.583	453.170	500.000	-9.4
Aroclor-1260-3	7.840	7.740	7.940	503.620	500.000	0.7
Aroclor-1260-4	8.063	7.964	8.164	511.200	500.000	2.2
Aroclor-1260-5	8.381	8.283	8.483	510.220	500.000	2.0
Decachlorobiphenyl	10.174	10.076	10.276	52.950	50.000	5.9
Tetrachloro-m-xylene	4.488	4.388	4.588	53.000	50.000	6.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2555
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/11/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073726.D **Time Analyzed:** 16:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.857	4.758	4.958	514.730	500.000	2.9
Aroclor-1016-2	4.874	4.775	4.975	511.340	500.000	2.3
Aroclor-1016-3	5.051	4.951	5.151	518.080	500.000	3.6
Aroclor-1016-4	5.093	4.993	5.193	514.910	500.000	3.0
Aroclor-1016-5	5.306	5.207	5.407	589.400	500.000	17.9
Aroclor-1260-1	6.335	6.237	6.437	516.330	500.000	3.3
Aroclor-1260-2	6.524	6.425	6.625	536.800	500.000	7.4
Aroclor-1260-3	6.676	6.577	6.777	502.940	500.000	0.6
Aroclor-1260-4	7.145	7.046	7.246	520.580	500.000	4.1
Aroclor-1260-5	7.388	7.289	7.489	512.150	500.000	2.4
Decachlorobiphenyl	8.778	8.680	8.880	54.240	50.000	8.5
Tetrachloro-m-xylene	3.779	3.678	3.878	51.270	50.000	2.5