

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2555
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/11/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073741.D **Time Analyzed:** 21:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.640	5.539	5.739	496.340	500.000	-0.7
Aroclor-1016-2	5.661	5.560	5.760	513.030	500.000	2.6
Aroclor-1016-3	5.723	5.622	5.822	515.830	500.000	3.2
Aroclor-1016-4	5.821	5.720	5.920	524.450	500.000	4.9
Aroclor-1016-5	6.113	6.012	6.212	538.900	500.000	7.8
Aroclor-1260-1	7.229	7.129	7.329	525.470	500.000	5.1
Aroclor-1260-2	7.483	7.383	7.583	449.140	500.000	-10.2
Aroclor-1260-3	7.841	7.740	7.940	496.710	500.000	-0.7
Aroclor-1260-4	8.064	7.964	8.164	508.030	500.000	1.6
Aroclor-1260-5	8.382	8.283	8.483	505.700	500.000	1.1
Decachlorobiphenyl	10.175	10.076	10.276	52.150	50.000	4.3
Tetrachloro-m-xylene	4.489	4.388	4.588	52.240	50.000	4.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** NOBI03
Lab Code: ACE **SDG NO.:** Q2555
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/11/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073741.D **Time Analyzed:** 21:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.856	4.758	4.958	501.980	500.000	0.4
Aroclor-1016-2	4.874	4.775	4.975	509.230	500.000	1.8
Aroclor-1016-3	5.051	4.951	5.151	513.940	500.000	2.8
Aroclor-1016-4	5.093	4.993	5.193	505.170	500.000	1.0
Aroclor-1016-5	5.306	5.207	5.407	588.380	500.000	17.7
Aroclor-1260-1	6.336	6.237	6.437	510.450	500.000	2.1
Aroclor-1260-2	6.524	6.425	6.625	526.540	500.000	5.3
Aroclor-1260-3	6.675	6.577	6.777	491.100	500.000	-1.8
Aroclor-1260-4	7.144	7.046	7.246	512.680	500.000	2.5
Aroclor-1260-5	7.387	7.289	7.489	507.870	500.000	1.6
Decachlorobiphenyl	8.778	8.680	8.880	57.670	50.000	15.3
Tetrachloro-m-xylene	3.779	3.678	3.878	51.100	50.000	2.2