

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2555

Continuing Calib Date: 07/11/2025

Initial Calibration Date(s): 06/17/2025

06/17/2025

Continuing Calib Time: 18:16

Initial Calibration Time(s): 15:52

16:47

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.07	9.07	8.97	9.17	0.00
Tetrachloro-m-xylene	3.55	3.55	3.45	3.65	0.00
alpha-BHC	4.00	4.00	3.90	4.10	0.00
beta-BHC	4.52	4.52	4.42	4.62	0.01
delta-BHC	4.76	4.76	4.66	4.86	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.93	4.93	4.83	5.03	0.00
Aldrin	5.27	5.27	5.17	5.37	0.00
Heptachlor epoxide	5.69	5.69	5.59	5.79	0.00
Endosulfan I	6.07	6.08	5.98	6.18	0.01
Dieldrin	6.35	6.35	6.25	6.45	0.01
4,4'-DDE	6.19	6.20	6.10	6.30	0.01
Endrin	6.57	6.58	6.48	6.68	0.01
Endosulfan II	6.79	6.79	6.69	6.89	0.01
4,4'-DDD	6.70	6.71	6.61	6.81	0.01
Endosulfan sulfate	7.15	7.15	7.05	7.25	0.00
4,4'-DDT	7.02	7.02	6.92	7.12	0.00
Methoxychlor	7.49	7.49	7.39	7.59	0.00
Endrin ketone	7.63	7.63	7.53	7.73	0.00
Endrin aldehyde	6.91	6.92	6.82	7.02	0.01
alpha-Chlordane	6.03	6.03	5.93	6.13	0.00
gamma-Chlordane	5.95	5.95	5.85	6.05	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: NOBI03

Lab Code: ACE

SDG NO.: Q2555

Continuing Calib Date: 07/11/2025

Initial Calibration Date(s): 06/17/2025

06/17/2025

Continuing Calib Time: 18:16

Initial Calibration Time(s): 15:52

16:47

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.07	8.07	7.97	8.17	0.00
Tetrachloro-m-xylene	2.88	2.88	2.78	2.98	0.00
alpha-BHC	3.39	3.39	3.29	3.49	0.00
beta-BHC	4.03	4.03	3.93	4.13	0.00
delta-BHC	4.26	4.26	4.16	4.36	0.00
gamma-BHC (Lindane)	3.73	3.73	3.63	3.83	0.00
Heptachlor	4.08	4.08	3.98	4.18	0.00
Aldrin	4.37	4.37	4.27	4.47	0.00
Heptachlor epoxide	4.87	4.87	4.77	4.97	0.00
Endosulfan I	5.25	5.25	5.15	5.35	0.01
Dieldrin	5.51	5.51	5.41	5.61	0.00
4,4'-DDE	5.37	5.38	5.28	5.48	0.01
Endrin	5.79	5.79	5.69	5.89	0.00
Endosulfan II	6.08	6.08	5.98	6.18	0.00
4,4'-DDD	5.93	5.93	5.83	6.03	0.00
Endosulfan sulfate	6.48	6.48	6.38	6.58	0.00
4,4'-DDT	6.18	6.18	6.08	6.28	0.00
Methoxychlor	6.75	6.76	6.66	6.86	0.01
Endrin ketone	6.99	6.99	6.89	7.09	0.00
Endrin aldehyde	6.26	6.26	6.16	6.36	0.00
alpha-Chlordane	5.19	5.19	5.09	5.29	0.00
gamma-Chlordane	5.12	5.13	5.03	5.23	0.01