

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD061825

Review By	Abdul	Review On	6/18/2025 8:40:56 AM
Supervise By	mohammad	Supervise On	6/19/2025 2:43:55 AM
SubDirectory	PD061825	HP Acquire Method	HP Processing Method pd061825 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24651
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284
CCC	PP24261,PP24273,PP24279,PP24284
Internal Standard/PEM	
ICV/I.BLK	PP24273,PP24279,PP24284
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD088989.D	17 Jun 2025 14:58		AR\AJ	Ok
2	I.BLK	I.BLK	PD088990.D	17 Jun 2025 15:11		AR\AJ	Ok
3	PEM	PEM	PD088991.D	17 Jun 2025 15:25		AR\AJ	Ok,M
4	RESCHK	RESCHK	PD088992.D	17 Jun 2025 15:39		AR\AJ	Ok
5	PSTDICC100	PSTDICC100	PD088993.D	17 Jun 2025 15:52		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PD088994.D	17 Jun 2025 16:06		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PD088995.D	17 Jun 2025 16:20		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PD088996.D	17 Jun 2025 16:33		AR\AJ	Ok,M
9	PSTDICC005	PSTDICC005	PD088997.D	17 Jun 2025 16:47		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PD088998.D	17 Jun 2025 17:00		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PD088999.D	17 Jun 2025 17:14		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PD089000.D	17 Jun 2025 17:28		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PD089001.D	17 Jun 2025 17:41		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PD089002.D	17 Jun 2025 17:55		AR\AJ	Ok
15	PTOXICC1000	PTOXICC1000	PD089003.D	17 Jun 2025 18:09		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PD089004.D	17 Jun 2025 18:22		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PD089005.D	17 Jun 2025 18:36		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PD089006.D	17 Jun 2025 18:50		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD061825

Review By	Abdul	Review On	6/18/2025 8:40:56 AM
Supervise By	mohammad	Supervise On	6/19/2025 2:43:55 AM
SubDirectory	PD061825	HP Acquire Method	HP Processing Method pd061825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PD089007.D	17 Jun 2025 19:03		AR\AJ	Ok,M
20	PSTDICV050	ICVPD061825	PD089008.D	17 Jun 2025 19:17		AR\AJ	Ok
21	PCHLORICV500	ICVPD061825CHLOR	PD089009.D	17 Jun 2025 19:31		AR\AJ	Ok
22	PTOXICV500	ICVPD061825TOX	PD089010.D	17 Jun 2025 19:44		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD071125

Review By	Abdul	Review On	7/14/2025 8:39:05 AM
Supervise By	mohammad	Supervise On	7/15/2025 1:40:55 AM
SubDirectory	PD071125	HP Acquire Method	HP Processing Method pd061825 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24651
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284
CCC	PP24261,PP24273,PP24279,PP24284
Internal Standard/PEM	
ICV/I.BLK	PP24273,PP24279,PP24284
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD089449.D	11 Jul 2025 09:53		AR\AJ	Ok
2	I.BLK	I.BLK	PD089450.D	11 Jul 2025 10:06		AR\AJ	Ok
3	PEM	PEM	PD089451.D	11 Jul 2025 10:20		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PD089452.D	11 Jul 2025 11:25		AR\AJ	Ok,M
5	PB168810BL	PB168810BL	PD089453.D	11 Jul 2025 13:01		AR\AJ	Ok,M
6	PB168810BS	PB168810BS	PD089454.D	11 Jul 2025 13:15		AR\AJ	Ok,M
7	Q2559-01	500-3B CONCRETE C	PD089455.D	11 Jul 2025 13:28		AR\AJ	Ok
8	Q2560-01	LP-7102025	PD089456.D	11 Jul 2025 13:42		AR\AJ	Ok,M
9	Q2560-01MS	LP-7102025MS	PD089457.D	11 Jul 2025 13:56		AR\AJ	Ok,M
10	Q2560-01MSD	LP-7102025MSD	PD089458.D	11 Jul 2025 14:09		AR\AJ	Ok,M
11	Q2564-01	ARS20-0030	PD089459.D	11 Jul 2025 14:23		AR\AJ	Ok,M
12	Q2564-03	ARS20-0013	PD089460.D	11 Jul 2025 14:36		AR\AJ	Ok,M
13	Q2564-05	ARS20-0039	PD089461.D	11 Jul 2025 14:50	need dilution	AR\AJ	Dilution
14	I.BLK	I.BLK	PD089462.D	11 Jul 2025 15:04		AR\AJ	Ok,M
15	PSTDCCC050	PSTDCCC050	PD089463.D	11 Jul 2025 15:17		AR\AJ	Ok,M
16	Q2565-02	MOO-25-0194-0195	PD089464.D	11 Jul 2025 15:31		AR\AJ	Ok,M
17	Q2565-04	MOO-25-0196	PD089465.D	11 Jul 2025 15:45		AR\AJ	Ok,M
18	Q2565-05	MOO-25-0180	PD089466.D	11 Jul 2025 15:58	TCMX high in 2nd column	AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD071125

Review By	Abdul	Review On	7/14/2025 8:39:05 AM
Supervise By	mohammad	Supervise On	7/15/2025 1:40:55 AM
SubDirectory	PD071125	HP Acquire Method	HP Processing Method pd061825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q2565-03	MOO-25-0191	PD089467.D	11 Jul 2025 16:12		AR\AJ	Ok,M
20	Q2571-01	TP-18	PD089468.D	11 Jul 2025 16:26		AR\AJ	Ok,M
21	Q2571-05	TP-17	PD089469.D	11 Jul 2025 16:40		AR\AJ	Ok,M
22	I.BLK	I.BLK	PD089470.D	11 Jul 2025 16:53		AR\AJ	Ok
23	PSTDCCC050	PSTDCCC050	PD089471.D	11 Jul 2025 18:16		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD071425

Review By	Abdul	Review On	7/16/2025 9:15:27 AM
Supervise By	mohammad	Supervise On	7/22/2025 9:35:12 AM
SubDirectory	PD071425	HP Acquire Method	HP Processing Method pd061825 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24651
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284
CCC	PP24261,PP24273,PP24279,PP24284
Internal Standard/PEM	
ICV/I.BLK	PP24273,PP24279,PP24284
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD089472.D	14 Jul 2025 09:45		AR\AJ	Ok
2	I.BLK	I.BLK	PD089473.D	14 Jul 2025 10:13		AR\AJ	Ok
3	PEM	PEM	PD089474.D	14 Jul 2025 10:57		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PD089475.D	14 Jul 2025 11:11		AR\AJ	Ok
5	PCHLORCCC500	PCHLORCCC500	PD089476.D	14 Jul 2025 12:42		AR\AJ	Ok,M
6	PTOXCCC500	PTOXCCC500	PD089477.D	14 Jul 2025 14:11		AR\AJ	Ok,M
7	PB168810BS	PB168810BS	PD089478.D	14 Jul 2025 14:28	CHLOR BS	AR\AJ	Ok,M
8	PB168810BS	PB168810BS	PD089479.D	14 Jul 2025 16:00	TOX BS	AR\AJ	Ok,M
9	I.BLK	I.BLK	PD089480.D	14 Jul 2025 16:44		AR\AJ	Ok
10	PSTDCCC050	PSTDCCC050	PD089481.D	14 Jul 2025 16:58		AR\AJ	Ok
11	PCHLORCCC500	PCHLORCCC500	PD089482.D	14 Jul 2025 17:39		AR\AJ	Ok,M
12	PTOXCCC500	PTOXCCC500	PD089483.D	14 Jul 2025 17:54		AR\AJ	Ok
13	Q2564-05DL	ARS20-0039DL	PD089484.D	14 Jul 2025 18:43		AR\AJ	Ok,M
14	Q2555-01	OU4-TS-29-070925	PD089485.D	14 Jul 2025 19:01	DCB Low in both column	AR\AJ	ReRun
15	Q2555-03	OU4-TS-30-070925	PD089486.D	14 Jul 2025 19:14	DCB Low in both column	AR\AJ	ReRun
16	Q2558-01	OU4-TS-Denali-070925	PD089487.D	14 Jul 2025 19:28		AR\AJ	Ok,M
17	Q2558-03	OU4-TS-Grillo-OG-070925	PD089488.D	14 Jul 2025 19:42	DCB Low in both column	AR\AJ	ReRun
18	I.BLK	I.BLK	PD089489.D	14 Jul 2025 19:55		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD071425

Review By	Abdul	Review On	7/16/2025 9:15:27 AM
Supervise By	mohammad	Supervise On	7/22/2025 9:35:12 AM
SubDirectory	PD071425	HP Acquire Method	HP Processing Method
			pd061825 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24651
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,P P24278,PP24279,PP24280,PP24281,PP24282,PP24283,PP24284
CCC	PP24261,PP24273,PP24279,PP24284
Internal Standard/PEM	
ICV/I.BLK	PP24273,PP24279,PP24284
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	PSTDCCC050	PSTDCCC050	PD089490.D	14 Jul 2025 20:09		AR\AJ	Ok
20	PCHLORCCC500	PCHLORCCC500	PD089491.D	14 Jul 2025 20:23		AR\AJ	Ok,M
21	PTOXCCC500	PTOXCCC500	PD089492.D	14 Jul 2025 20:36	CCC Fail	AR\AJ	Not Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD072125

Review By	Abdul	Review On	7/22/2025 8:03:16 AM
Supervise By	mohammad	Supervise On	7/23/2025 1:33:20 AM
SubDirectory	PD072125	HP Acquire Method	HP Processing Method pd072125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24744,PP24750,PP24751,PP24752,PP24753,PP24746,PP24755,PP24756,PP24757,PP24758,PP24748,PP24760,PP24761,PP24762,PP24763		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24764		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD089536.D	21 Jul 2025 11:27		AR\AJ	Ok
2	I.BLK	I.BLK	PD089537.D	21 Jul 2025 12:08		AR\AJ	Ok
3	PEM	PEM	PD089538.D	21 Jul 2025 12:22		AR\AJ	Ok,M
4	RESCHK	RESCHK	PD089539.D	21 Jul 2025 12:35		AR\AJ	Ok
5	PSTDICC100	PSTDICC100	PD089540.D	21 Jul 2025 12:49		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PD089541.D	21 Jul 2025 13:03		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PD089542.D	21 Jul 2025 13:16		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PD089543.D	21 Jul 2025 13:30		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PD089544.D	21 Jul 2025 13:44		AR\AJ	Ok
10	PCHLORICC1000	PCHLORICC1000	PD089545.D	21 Jul 2025 13:57		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PD089546.D	21 Jul 2025 14:11		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PD089547.D	21 Jul 2025 14:25		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PD089548.D	21 Jul 2025 14:38		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PD089549.D	21 Jul 2025 14:52		AR\AJ	Ok
15	PTOXICC1000	PTOXICC1000	PD089550.D	21 Jul 2025 15:05		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PD089551.D	21 Jul 2025 15:19		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PD089552.D	21 Jul 2025 15:32		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PD089553.D	21 Jul 2025 15:46		AR\AJ	Ok,M

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD072125

Review By	Abdul	Review On	7/22/2025 8:03:16 AM
Supervise By	mohammad	Supervise On	7/23/2025 1:33:20 AM
SubDirectory	PD072125	HP Acquire Method	HP Processing Method pd072125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24744,PP24750,PP24751,PP24752,PP24753,PP24746,PP24755,PP24756,PP24757,PP24758,PP24748,PP24760,PP24761,PP24762,PP24763		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24764		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PD089554.D	21 Jul 2025 15:59		AR\AJ	Ok
20	PSTDICV050	ICVPD072125	PD089555.D	21 Jul 2025 16:25		AR\AJ	Ok
21	PCHLORICV500	ICVPD072125CHLOR	PD089556.D	21 Jul 2025 16:39		AR\AJ	Ok
22	PTOXICV500	ICVPD072125TOX	PD089557.D	21 Jul 2025 16:52		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD072425

Review By	Abdul	Review On	7/24/2025 9:20:07 AM
Supervise By	mohammad	Supervise On	7/24/2025 9:21:10 AM
SubDirectory	PD072425	HP Acquire Method	HP Processing Method pd072125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24651
Initial Calibration Stds	PP24744,PP24750,PP24751,PP24752,PP24753,PP24746,PP24755,PP24756,PP24757,PP24758,PP24748,PP24760,PP24761,PP24762,PP24763
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24764
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD089590.D	23 Jul 2025 12:22		AR\AJ	Ok
2	I.BLK	I.BLK	PD089591.D	23 Jul 2025 12:35		AR\AJ	Ok
3	PEM	PEM	PD089592.D	23 Jul 2025 12:49		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PD089593.D	23 Jul 2025 13:02		AR\AJ	Ok
5	PCHLORCCC500	PCHLORCCC500	PD089594.D	23 Jul 2025 13:39		AR\AJ	Ok
6	PTOXCCC500	PTOXCCC500	PD089595.D	23 Jul 2025 13:57		AR\AJ	Ok
7	PB168922BL	PB168922BL	PD089596.D	23 Jul 2025 14:18	TYPO PB168928BL, corrected.	AR\AJ	Ok
8	PB168922BS	PB168922BS	PD089597.D	23 Jul 2025 15:19	TYPO PB168928BS, corrected.	AR\AJ	Ok
9	PB168922BS	PB168922BS	PD089598.D	23 Jul 2025 15:32	TYPO PB168928BS, corrected.	AR\AJ	Ok
10	PB168922BS	PB168922BS	PD089599.D	23 Jul 2025 15:46	TYPO PB168928BS, corrected.	AR\AJ	Ok,M
11	I.BLK	I.BLK	PD089600.D	23 Jul 2025 16:00		AR\AJ	Ok
12	PSTDCCC050	PSTDCCC050	PD089601.D	23 Jul 2025 16:13		AR\AJ	Ok
13	PCHLORCCC500	PCHLORCCC500	PD089602.D	23 Jul 2025 16:27		AR\AJ	Ok
14	PTOXCCC500	PTOXCCC500	PD089603.D	23 Jul 2025 16:41		AR\AJ	Ok,M
15	Q2555-01RE	OU4-TS-29-070925RE	PD089604.D	23 Jul 2025 16:54	DCB Low in both column	AR\AJ	Confirms
16	Q2555-03RE	OU4-TS-30-070925RE	PD089605.D	23 Jul 2025 17:08	DCB Low in both column	AR\AJ	Confirms
17	Q2633-02	FIBER-GLASS-TANK	PD089606.D	23 Jul 2025 17:21		AR\AJ	Ok,M
18	PB168972BL	PB168972BL	PD089607.D	23 Jul 2025 17:35		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD072425

Review By	Abdul	Review On	7/24/2025 9:20:07 AM
Supervise By	mohammad	Supervise On	7/24/2025 9:21:10 AM
SubDirectory	PD072425	HP Acquire Method	HP Processing Method pd072125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24744,PP24750,PP24751,PP24752,PP24753,PP24746,PP24755,PP24756,PP24757,PP24758,PP24748,PP24760,PP24761,PP24762,PP24763		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24764		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PB168972BS	PB168972BS	PD089608.D	23 Jul 2025 17:48	some compound recovery fail	AR\AJ	Not Ok
20	PB168972BSD	PB168972BSD	PD089609.D	23 Jul 2025 18:02	some compound recovery fail	AR\AJ	Not Ok
21	I.BLK	I.BLK	PD089610.D	23 Jul 2025 18:16		AR\AJ	Ok
22	PSTDCCC050	PSTDCCC050	PD089611.D	23 Jul 2025 18:29		AR\AJ	Ok
23	Q2639-01	OU4-TS-38-071725	PD089612.D	23 Jul 2025 18:57	DCB Low in 2nd column	AR\AJ	Ok,M
24	Q2639-03	OU4-TS-39-071725	PD089613.D	23 Jul 2025 19:10	DCB Low in 2nd column	AR\AJ	Ok,M
25	Q2639-05	OU4-TS-40-071725	PD089614.D	23 Jul 2025 19:24	DCB Low in 2nd column	AR\AJ	Ok,M
26	Q2639-07	OU4-TS-41-071725	PD089615.D	23 Jul 2025 19:38	DCB Low in 2nd column	AR\AJ	Ok,M
27	Q2639-09	OU4-TS-42-071725	PD089616.D	23 Jul 2025 19:51	DCB Low in both column	AR\AJ	ReRun
28	Q2639-11	OU4-TS-43-071725	PD089617.D	23 Jul 2025 20:05	DCB Low in both column	AR\AJ	ReRun
29	Q2639-13	OU4-TS-44-071725	PD089618.D	23 Jul 2025 20:19	DCB Low in both column	AR\AJ	ReRun
30	I.BLK	I.BLK	PD089619.D	23 Jul 2025 20:46		AR\AJ	Ok
31	PSTDCCC050	PSTDCCC050	PD089620.D	23 Jul 2025 22:08		AR\AJ	Ok

M : Manual Integration