

Surrogate Summary

SDG No.: **Q2555**

Client: **Nobis Group**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD088990.D	PIBLK-PD088990.D	Decachlorobiphen	1	20	18.2	91		30	135
		Tetrachloro-m-xyl	1	20	16.1	80		44	124
		Decachlorobiphen	2	20	18.6	93		30	135
		Tetrachloro-m-xyl	2	20	17.7	88		44	124
I.BLK-PD089450.D	PIBLK-PD089450.D	Decachlorobiphen	1	20	20.2	101		30	135
		Tetrachloro-m-xyl	1	20	19.5	98		44	124
		Decachlorobiphen	2	20	20.1	101		30	135
		Tetrachloro-m-xyl	2	20	23.3	117		44	124
PB168810BL	PB168810BL	Decachlorobiphen	1	20	16.9	84		55	130
		Tetrachloro-m-xyl	1	20	16.5	83		42	129
		Decachlorobiphen	2	20	17.3	86		55	130
		Tetrachloro-m-xyl	2	20	19.0	95		42	129
PB168810BS	PB168810BS	Decachlorobiphen	1	20	16.5	83		55	130
		Tetrachloro-m-xyl	1	20	15.7	79		42	129
		Decachlorobiphen	2	20	16.9	84		55	130
		Tetrachloro-m-xyl	2	20	19.6	98		42	129
Q2560-01MS	LP-7102025MS	Decachlorobiphen	1	20	17.5	88		55	130
		Tetrachloro-m-xyl	1	20	15.8	79		42	129
		Decachlorobiphen	2	20	16.4	82		55	130
		Tetrachloro-m-xyl	2	20	20.1	101		42	129
Q2560-01MSD	LP-7102025MSD	Decachlorobiphen	1	20	17.0	85		55	130
		Tetrachloro-m-xyl	1	20	15.1	76		42	129
		Decachlorobiphen	2	20	16.3	81		55	130
		Tetrachloro-m-xyl	2	20	20.4	102		42	129
I.BLK-PD089462.D	PIBLK-PD089462.D	Decachlorobiphen	1	20	17.9	89		30	135
		Tetrachloro-m-xyl	1	20	19.4	97		44	124
		Decachlorobiphen	2	20	17.6	88		30	135
		Tetrachloro-m-xyl	2	20	22.6	113		44	124
I.BLK-PD089480.D	PIBLK-PD089480.D	Decachlorobiphen	1	20	21.2	106		30	135
		Tetrachloro-m-xyl	1	20	20.6	103		44	124
		Decachlorobiphen	2	20	23.3	117		30	135
		Tetrachloro-m-xyl	2	20	23.6	118		44	124
Q2555-01	OU4-TS-29-070925	Decachlorobiphen	1	20	7.37	37	*	55	130
		Tetrachloro-m-xyl	1	20	13.5	67		42	129
		Decachlorobiphen	2	20	8.29	41	*	55	130
		Tetrachloro-m-xyl	2	20	11.9	60		42	129
Q2555-03	OU4-TS-30-070925	Decachlorobiphen	1	20	6.86	34	*	55	130
		Tetrachloro-m-xyl	1	20	8.98	45		42	129
		Decachlorobiphen	2	20	7.60	38	*	55	130
		Tetrachloro-m-xyl	2	20	10.1	50		42	129
I.BLK-PD089489.D	PIBLK-PD089489.D	Decachlorobiphen	1	20	18.5	93		30	135

Surrogate Summary

SDG No.: Q2555

Client: Nobis Group

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD089489.D	PIBLK-PD089489.D	Tetrachloro-m-xyl	1	20	20.1	100		44	124
		Decachlorobiphen	2	20	20.6	103		30	135
I.BLK-PD089537.D	PIBLK-PD089537.D	Tetrachloro-m-xyl	2	20	23.4	117		44	124
		Decachlorobiphen	1	20	21.2	106		30	135
		Tetrachloro-m-xyl	1	20	18.5	92		44	124
		Decachlorobiphen	2	20	20.5	103		30	135
I.BLK-PD089600.D	PIBLK-PD089600.D	Tetrachloro-m-xyl	2	20	19.9	100		44	124
		Decachlorobiphen	1	20	23.1	115		30	135
		Tetrachloro-m-xyl	1	20	20.2	101		44	124
		Decachlorobiphen	2	20	22.4	112		30	135
Q2555-01RE	OU4-TS-29-070925RE	Tetrachloro-m-xyl	2	20	21.7	109		44	124
		Decachlorobiphen	1	20	8.01	40	*	55	130
		Tetrachloro-m-xyl	1	20	12.9	64		42	129
		Decachlorobiphen	2	20	7.93	40	*	55	130
Q2555-03RE	OU4-TS-30-070925RE	Tetrachloro-m-xyl	2	20	10.1	50		42	129
		Decachlorobiphen	1	20	7.16	36	*	55	130
		Tetrachloro-m-xyl	1	20	8.63	43		42	129
		Decachlorobiphen	2	20	7.05	35	*	55	130
I.BLK-PD089610.D	PIBLK-PD089610.D	Tetrachloro-m-xyl	2	20	8.80	44		42	129
		Decachlorobiphen	1	20	23.1	116		30	135
		Tetrachloro-m-xyl	1	20	20.2	101		44	124
		Decachlorobiphen	2	20	22.5	112		30	135
		Tetrachloro-m-xyl	2	20	21.9	110		44	124