

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF071525

Review By	Rahul	Review On	7/16/2025 11:47:09 AM
Supervise By	Jagrut	Supervise On	7/16/2025 1:41:28 PM
SubDirectory	BF071525	HP Acquire Method	BNA_F
		HP Processing Method	Bf071525

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12674,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143092.D	15 Jul 2025 12:35		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF143093.D	15 Jul 2025 13:04		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF143094.D	15 Jul 2025 13:35		RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BF143095.D	15 Jul 2025 14:05		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BF143096.D	15 Jul 2025 14:36	Compound#26,32,48,51,57,80,84,85 Kept on LR	RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BF143097.D	15 Jul 2025 15:05	Compound#54,65 Kept on QR	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BF143098.D	15 Jul 2025 15:35		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BF143099.D	15 Jul 2025 16:05		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BF143100.D	15 Jul 2025 16:35		RC/JU	Ok,M
10	SSTDICV040	ICVBF071525	BF143101.D	15 Jul 2025 17:27		RC/JU	Ok,M
11	PB168737BL	PB168737BL	BF143102.D	15 Jul 2025 17:57		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF071625

Review By	Rahul	Review On	7/17/2025 11:19:37 AM
Supervise By	Jagrut	Supervise On	7/17/2025 11:19:59 AM
SubDirectory	BF071625	HP Acquire Method	BNA_F
		HP Processing Method	Bf071525

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12674,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143103.D	16 Jul 2025 09:06		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF143104.D	16 Jul 2025 10:04		RC/JU	Ok
3	PB168787BL	PB168787BL	BF143105.D	16 Jul 2025 10:34		RC/JU	Ok
4	Q2126-02	LOQ-SOIL-02-QT2-202	BF143106.D	16 Jul 2025 11:03	LOQ-SOIL-5 PPM	RC/JU	Ok
5	Q2126-02	LOQ-SOIL-02-QT2-202	BF143107.D	16 Jul 2025 11:37	LOQ-SOIL-10 PPM	RC/JU	Ok
6	Q2126-08	LOQ-WATER-02-QT2-2	BF143108.D	16 Jul 2025 12:07	LOQ-WATER-5 PPM	RC/JU	Ok
7	Q2126-08	LOQ-WATER-02-QT2-2	BF143109.D	16 Jul 2025 12:36	LOQ-WATER-10 PPM	RC/JU	Ok
8	Q2126-01	LOD-MDL-SOIL-03-QT	BF143110.D	16 Jul 2025 13:05	LOD-MDL-SOIL-4 PPM	RC/JU	Ok
9	Q2126-01	LOD-MDL-SOIL-03-QT	BF143111.D	16 Jul 2025 13:35	LOD-MDL-SOIL-8 PPM	RC/JU	Ok
10	Q2126-07	LOD-MDL-WATER-01-0	BF143112.D	16 Jul 2025 14:05	LOD-MDL-WATER-4 PPM	RC/JU	Ok
11	Q2126-07	LOD-MDL-WATER-01-0	BF143113.D	16 Jul 2025 14:35	LOD-MDL-WATER-8 PPM	RC/JU	Ok
12	PB168802BL	PB168802BL	BF143114.D	16 Jul 2025 15:13		RC/JU	Ok
13	SSTDCCC040	SSTDCCC040EC	BF143115.D	16 Jul 2025 15:45		RC/JU	Ok,M
14	DFTPP	DFTPP	BF143116.D	16 Jul 2025 19:09		RC/JU	Ok
15	SSTDCCC040	SSTDCCC040	BF143117.D	16 Jul 2025 19:39		RC/JU	Ok
16	PB168854BL	PB168854BL	BF143118.D	16 Jul 2025 20:09		RC/JU	Ok
17	PB168854BS	PB168854BS	BF143119.D	16 Jul 2025 20:39		RC/JU	Ok,M
18	PB168847TB	PB168847TB	BF143120.D	16 Jul 2025 21:09		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF071625

Review By	Rahul	Review On	7/17/2025 11:19:37 AM		
Supervise By	Jagrut	Supervise On	7/17/2025 11:19:59 AM		
SubDirectory	BF071625	HP Acquire Method	BNA_F	HP Processing Method	Bf071525
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840				
CCC	SP6836				
Internal Standard/PEM	S12674,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2592-02	WC-SOIL-20250711	BF143121.D	16 Jul 2025 21:39		RC/JU	Ok
20	Q2592-02MS	WC-SOIL-20250711MS	BF143122.D	16 Jul 2025 22:09		RC/JU	Ok,M
21	Q2592-02MSD	WC-SOIL-20250711MS	BF143123.D	16 Jul 2025 22:38		RC/JU	Ok,M
22	Q2600-02	TRENCH	BF143124.D	16 Jul 2025 23:08		RC/JU	Ok
23	Q2600-06	STOCK-PILE	BF143125.D	16 Jul 2025 23:37		RC/JU	Ok
24	Q2600-10	END-OF-TRENCH	BF143126.D	17 Jul 2025 00:07		RC/JU	Ok,M
25	Q2614-06	HR-MCN-COMP-01	BF143127.D	17 Jul 2025 00:37		RC/JU	Ok
26	Q2605-03	VB15061	BF143128.D	17 Jul 2025 01:05		RC/JU	Ok
27	Q2605-04	V897	BF143129.D	17 Jul 2025 01:35		RC/JU	Ok
28	Q2571-05	TP-17	BF143130.D	17 Jul 2025 02:04		RC/JU	Ok
29	Q2571-05MS	TP-17MS	BF143131.D	17 Jul 2025 02:34		RC/JU	Ok
30	Q2571-05MSD	TP-17MSD	BF143132.D	17 Jul 2025 03:03		RC/JU	Ok
31	Q2555-01	OU4-TS-29-070925	BF143133.D	17 Jul 2025 03:32		RC/JU	Ok,M
32	Q2558-01	OU4-TS-Denali-070925	BF143134.D	17 Jul 2025 04:02		RC/JU	Ok,M
33	Q2555-03	OU4-TS-30-070925	BF143135.D	17 Jul 2025 04:31		RC/JU	Ok,M
34	Q2558-03	OU4-TS-Grillo-OG-070925	BF143136.D	17 Jul 2025 05:00		RC/JU	Ok,M
35	SSTDCCC040	SSTDCCC040EC	BF143137.D	17 Jul 2025 05:30		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF071725

Review By	Rahul	Review On	7/18/2025 11:02:11 AM
Supervise By	Jagrut	Supervise On	7/18/2025 1:30:54 PM
SubDirectory	BF071725	HP Acquire Method	BNA_F
		HP Processing Method	bf071725

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12674,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143138.D	17 Jul 2025 09:58		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF143139.D	17 Jul 2025 10:28	A Fresh Calibration is required.	RC/JU	Not Ok
3	SSTDICC2.5	SSTDICC2.5	BF143140.D	17 Jul 2025 11:04		RC/JU	Ok
4	SSTDICC005	SSTDICC005	BF143141.D	17 Jul 2025 11:34		RC/JU	Ok,M
5	SSTDICC010	SSTDICC010	BF143142.D	17 Jul 2025 12:04		RC/JU	Ok,M
6	SSTDICC020	SSTDICC020	BF143143.D	17 Jul 2025 12:34		RC/JU	Ok,M
7	SSTDICCC040	SSTDICCC040	BF143144.D	17 Jul 2025 13:03	Compound #32,54,65 Kept on LR	RC/JU	Ok,M
8	SSTDICC050	SSTDICC050	BF143145.D	17 Jul 2025 13:33		RC/JU	Ok,M
9	SSTDICC060	SSTDICC060	BF143146.D	17 Jul 2025 14:04		RC/JU	Ok,M
10	SSTDICC080	SSTDICC080	BF143147.D	17 Jul 2025 14:34	Compound#77 removed from 80 ppm	RC/JU	Ok,M
11	SSTDICV040	ICVBF071725	BF143148.D	17 Jul 2025 15:06		RC/JU	Ok,M
12	PB168813BL	PB168813BL	BF143149.D	17 Jul 2025 15:36		RC/JU	Ok
13	Q2126-03	MDL-SOIL-03-QT2-202	BF143150.D	17 Jul 2025 16:06	MDL-SOIL 4 ppm	RC/JU	Ok,M
14	Q2126-03	MDL-SOIL-03-QT2-202	BF143151.D	17 Jul 2025 16:36	MDL-SOIL 8 ppm	RC/JU	Ok,M
15	Q2126-09	MDL-WATER-03-QT2-2	BF143152.D	17 Jul 2025 17:07	MDL-WATER 4 ppm	RC/JU	Ok,M
16	Q2126-09	MDL-WATER-03-QT2-2	BF143153.D	17 Jul 2025 17:37	MDL-WATER 8 ppm	RC/JU	Ok,M
17	PB168816BL	PB168816BL	BF143154.D	17 Jul 2025 18:07		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF071725

Review By	Rahul	Review On	7/18/2025 11:02:11 AM		
Supervise By	Jagrut	Supervise On	7/18/2025 1:30:54 PM		
SubDirectory	BF071725	HP Acquire Method	BNA_F	HP Processing Method	bf071725

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12674,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	PB167904BL	PB167904BL	BF143155.D	17 Jul 2025 18:37		RC/JU	Ok
19	PB167904BS	PB167904BS	BF143156.D	17 Jul 2025 19:07		RC/JU	Ok,M
20	SSTDCCC040	SSTDCCC040EC	BF143157.D	17 Jul 2025 19:37		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF072225

Review By	Rahul	Review On	7/23/2025 8:23:11 AM
Supervise By	Jagrut	Supervise On	7/23/2025 1:06:52 PM
SubDirectory	BF072225	HP Acquire Method	BNA_F
		HP Processing Method	bf071725

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S13168,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143176.D	22 Jul 2025 07:48		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF143177.D	22 Jul 2025 08:46		RC/JU	Ok
3	PB168929BL	PB168929BL	BF143178.D	22 Jul 2025 09:15		RC/JU	Ok
4	PB168813BS	PB168813BS	BF143179.D	22 Jul 2025 09:44		RC/JU	Ok,M
5	Q2635-01MS	RT2286MS	BF143180.D	22 Jul 2025 10:26		RC/JU	Ok,M
6	Q2635-01MSD	RT2286MSD	BF143181.D	22 Jul 2025 10:55		RC/JU	Ok,M
7	Q2638-01	OU4-TS-31-071725	BF143182.D	22 Jul 2025 11:25		RC/JU	Ok,M
8	Q2638-03	OU4-TS-32-071725	BF143183.D	22 Jul 2025 11:54		RC/JU	Ok,M
9	Q2638-05	OU4-TS-33-071725	BF143184.D	22 Jul 2025 12:23		RC/JU	Ok,M
10	Q2638-07	OU4-TS-34-071725	BF143185.D	22 Jul 2025 12:52	Surrogate Fail	RC/JU	ReRun
11	Q2638-09	OU4-TS-35-071725	BF143186.D	22 Jul 2025 13:21		RC/JU	Ok,M
12	SSTDCCC040	SSTDCCC040EC	BF143187.D	22 Jul 2025 13:55		RC/JU	Ok
13	DFTPP	DFTPP	BF143188.D	22 Jul 2025 14:54		RC/JU	Ok
14	SSTDCCC040	SSTDCCC040	BF143189.D	22 Jul 2025 15:23		RC/JU	Ok
15	PB168954BL	PB168954BL	BF143190.D	22 Jul 2025 15:52		RC/JU	Ok
16	Q2641-02	P001-CONCRETE001	BF143191.D	22 Jul 2025 16:26		RC/JU	Ok
17	Q2649-04	WC-1	BF143192.D	22 Jul 2025 16:55		RC/JU	Ok,M
18	Q2649-08	WC-2	BF143193.D	22 Jul 2025 17:25		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF072225

Review By	Rahul	Review On	7/23/2025 8:23:11 AM		
Supervise By	Jagrut	Supervise On	7/23/2025 1:06:52 PM		
SubDirectory	BF072225	HP Acquire Method	BNA_F	HP Processing Method	bf071725
STD. NAME		STD REF.#			
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840				
CCC	SP6836				
Internal Standard/PEM	S13168,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2649-12	WC-3	BF143194.D	22 Jul 2025 17:54		RC/JU	Ok
20	Q2649-16	WC-4	BF143195.D	22 Jul 2025 18:23		RC/JU	Ok
21	Q2649-20	WC-5	BF143196.D	22 Jul 2025 18:53		RC/JU	Ok
22	Q2649-24	WC-6	BF143197.D	22 Jul 2025 19:22		RC/JU	Ok
23	Q2646-03	FRAC TANK	BF143198.D	22 Jul 2025 19:51	Internal standard fail.	RC/JU	Ok
24	Q2646-03MS	FRAC TANKMS	BF143199.D	22 Jul 2025 20:20		RC/JU	Ok,M
25	Q2646-03MSD	FRAC TANKMSD	BF143200.D	22 Jul 2025 20:49		RC/JU	Ok,M
26	Q2649-21	WC-6	BF143201.D	22 Jul 2025 21:18		RC/JU	Ok
27	Q2649-13	WC-4	BF143202.D	22 Jul 2025 21:48		RC/JU	Ok
28	Q2658-02	PAD-072125-2	BF143203.D	22 Jul 2025 22:17		RC/JU	Ok
29	Q2658-01	PAD-072125-1	BF143204.D	22 Jul 2025 22:46	Internal standard fail.	RC/JU	Ok,M
30	Q2660-01	MOO-25-0205	BF143205.D	22 Jul 2025 23:15	Internal standard fail.	RC/JU	Ok
31	Q2649-17	WC-5	BF143206.D	22 Jul 2025 23:43		RC/JU	Ok,M
32	Q2649-01	WC-1	BF143207.D	23 Jul 2025 00:12		RC/JU	Ok,M
33	Q2660-03	MOO-25-0218	BF143208.D	23 Jul 2025 00:41	Internal standard fail	RC/JU	Ok

M : Manual Integration