

Surrogate Summary

SDG No.: **Q2555**

Client: **Nobis Group**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2555-01	OU4-TS-29-070925	1,2-Dichloroethane-d4	50	54.1	108		71	136
		Dibromofluoromethane	50	47.6	95		78	119
		Toluene-d8	50	45.3	91		85	116
		4-Bromofluorobenzene	50	43.4	87		79	119
Q2555-03	OU4-TS-30-070925	1,2-Dichloroethane-d4	50	55.2	110		71	136
		Dibromofluoromethane	50	46.9	94		78	119
		Toluene-d8	50	45.9	92		85	116
		4-Bromofluorobenzene	50	46.5	93		79	119
VW0710SBL01	VW0710SBL01	1,2-Dichloroethane-d4	50	55.9	112		71	136
		Dibromofluoromethane	50	46.0	92		78	119
		Toluene-d8	50	45.5	91		85	116
		4-Bromofluorobenzene	50	45.6	91		79	119
VW0710SBS01	VW0710SBS01	1,2-Dichloroethane-d4	50	48.8	98		71	136
		Dibromofluoromethane	50	47.6	95		78	119
		Toluene-d8	50	49.3	99		85	116
		4-Bromofluorobenzene	50	51.3	103		79	119
VW0710SBSD0	VW0710SBSD01	1,2-Dichloroethane-d4	50	51.5	103		71	136
		Dibromofluoromethane	50	49.9	100		78	119
		Toluene-d8	50	49.8	100		85	116
		4-Bromofluorobenzene	50	51.9	104		79	119