

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: NOBI03
 Lab Code: ACE SDG NO.: Q2555
 Lab File ID: VW031791.D Date Analyzed: 07/10/2025
 Instrument ID: MSVOA_W Time Analyzed: 08:47
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	245114	7.97	431816	8.85	372703	11.63
UPPER LIMIT	490228	8.465	863632	9.349	745406	12.129
LOWER LIMIT	122557	7.465	215908	8.349	186352	11.129
EPA SAMPLE NO.						
OU4-TS-29-070925	166057	7.97	358196	8.85	316208	11.63
OU4-TS-30-070925	164102	7.97	356847	8.86	318986	11.64
VW0710SBL01	178449	7.96	403423	8.85	368923	11.64
VW0710SBS01	227800	7.97	411611	8.86	356304	11.64
VW0710SBSD01	226314	7.97	415001	8.85	365558	11.63

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.