

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: NOBI03
 Lab Code: ACE SDG NO.: Q2555
 Lab File ID: VW031791.D Date Analyzed: 07/10/2025
 Instrument ID: MSVOA_W Time Analyzed: 08:47
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	173821	13.556				
UPPER LIMIT	347642	14.056				
LOWER LIMIT	86910.5	13.056				
EPA SAMPLE NO.						
OU4-TS-29-070925	137883	13.56				
OU4-TS-30-070925	145135	13.56				
VW0710SBL01	171872	13.56				
VW0710SBS01	178658	13.56				
VW0710SBSD01	173637	13.56				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.