



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Nobis Group

Project Name: Raymark Superfund Site

Project # N/A

Order ID # Q2558

Test Name: SPLP Mercury,SPLP MetalGroup3

A. Number of Samples and Date of Receipt:

4 Solid samples were received on 07/10/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Cyanide, Herbicide Group1, Mercury, Metals ICP-TAL, METALS-TAL, PCB, Pesticide-TCL, SPLP Extraction, SPLP Mercury, SPLP MetalGroup3, SPLP MetalGroup4, SPLP MetalGroup6, SVOCMS Group3 and VOCMS Group3. This data package contains results for SPLP Mercury,SPLP MetalGroup3.

C. Analytical Techniques:

The analysis of SPLP MetalGroup3 was based on method 6020B, digestion based on method 3010 (water). The analysis of SPLP Mercury was based on method 7470A and digestion was based on method 7470 (water).

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all compounds.

The Duplicate analysis met criteria for all compounds.

The Matrix Spike (OU4-TS-29-070925MS) analysis met criteria for all compounds except for Barium, and Silver due to Chemical Interference during Digestion Process.

The Matrix Spike Duplicate (OU4-TS-29-070925MSD) analysis met criteria for all compounds except for Barium and Silver due to Chemical Interference during Digestion Process.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

E. Additional Comments:

The Post Digest Spike (OU4-TS-29-070925A) analysis met criteria for all compounds except for Barium and Silver due to unknown chemical interference of matrix with the addition of spike amount after digestion and before analysis; matrix has suppression effect during addition of spike.

Q2558 SPLP all samples diluted 5X Straight due to SPLP fluid which cannot be injected as is without dilution to avoid damage to detector of instrument.



Internal standard 89Y(1) was outside qc limit for samples Q2558-02 in Original so for these samples affected parameters are reported from its Dilution.

In analytical Sequence LB136601, The % Recovery Outside limit for Beryllium of ICV01, CCV01 and LLICV01 but, no any sample parameter affected under these calibration. Only Internal standard 89Y(1 and 2) affected parameters are reported from this analytical sequence.

Fax and hard copy is not matching for Q2558-02 for selenium parameter due to at time of fax analysis its internal standard frailer dilutions are not analyzed in sequence so as corrective action lab analyzed its dilution in another sequence. Hard copy is reported correct.

Collision cell is being used to remove potential interferences. The analytes Na, Mg, Al, K, V, Cr, Mn, Fe, Co, Ni, Cu, Zn, As are being analyzed with collision cell and analytes Be, B, Ca, Ti, Se, Sr, Zr, Mo, Ag, Cd, Sn, Sb, Ba, Tl, Pb, U are being analyzed with Non-Collision Cell. Helium gas is used for the Collision Cell analysis.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____