

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136438

Review By	MOHAN	Review On	7/14/2025 1:14:06 PM
Supervise By	jaswal	Supervise On	7/14/2025 11:09:31 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86338,MP86339,MP86340,MP86341,MP86342,MP86343		
ICV Standard	MP86344		
CCV Standard	MP86346		
ICSA Standard			
CRI Standard	MP86348		
LCS Standard			
Chk Standard	MP86345,MP86347,MP86349,MP86353		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	07/11/25 13:56		MOHAN	OK
2	S0.2	S0.2	CAL2	07/11/25 13:58		MOHAN	OK
3	S2.5	S2.5	CAL3	07/11/25 14:00		MOHAN	OK
4	S5	S5	CAL4	07/11/25 14:03		MOHAN	OK
5	S7.5	S7.5	CAL5	07/11/25 14:05		MOHAN	OK
6	S10	S10	CAL6	07/11/25 14:10		MOHAN	OK
7	ICV25	ICV25	ICV	07/11/25 14:13		MOHAN	OK
8	ICB25	ICB25	ICB	07/11/25 14:15		MOHAN	OK
9	CCV82	CCV82	CCV	07/11/25 14:18		MOHAN	OK
10	CCB82	CCB82	CCB	07/11/25 14:25		MOHAN	OK
11	CRA	CRA	CRDL	07/11/25 14:28		MOHAN	OK
12	HighStd	HighStd	HIGH STD	07/11/25 14:30		MOHAN	OK
13	ChkStd	ChkStd	SAM	07/11/25 14:32		MOHAN	OK
14	PB168821BL	PB168821BL	MB	07/11/25 14:35		MOHAN	OK
15	PB168821BS	PB168821BS	LCS	07/11/25 14:40		MOHAN	OK
16	Q2537-01	TR-04-07092025	SAM	07/11/25 14:42		MOHAN	OK
17	Q2539-01	HACKENSACK	SAM	07/11/25 14:44		MOHAN	OK
18	Q2542-01	VNJ-245	SAM	07/11/25 14:47		MOHAN	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136438

Review By	MOHAN	Review On	7/14/2025 1:14:06 PM
Supervise By	jaswal	Supervise On	7/14/2025 11:09:31 PM

STD. NAME	STD REF.#
ICAL Standard	MP86338,MP86339,MP86340,MP86341,MP86342,MP86343
ICV Standard	MP86344
CCV Standard	MP86346
ICSA Standard	
CRI Standard	MP86348
LCS Standard	
Chk Standard	MP86345,MP86347,MP86349,MP86353

19	Q2543-01	TP-41	SAM	07/11/25 14:49		MOHAN	OK
20	Q2543-02	TP-49	SAM	07/11/25 14:51		MOHAN	OK
21	CCV83	CCV83	CCV	07/11/25 14:56		MOHAN	OK
22	CCB83	CCB83	CCB	07/11/25 15:01		MOHAN	OK
23	Q2543-03	TP-23	SAM	07/11/25 15:04		MOHAN	OK
24	Q2543-04	TP-23-99	SAM	07/11/25 15:06		MOHAN	OK
25	Q2546-01	HD-02-7-9-2025	SAM	07/11/25 15:08		MOHAN	OK
26	Q2550-02	VNJ-203	SAM	07/11/25 15:10		MOHAN	OK
27	Q2551-02	VNJ-204	SAM	07/11/25 15:13		MOHAN	OK
28	Q2555-01	OU4-TS-29-070925	SAM	07/11/25 15:15		MOHAN	OK
29	Q2555-03	OU4-TS-30-070925	SAM	07/11/25 15:17		MOHAN	OK
30	Q2556-01	RT3997	SAM	07/11/25 15:20		MOHAN	OK
31	Q2557-01	OILY SPILL DEBRIS	SAM	07/11/25 15:22		MOHAN	OK
32	Q2558-01	OU4-TS-Denali-070925	SAM	07/11/25 15:24		MOHAN	OK
33	CCV84	CCV84	CCV	07/11/25 15:26		MOHAN	OK
34	CCB84	CCB84	CCB	07/11/25 15:29		MOHAN	OK
35	Q2558-03	OU4-TS-Grillo-OG-07	SAM	07/11/25 15:36		MOHAN	OK
36	Q2558-03DUP	OU4-TS-Grillo-OG-07	DUP	07/11/25 15:38		MOHAN	OK
37	Q2558-03MS	OU4-TS-Grillo-OG-07	MS	07/11/25 15:41		MOHAN	OK
38	Q2558-03MSD	OU4-TS-Grillo-OG-07	MSD	07/11/25 15:43		MOHAN	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136438

Review By	MOHAN	Review On	7/14/2025 1:14:06 PM
Supervise By	jaswal	Supervise On	7/14/2025 11:09:31 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86338,MP86339,MP86340,MP86341,MP86342,MP86343		
ICV Standard	MP86344		
CCV Standard	MP86346		
ICSA Standard			
CRI Standard	MP86348		
LCS Standard			
Chk Standard	MP86345,MP86347,MP86349,MP86353		

39	Q2560-01	LP-7102025	SAM	07/11/25 15:45		MOHAN	OK
40	PB168822BL	PB168822BL	MB	07/11/25 15:48		MOHAN	OK
41	PB168822BS	PB168822BS	LCS	07/11/25 15:50		MOHAN	OK
42	Q2559-01	500-3B CONCRETE C	SAM	07/11/25 15:52		MOHAN	OK
43	Q2564-01	ARS20-0030	SAM	07/11/25 15:55		MOHAN	OK
44	Q2564-03	ARS20-0013	SAM	07/11/25 15:57		MOHAN	OK
45	CCV85	CCV85	CCV	07/11/25 16:02		MOHAN	OK
46	CCB85	CCB85	CCB	07/11/25 16:04		MOHAN	OK
47	Q2564-05	ARS20-0039	SAM	07/11/25 16:06		MOHAN	OK
48	Q2564-05DUP	ARS20-0039DUP	DUP	07/11/25 16:09		MOHAN	OK
49	Q2564-05MS	ARS20-0039MS	MS	07/11/25 16:11		MOHAN	OK
50	Q2564-05MSD	ARS20-0039MSD	MSD	07/11/25 16:16		MOHAN	OK
51	Q2571-01	TP-18	SAM	07/11/25 16:18		MOHAN	OK
52	Q2571-05	TP-17	SAM	07/11/25 16:21		MOHAN	OK
53	Q2558-03L	OU4-TS-Grillo-OG-07	SD	07/11/25 16:23		MOHAN	OK
54	Q2558-03A	OU4-TS-Grillo-OG-07	PS	07/11/25 16:25		MOHAN	OK
55	Q2564-05L	ARS20-0039L	SD	07/11/25 16:27		MOHAN	OK
56	Q2564-05A	ARS20-0039A	PS	07/11/25 16:30		MOHAN	OK
57	CCV86	CCV86	CCV	07/11/25 16:32		MOHAN	OK
58	CCB86	CCB86	CCB	07/11/25 16:34		MOHAN	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136469

Review By	Janvi	Review On	7/15/2025 11:43:51 AM
Supervise By	jaswal	Supervise On	7/16/2025 6:41:06 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86213		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86215		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	07/14/25 13:05		Janvi	OK
2	S1	S1	CAL2	07/14/25 13:09		Janvi	OK
3	S2	S2	CAL3	07/14/25 13:14		Janvi	OK
4	S3	S3	CAL4	07/14/25 13:18		Janvi	OK
5	S4	S4	CAL5	07/14/25 13:22		Janvi	OK
6	S5	S5	CAL6	07/14/25 13:26		Janvi	OK
7	ICV01	ICV01	ICV	07/14/25 13:44		Janvi	OK
8	ICB01	ICB01	ICB	07/14/25 14:16		Janvi	OK
9	CRI01	CRI01	CRDL	07/14/25 14:26		Janvi	OK
10	ICSA01	ICSA01	ICSA	07/14/25 14:37		Janvi	OK
11	ICSAB01	ICSAB01	ICSAB	07/14/25 14:41		Janvi	OK
12	ICSADL	ICSADL	ICSA	07/14/25 14:46		Janvi	OK
13	ICSABDL	ICSABDL	ICSAB	07/14/25 14:50		Janvi	OK
14	CCV01	CCV01	CCV	07/14/25 14:54		Janvi	OK
15	CCB01	CCB01	CCB	07/14/25 14:59		Janvi	OK
16	Q2540-01	A597	SAM	07/14/25 15:05		Janvi	OK
17	Q2487-11DL2	G3(0-6)DL2	SAM	07/14/25 15:09	10x for Zn	Janvi	Confirms
18	Q2487-16DL2	G1(6-12)DL2	SAM	07/14/25 15:14	10x for Zn	Janvi	Confirms

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136469

Review By	Janvi	Review On	7/15/2025 11:43:51 AM
Supervise By	jaswal	Supervise On	7/16/2025 6:41:06 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86213		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86215		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

19	Q2545-01DUP	3918DUP	DUP	07/14/25 15:24		Janvi	OK
20	Q2545-01	3918	SAM	07/14/25 15:28		Janvi	OK
21	Q2545-01L	3918L	SD	07/14/25 15:39		Janvi	OK
22	Q2580-02	WC-URBAN-FILL-B2	SAM	07/14/25 15:43		Janvi	OK
23	Q2545-01MS	3918MS	MS	07/14/25 15:48		Janvi	OK
24	Q2545-01MSD	3918MSD	MSD	07/14/25 15:52		Janvi	OK
25	Q2545-01A	3918A	PS	07/14/25 15:56	0.1ml of M6013,M6004 in 10ml sample.	Janvi	OK
26	CCV02	CCV02	CCV	07/14/25 16:01		Janvi	OK
27	LLCCV01	LLCCV01	LLCCV	07/14/25 16:27		Janvi	OK
28	CCB02	CCB02	CCB	07/14/25 16:57		Janvi	OK
29	PB168793BL	PB168793BL	MB	07/14/25 17:03		Janvi	OK
30	PB168793BS	PB168793BS	LCS	07/14/25 17:07		Janvi	OK
31	PB168808BL	PB168808BL	MB	07/14/25 17:12		Janvi	OK
32	Q2557-02	OILY WATER	SAM	07/14/25 17:20		Janvi	OK
33	PB168808BS	PB168808BS	LCS	07/14/25 17:34		Janvi	OK
34	Q2557-02DUP	OILY WATERDUP	DUP	07/14/25 17:38	K oversaturated	Janvi	Dilution
35	Q2557-02L	OILY WATERL	SD	07/14/25 17:43		Janvi	OK
36	Q2557-02MS	OILY WATERMS	MS	07/14/25 17:47	MS/MSD & DUP RPD fail for more than 50% parameters	Janvi	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136469

Review By	Janvi	Review On	7/15/2025 11:43:51 AM
Supervise By	jaswal	Supervise On	7/16/2025 6:41:06 PM

STD. NAME	STD REF.#
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212
ICV Standard	MP86213
CCV Standard	MP86216
ICSA Standard	MP86214,MP86215
CRI Standard	MP86212
LCS Standard	
Chk Standard	MP86217,MP86218

37	Q2557-02MSD	OILY WATERMSD	MSD	07/14/25 17:52	MS/MSD & DUP RPD fail for more than 50% parameters	Janvi	OK
38	Q2580-02DL	WC-URBAN-FILL-B2	SAM	07/14/25 17:56	NOT USE	Janvi	Not Ok
39	CCV03	CCV03	CCV	07/14/25 18:01		Janvi	OK
40	CCB03	CCB03	CCB	07/14/25 18:05		Janvi	OK
41	Q2557-02A	OILY WATERA	PS	07/14/25 18:09	0.1ml of M6013,M6004 in 10ml sample.	Janvi	OK
42	PB168806BL	PB168806BL	MB	07/14/25 18:15		Janvi	OK
43	Q2556-01	RT3997	SAM	07/14/25 18:24		Janvi	OK
44	Q2557-01	OILY SPILL DEBRIS	SAM	07/14/25 18:28		Janvi	OK
45	PB168806BS	PB168806BS	LCS	07/14/25 18:33		Janvi	OK
46	Q2559-01	500-3B CONCRETE C	SAM	07/14/25 18:37	Not use	Janvi	Not Ok
47	Q2560-01	LP-7102025	SAM	07/14/25 18:42		Janvi	OK
48	Q2564-01	ARS20-0030	SAM	07/14/25 18:46		Janvi	OK
49	Q2564-03	ARS20-0013	SAM	07/14/25 18:51		Janvi	OK
50	Q2564-05	ARS20-0039	SAM	07/14/25 18:55	Not Use	Janvi	Not Ok
51	CCV04	CCV04	CCV	07/14/25 19:00		Janvi	OK
52	CCB04	CCB04	CCB	07/14/25 19:04		Janvi	OK
53	Q2571-01	TP-18	SAM	07/14/25 19:09		Janvi	OK
54	Q2564-05DUP	ARS20-0039DUP	DUP	07/14/25 19:13	Not Use	Janvi	Not Ok
55	Q2564-05L	ARS20-0039L	SD	07/14/25 19:18	Not Use	Janvi	Not Ok

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136469

Review By	Janvi	Review On	7/15/2025 11:43:51 AM
Supervise By	jaswal	Supervise On	7/16/2025 6:41:06 PM

STD. NAME	STD REF.#
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212
ICV Standard	MP86213
CCV Standard	MP86216
ICSA Standard	MP86214,MP86215
CRI Standard	MP86212
LCS Standard	
Chk Standard	MP86217,MP86218

56	Q2564-05MS	ARS20-0039MS	MS	07/14/25 19:22	Not Use	Janvi	Not Ok
57	Q2564-05MSD	ARS20-0039MSD	MSD	07/14/25 19:26	Not Use	Janvi	Not Ok
58	Q2564-05A	ARS20-0039A	PS	07/14/25 19:31	Not Use	Janvi	Not Ok
59	Q2555-01	OU4-TS-29-070925	SAM	07/14/25 19:35		Janvi	OK
60	Q2555-03	OU4-TS-30-070925	SAM	07/14/25 19:40		Janvi	OK
61	Q2558-01	OU4-TS-Denali-070925	SAM	07/14/25 19:44		Janvi	OK
62	Q2558-03	OU4-TS-Grillo-OG-07	SAM	07/14/25 19:49		Janvi	OK
63	CCV05	CCV05	CCV	07/14/25 19:53		Janvi	OK
64	CCB05	CCB05	CCB	07/14/25 19:58		Janvi	OK
65	Q2571-05	TP-17	SAM	07/14/25 20:02		Janvi	OK
66	CCV06	CCV06	CCV	07/14/25 20:07		Janvi	OK
67	CCB06	CCB06	CCB	07/14/25 20:11		Janvi	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136488

Review By	Janvi	Review On	7/16/2025 12:40:30 PM
Supervise By	jaswal	Supervise On	7/16/2025 6:31:51 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212		
ICV Standard	MP86213		
CCV Standard	MP86216		
ICSA Standard	MP86214,MP86215		
CRI Standard	MP86212		
LCS Standard			
Chk Standard	MP86217,MP86218		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	07/15/25 19:01		Janvi	OK
2	S1	S1	CAL2	07/15/25 19:05		Janvi	OK
3	S2	S2	CAL3	07/15/25 19:10		Janvi	OK
4	S3	S3	CAL4	07/15/25 19:14		Janvi	OK
5	S4	S4	CAL5	07/15/25 19:18		Janvi	OK
6	S5	S5	CAL6	07/15/25 19:22		Janvi	OK
7	ICV01	ICV01	ICV	07/15/25 19:27		Janvi	OK
8	LLICV01	LLICV01	LLICV	07/15/25 19:37		Janvi	OK
9	ICB01	ICB01	ICB	07/15/25 19:42		Janvi	OK
10	CRI01	CRI01	CRDL	07/15/25 19:56		Janvi	OK
11	ICSA01	ICSA01	ICSA	07/15/25 20:01		Janvi	OK
12	ICSAB01	ICSAB01	ICSAB	07/15/25 20:05		Janvi	OK
13	ICSADL	ICSADL	ICSA	07/15/25 20:10		Janvi	OK
14	ICSABDL	ICSABDL	ICSAB	07/15/25 20:20		Janvi	OK
15	CCV01	CCV01	CCV	07/15/25 20:25		Janvi	OK
16	CCB01	CCB01	CCB	07/15/25 20:29		Janvi	OK
17	PB168772BL	PB168772BL	MB	07/15/25 20:34		Janvi	OK
18	PB168772BS	PB168772BS	LCS	07/15/25 21:01		Janvi	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB136488

Review By	Janvi	Review On	7/16/2025 12:40:30 PM
Supervise By	jaswal	Supervise On	7/16/2025 6:31:51 PM

STD. NAME	STD REF.#
ICAL Standard	MP86192,MP86193,MP86194,MP86195,MP86196,MP86212
ICV Standard	MP86213
CCV Standard	MP86216
ICSA Standard	MP86214,MP86215
CRI Standard	MP86212
LCS Standard	
Chk Standard	MP86217,MP86218

19	Q2564-05	ARS20-0039	SAM	07/15/25 21:06		Janvi	OK
20	Q2564-05DUP	ARS20-0039DUP	DUP	07/15/25 21:11		Janvi	OK
21	Q2564-05L	ARS20-0039L	SD	07/15/25 21:15		Janvi	OK
22	Q2564-05MS	ARS20-0039MS	MS	07/15/25 21:20		Janvi	OK
23	Q2564-05MSD	ARS20-0039MSD	MSD	07/15/25 21:24		Janvi	OK
24	Q2564-05A	ARS20-0039A	PS	07/15/25 21:29	0.1ml of M6013,M6004 in 10 ml of sample.	Janvi	OK
25	Q2559-01	500-3B CONCRETE C	SAM	07/15/25 21:33	Ca High	Janvi	Dilution
26	CCV02	CCV02	CCV	07/15/25 22:12		Janvi	OK
27	CCB02	CCB02	CCB	07/15/25 22:17		Janvi	OK
28	Q2559-01DL	500-3B CONCRETE C	SAM	07/15/25 22:21	5X For Ca	Janvi	Confirms
29	Q2520-06	A-6	SAM	07/15/25 22:25		Janvi	OK
30	Q2520-04	A-4	SAM	07/15/25 22:30		Janvi	OK
31	CCV03	CCV03	CCV	07/15/25 22:39		Janvi	OK
32	CCB03	CCB03	CCB	07/15/25 22:43		Janvi	OK