



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2558

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB25	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	07/11/2025	14:15	LB136438

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2558

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB82	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	07/11/2025	14:25	LB136438
CCB83	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	07/11/2025	15:01	LB136438
CCB84	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	07/11/2025	15:29	LB136438
CCB85	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	07/11/2025	16:04	LB136438
CCB86	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	07/11/2025	16:34	LB136438

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2558

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	27.3	+/-50	J	80.0	100	P	07/14/2025	14:16	LB136469
	Antimony	6.76	+/-25	U	12.5	50.0	P	07/14/2025	14:16	LB136469
	Arsenic	5.12	+/-10	U	16.0	20.0	P	07/14/2025	14:16	LB136469
	Barium	14.6	+/-50	U	25.0	100	P	07/14/2025	14:16	LB136469
	Beryllium	0.56	+/-3	U	1.50	6.00	P	07/14/2025	14:16	LB136469
	Cadmium	0.50	+/-3	U	1.50	6.00	P	07/14/2025	14:16	LB136469
	Calcium	234	+/-1000	U	500	2000	P	07/14/2025	14:16	LB136469
	Chromium	2.12	+/-5	U	5.00	10.0	P	07/14/2025	14:16	LB136469
	Cobalt	2.26	+/-15	U	7.50	30.0	P	07/14/2025	14:16	LB136469
	Copper	4.60	+/-10	U	16.0	20.0	P	07/14/2025	14:16	LB136469
	Iron	23.4	+/-50	U	80.0	100	P	07/14/2025	14:16	LB136469
	Lead	2.30	+/-6	U	9.60	12.0	P	07/14/2025	14:16	LB136469
	Magnesium	244	+/-1000	U	500	2000	P	07/14/2025	14:16	LB136469
	Manganese	5.94	+/-10	U	5.00	20.0	P	07/14/2025	14:16	LB136469
	Nickel	3.06	+/-20	U	10.0	40.0	P	07/14/2025	14:16	LB136469
	Potassium	918	+/-1000	U	1600	2000	P	07/14/2025	14:16	LB136469
	Selenium	9.64	+/-10	U	16.0	20.0	P	07/14/2025	14:16	LB136469
	Silver	1.62	+/-5	U	5.00	10.0	P	07/14/2025	14:16	LB136469
	Sodium	868	+/-1000	U	1000	2000	P	07/14/2025	14:16	LB136469
	Thallium	4.38	+/-20	U	20.0	40.0	P	07/14/2025	14:16	LB136469
	Vanadium	6.26	+/-20	U	20.0	40.0	P	07/14/2025	14:16	LB136469
	Zinc	3.50	+/-20	U	15.0	40.0	P	07/14/2025	14:16	LB136469

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2558

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	80.0	100	P	07/14/2025	14:59	LB136469
	Antimony	6.76	+/-25	U	12.5	50.0	P	07/14/2025	14:59	LB136469
	Arsenic	5.12	+/-10	U	16.0	20.0	P	07/14/2025	14:59	LB136469
	Barium	14.6	+/-50	U	25.0	100	P	07/14/2025	14:59	LB136469
	Beryllium	0.56	+/-3	U	1.50	6.00	P	07/14/2025	14:59	LB136469
	Cadmium	0.50	+/-3	U	1.50	6.00	P	07/14/2025	14:59	LB136469
	Calcium	234	+/-1000	U	500	2000	P	07/14/2025	14:59	LB136469
	Chromium	2.12	+/-5	U	5.00	10.0	P	07/14/2025	14:59	LB136469
	Cobalt	2.26	+/-15	U	7.50	30.0	P	07/14/2025	14:59	LB136469
	Copper	4.60	+/-10	U	16.0	20.0	P	07/14/2025	14:59	LB136469
	Iron	23.4	+/-50	U	80.0	100	P	07/14/2025	14:59	LB136469
	Lead	2.30	+/-6	U	9.60	12.0	P	07/14/2025	14:59	LB136469
	Magnesium	244	+/-1000	U	500	2000	P	07/14/2025	14:59	LB136469
	Manganese	5.94	+/-10	U	5.00	20.0	P	07/14/2025	14:59	LB136469
	Nickel	3.06	+/-20	U	10.0	40.0	P	07/14/2025	14:59	LB136469
	Potassium	918	+/-1000	U	1600	2000	P	07/14/2025	14:59	LB136469
	Selenium	9.64	+/-10	U	16.0	20.0	P	07/14/2025	14:59	LB136469
	Silver	1.62	+/-5	U	5.00	10.0	P	07/14/2025	14:59	LB136469
	Sodium	868	+/-1000	U	1000	2000	P	07/14/2025	14:59	LB136469
	Thallium	4.38	+/-20	U	20.0	40.0	P	07/14/2025	14:59	LB136469
	Vanadium	6.26	+/-20	U	20.0	40.0	P	07/14/2025	14:59	LB136469
	Zinc	3.50	+/-20	U	15.0	40.0	P	07/14/2025	14:59	LB136469
CCB02	Aluminum	16.9	+/-50	J	80.0	100	P	07/14/2025	16:57	LB136469
	Antimony	6.76	+/-25	U	12.5	50.0	P	07/14/2025	16:57	LB136469
	Arsenic	5.12	+/-10	U	16.0	20.0	P	07/14/2025	16:57	LB136469
	Barium	14.6	+/-50	U	25.0	100	P	07/14/2025	16:57	LB136469
	Beryllium	0.56	+/-3	U	1.50	6.00	P	07/14/2025	16:57	LB136469
	Cadmium	0.50	+/-3	U	1.50	6.00	P	07/14/2025	16:57	LB136469
	Calcium	234	+/-1000	U	500	2000	P	07/14/2025	16:57	LB136469
	Chromium	2.12	+/-5	U	5.00	10.0	P	07/14/2025	16:57	LB136469
	Cobalt	2.26	+/-15	U	7.50	30.0	P	07/14/2025	16:57	LB136469
	Copper	4.60	+/-10	U	16.0	20.0	P	07/14/2025	16:57	LB136469
	Iron	23.4	+/-50	U	80.0	100	P	07/14/2025	16:57	LB136469
	Lead	2.30	+/-6	U	9.60	12.0	P	07/14/2025	16:57	LB136469
	Magnesium	244	+/-1000	U	500	2000	P	07/14/2025	16:57	LB136469
	Manganese	5.94	+/-10	U	5.00	20.0	P	07/14/2025	16:57	LB136469
	Nickel	3.06	+/-20	U	10.0	40.0	P	07/14/2025	16:57	LB136469
	Potassium	918	+/-1000	U	1600	2000	P	07/14/2025	16:57	LB136469
	Selenium	9.64	+/-10	U	16.0	20.0	P	07/14/2025	16:57	LB136469

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2558

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	5.00	10.0	P	07/14/2025	16:57	LB136469
	Sodium	868	+/-1000	U	1000	2000	P	07/14/2025	16:57	LB136469
	Thallium	4.38	+/-20	U	20.0	40.0	P	07/14/2025	16:57	LB136469
	Vanadium	6.26	+/-20	U	20.0	40.0	P	07/14/2025	16:57	LB136469
	Zinc	3.50	+/-20	U	15.0	40.0	P	07/14/2025	16:57	LB136469
CCB03	Aluminum	20.5	+/-50	J	80.0	100	P	07/14/2025	18:05	LB136469
	Antimony	6.76	+/-25	U	12.5	50.0	P	07/14/2025	18:05	LB136469
	Arsenic	5.12	+/-10	U	16.0	20.0	P	07/14/2025	18:05	LB136469
	Barium	14.6	+/-50	U	25.0	100	P	07/14/2025	18:05	LB136469
	Beryllium	0.56	+/-3	U	1.50	6.00	P	07/14/2025	18:05	LB136469
	Cadmium	0.50	+/-3	U	1.50	6.00	P	07/14/2025	18:05	LB136469
	Calcium	234	+/-1000	U	500	2000	P	07/14/2025	18:05	LB136469
	Chromium	2.12	+/-5	U	5.00	10.0	P	07/14/2025	18:05	LB136469
	Cobalt	2.26	+/-15	U	7.50	30.0	P	07/14/2025	18:05	LB136469
	Copper	4.60	+/-10	U	16.0	20.0	P	07/14/2025	18:05	LB136469
	Iron	23.4	+/-50	U	80.0	100	P	07/14/2025	18:05	LB136469
	Lead	2.30	+/-6	U	9.60	12.0	P	07/14/2025	18:05	LB136469
	Magnesium	244	+/-1000	U	500	2000	P	07/14/2025	18:05	LB136469
	Manganese	5.94	+/-10	U	5.00	20.0	P	07/14/2025	18:05	LB136469
	Nickel	3.06	+/-20	U	10.0	40.0	P	07/14/2025	18:05	LB136469
	Potassium	918	+/-1000	U	1600	2000	P	07/14/2025	18:05	LB136469
	Selenium	9.64	+/-10	U	16.0	20.0	P	07/14/2025	18:05	LB136469
	Silver	1.62	+/-5	U	5.00	10.0	P	07/14/2025	18:05	LB136469
	Sodium	868	+/-1000	U	1000	2000	P	07/14/2025	18:05	LB136469
	Thallium	4.38	+/-20	U	20.0	40.0	P	07/14/2025	18:05	LB136469
	Vanadium	6.26	+/-20	U	20.0	40.0	P	07/14/2025	18:05	LB136469
	Zinc	3.50	+/-20	U	15.0	40.0	P	07/14/2025	18:05	LB136469
CCB04	Aluminum	11.3	+/-50	U	80.0	100	P	07/14/2025	19:04	LB136469
	Antimony	6.76	+/-25	U	12.5	50.0	P	07/14/2025	19:04	LB136469
	Arsenic	5.12	+/-10	U	16.0	20.0	P	07/14/2025	19:04	LB136469
	Barium	14.6	+/-50	U	25.0	100	P	07/14/2025	19:04	LB136469
	Beryllium	0.56	+/-3	U	1.50	6.00	P	07/14/2025	19:04	LB136469
	Cadmium	0.50	+/-3	U	1.50	6.00	P	07/14/2025	19:04	LB136469
	Calcium	234	+/-1000	U	500	2000	P	07/14/2025	19:04	LB136469
	Chromium	2.12	+/-5	U	5.00	10.0	P	07/14/2025	19:04	LB136469
	Cobalt	2.26	+/-15	U	7.50	30.0	P	07/14/2025	19:04	LB136469
	Copper	4.60	+/-10	U	16.0	20.0	P	07/14/2025	19:04	LB136469
	Iron	23.4	+/-50	U	80.0	100	P	07/14/2025	19:04	LB136469
	Lead	2.30	+/-6	U	9.60	12.0	P	07/14/2025	19:04	LB136469

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2558

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	500	2000	P	07/14/2025	19:04	LB136469
	Manganese	5.94	+/-10	U	5.00	20.0	P	07/14/2025	19:04	LB136469
	Nickel	3.06	+/-20	U	10.0	40.0	P	07/14/2025	19:04	LB136469
	Potassium	918	+/-1000	U	1600	2000	P	07/14/2025	19:04	LB136469
	Selenium	9.64	+/-10	U	16.0	20.0	P	07/14/2025	19:04	LB136469
	Silver	1.62	+/-5	U	5.00	10.0	P	07/14/2025	19:04	LB136469
	Sodium	868	+/-1000	U	1000	2000	P	07/14/2025	19:04	LB136469
	Thallium	4.38	+/-20	U	20.0	40.0	P	07/14/2025	19:04	LB136469
	Vanadium	6.26	+/-20	U	20.0	40.0	P	07/14/2025	19:04	LB136469
	Zinc	3.50	+/-20	U	15.0	40.0	P	07/14/2025	19:04	LB136469
CCB05	Aluminum	18.1	+/-50	J	80.0	100	P	07/14/2025	19:58	LB136469
	Antimony	6.76	+/-25	U	12.5	50.0	P	07/14/2025	19:58	LB136469
	Arsenic	5.12	+/-10	U	16.0	20.0	P	07/14/2025	19:58	LB136469
	Barium	14.6	+/-50	U	25.0	100	P	07/14/2025	19:58	LB136469
	Beryllium	0.56	+/-3	U	1.50	6.00	P	07/14/2025	19:58	LB136469
	Cadmium	0.50	+/-3	U	1.50	6.00	P	07/14/2025	19:58	LB136469
	Calcium	234	+/-1000	U	500	2000	P	07/14/2025	19:58	LB136469
	Chromium	2.12	+/-5	U	5.00	10.0	P	07/14/2025	19:58	LB136469
	Cobalt	2.26	+/-15	U	7.50	30.0	P	07/14/2025	19:58	LB136469
	Copper	4.60	+/-10	U	16.0	20.0	P	07/14/2025	19:58	LB136469
	Iron	23.4	+/-50	U	80.0	100	P	07/14/2025	19:58	LB136469
	Lead	2.30	+/-6	U	9.60	12.0	P	07/14/2025	19:58	LB136469
	Magnesium	244	+/-1000	U	500	2000	P	07/14/2025	19:58	LB136469
	Manganese	5.94	+/-10	U	5.00	20.0	P	07/14/2025	19:58	LB136469
	Nickel	3.06	+/-20	U	10.0	40.0	P	07/14/2025	19:58	LB136469
	Potassium	918	+/-1000	U	1600	2000	P	07/14/2025	19:58	LB136469
	Selenium	9.64	+/-10	U	16.0	20.0	P	07/14/2025	19:58	LB136469
	Silver	1.62	+/-5	U	5.00	10.0	P	07/14/2025	19:58	LB136469
	Sodium	868	+/-1000	U	1000	2000	P	07/14/2025	19:58	LB136469
	Thallium	4.38	+/-20	U	20.0	40.0	P	07/14/2025	19:58	LB136469
	Vanadium	6.26	+/-20	U	20.0	40.0	P	07/14/2025	19:58	LB136469
	Zinc	3.50	+/-20	U	15.0	40.0	P	07/14/2025	19:58	LB136469
CCB06	Aluminum	11.3	+/-50	U	80.0	100	P	07/14/2025	20:11	LB136469
	Antimony	6.76	+/-25	U	12.5	50.0	P	07/14/2025	20:11	LB136469
	Arsenic	5.12	+/-10	U	16.0	20.0	P	07/14/2025	20:11	LB136469
	Barium	14.6	+/-50	U	25.0	100	P	07/14/2025	20:11	LB136469
	Beryllium	0.56	+/-3	U	1.50	6.00	P	07/14/2025	20:11	LB136469
	Cadmium	0.50	+/-3	U	1.50	6.00	P	07/14/2025	20:11	LB136469
	Calcium	234	+/-1000	U	500	2000	P	07/14/2025	20:11	LB136469

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2558

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	5.00	10.0	P	07/14/2025	20:11	LB136469
	Cobalt	2.26	+/-15	U	7.50	30.0	P	07/14/2025	20:11	LB136469
	Copper	4.60	+/-10	U	16.0	20.0	P	07/14/2025	20:11	LB136469
	Iron	23.4	+/-50	U	80.0	100	P	07/14/2025	20:11	LB136469
	Lead	2.30	+/-6	U	9.60	12.0	P	07/14/2025	20:11	LB136469
	Magnesium	244	+/-1000	U	500	2000	P	07/14/2025	20:11	LB136469
	Manganese	5.94	+/-10	U	5.00	20.0	P	07/14/2025	20:11	LB136469
	Nickel	3.06	+/-20	U	10.0	40.0	P	07/14/2025	20:11	LB136469
	Potassium	918	+/-1000	U	1600	2000	P	07/14/2025	20:11	LB136469
	Selenium	9.64	+/-10	U	16.0	20.0	P	07/14/2025	20:11	LB136469
	Silver	1.62	+/-5	U	5.00	10.0	P	07/14/2025	20:11	LB136469
	Sodium	868	+/-1000	U	1000	2000	P	07/14/2025	20:11	LB136469
	Thallium	4.38	+/-20	U	20.0	40.0	P	07/14/2025	20:11	LB136469
	Vanadium	6.26	+/-20	U	20.0	40.0	P	07/14/2025	20:11	LB136469
	Zinc	3.50	+/-20	U	15.0	40.0	P	07/14/2025	20:11	LB136469

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2558

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	80.0	100	P	07/15/2025	19:42	LB136488
	Antimony	6.76	+/-25	U	12.5	50.0	P	07/15/2025	19:42	LB136488
	Arsenic	5.12	+/-10	U	16.0	20.0	P	07/15/2025	19:42	LB136488
	Barium	14.6	+/-50	U	25.0	100	P	07/15/2025	19:42	LB136488
	Beryllium	0.56	+/-3	U	1.50	6.00	P	07/15/2025	19:42	LB136488
	Cadmium	0.50	+/-3	U	1.50	6.00	P	07/15/2025	19:42	LB136488
	Calcium	234	+/-1000	U	500	2000	P	07/15/2025	19:42	LB136488
	Chromium	2.12	+/-5	U	5.00	10.0	P	07/15/2025	19:42	LB136488
	Cobalt	2.26	+/-15	U	7.50	30.0	P	07/15/2025	19:42	LB136488
	Copper	4.60	+/-10	U	16.0	20.0	P	07/15/2025	19:42	LB136488
	Iron	23.4	+/-50	U	80.0	100	P	07/15/2025	19:42	LB136488
	Lead	2.30	+/-6	U	9.60	12.0	P	07/15/2025	19:42	LB136488
	Magnesium	244	+/-1000	U	500	2000	P	07/15/2025	19:42	LB136488
	Manganese	5.94	+/-10	U	5.00	20.0	P	07/15/2025	19:42	LB136488
	Nickel	3.06	+/-20	U	10.0	40.0	P	07/15/2025	19:42	LB136488
	Potassium	918	+/-1000	U	1600	2000	P	07/15/2025	19:42	LB136488
	Selenium	9.64	+/-10	U	16.0	20.0	P	07/15/2025	19:42	LB136488
	Silver	1.62	+/-5	U	5.00	10.0	P	07/15/2025	19:42	LB136488
	Sodium	868	+/-1000	U	1000	2000	P	07/15/2025	19:42	LB136488
	Thallium	4.38	+/-20	U	20.0	40.0	P	07/15/2025	19:42	LB136488
	Vanadium	6.26	+/-20	U	20.0	40.0	P	07/15/2025	19:42	LB136488
	Zinc	3.50	+/-20	U	15.0	40.0	P	07/15/2025	19:42	LB136488

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2558

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	80.0	100	P	07/15/2025	20:29	LB136488
	Antimony	6.76	+/-25	U	12.5	50.0	P	07/15/2025	20:29	LB136488
	Arsenic	5.12	+/-10	U	16.0	20.0	P	07/15/2025	20:29	LB136488
	Barium	14.6	+/-50	U	25.0	100	P	07/15/2025	20:29	LB136488
	Beryllium	0.56	+/-3	U	1.50	6.00	P	07/15/2025	20:29	LB136488
	Cadmium	0.50	+/-3	U	1.50	6.00	P	07/15/2025	20:29	LB136488
	Calcium	234	+/-1000	U	500	2000	P	07/15/2025	20:29	LB136488
	Chromium	2.12	+/-5	U	5.00	10.0	P	07/15/2025	20:29	LB136488
	Cobalt	2.26	+/-15	U	7.50	30.0	P	07/15/2025	20:29	LB136488
	Copper	4.60	+/-10	U	16.0	20.0	P	07/15/2025	20:29	LB136488
	Iron	23.4	+/-50	U	80.0	100	P	07/15/2025	20:29	LB136488
	Lead	2.30	+/-6	U	9.60	12.0	P	07/15/2025	20:29	LB136488
	Magnesium	244	+/-1000	U	500	2000	P	07/15/2025	20:29	LB136488
	Manganese	5.94	+/-10	U	5.00	20.0	P	07/15/2025	20:29	LB136488
	Nickel	3.06	+/-20	U	10.0	40.0	P	07/15/2025	20:29	LB136488
	Potassium	918	+/-1000	U	1600	2000	P	07/15/2025	20:29	LB136488
	Selenium	9.64	+/-10	U	16.0	20.0	P	07/15/2025	20:29	LB136488
	Silver	1.62	+/-5	U	5.00	10.0	P	07/15/2025	20:29	LB136488
	Sodium	868	+/-1000	U	1000	2000	P	07/15/2025	20:29	LB136488
	Thallium	4.38	+/-20	U	20.0	40.0	P	07/15/2025	20:29	LB136488
	Vanadium	6.26	+/-20	U	20.0	40.0	P	07/15/2025	20:29	LB136488
	Zinc	3.50	+/-20	U	15.0	40.0	P	07/15/2025	20:29	LB136488
CCB02	Aluminum	11.3	+/-50	U	80.0	100	P	07/15/2025	22:17	LB136488
	Antimony	6.76	+/-25	U	12.5	50.0	P	07/15/2025	22:17	LB136488
	Arsenic	5.12	+/-10	U	16.0	20.0	P	07/15/2025	22:17	LB136488
	Barium	14.6	+/-50	U	25.0	100	P	07/15/2025	22:17	LB136488
	Beryllium	0.56	+/-3	U	1.50	6.00	P	07/15/2025	22:17	LB136488
	Cadmium	0.50	+/-3	U	1.50	6.00	P	07/15/2025	22:17	LB136488
	Calcium	234	+/-1000	U	500	2000	P	07/15/2025	22:17	LB136488
	Chromium	2.12	+/-5	U	5.00	10.0	P	07/15/2025	22:17	LB136488
	Cobalt	2.26	+/-15	U	7.50	30.0	P	07/15/2025	22:17	LB136488
	Copper	4.60	+/-10	U	16.0	20.0	P	07/15/2025	22:17	LB136488
	Iron	23.4	+/-50	U	80.0	100	P	07/15/2025	22:17	LB136488
	Lead	2.30	+/-6	U	9.60	12.0	P	07/15/2025	22:17	LB136488
	Magnesium	244	+/-1000	U	500	2000	P	07/15/2025	22:17	LB136488
	Manganese	5.94	+/-10	U	5.00	20.0	P	07/15/2025	22:17	LB136488
	Nickel	3.06	+/-20	U	10.0	40.0	P	07/15/2025	22:17	LB136488
	Potassium	918	+/-1000	U	1600	2000	P	07/15/2025	22:17	LB136488
	Selenium	9.64	+/-10	U	16.0	20.0	P	07/15/2025	22:17	LB136488

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2558

Contract: NOBI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	5.00	10.0	P	07/15/2025	22:17	LB136488
	Sodium	868	+/-1000	U	1000	2000	P	07/15/2025	22:17	LB136488
	Thallium	4.38	+/-20	U	20.0	40.0	P	07/15/2025	22:17	LB136488
	Vanadium	6.26	+/-20	U	20.0	40.0	P	07/15/2025	22:17	LB136488
	Zinc	3.50	+/-20	U	15.0	40.0	P	07/15/2025	22:17	LB136488
CCB03	Aluminum	11.3	+/-50	U	80.0	100	P	07/15/2025	22:43	LB136488
	Antimony	6.76	+/-25	U	12.5	50.0	P	07/15/2025	22:43	LB136488
	Arsenic	5.12	+/-10	U	16.0	20.0	P	07/15/2025	22:43	LB136488
	Barium	14.6	+/-50	U	25.0	100	P	07/15/2025	22:43	LB136488
	Beryllium	0.56	+/-3	U	1.50	6.00	P	07/15/2025	22:43	LB136488
	Cadmium	0.50	+/-3	U	1.50	6.00	P	07/15/2025	22:43	LB136488
	Calcium	234	+/-1000	U	500	2000	P	07/15/2025	22:43	LB136488
	Chromium	2.12	+/-5	U	5.00	10.0	P	07/15/2025	22:43	LB136488
	Cobalt	2.26	+/-15	U	7.50	30.0	P	07/15/2025	22:43	LB136488
	Copper	4.60	+/-10	U	16.0	20.0	P	07/15/2025	22:43	LB136488
	Iron	23.4	+/-50	U	80.0	100	P	07/15/2025	22:43	LB136488
	Lead	2.30	+/-6	U	9.60	12.0	P	07/15/2025	22:43	LB136488
	Magnesium	244	+/-1000	U	500	2000	P	07/15/2025	22:43	LB136488
	Manganese	5.94	+/-10	U	5.00	20.0	P	07/15/2025	22:43	LB136488
	Nickel	3.06	+/-20	U	10.0	40.0	P	07/15/2025	22:43	LB136488
	Potassium	918	+/-1000	U	1600	2000	P	07/15/2025	22:43	LB136488
	Selenium	9.64	+/-10	U	16.0	20.0	P	07/15/2025	22:43	LB136488
	Silver	1.62	+/-5	U	5.00	10.0	P	07/15/2025	22:43	LB136488
	Sodium	868	+/-1000	U	1000	2000	P	07/15/2025	22:43	LB136488
	Thallium	4.38	+/-20	U	20.0	40.0	P	07/15/2025	22:43	LB136488
	Vanadium	6.26	+/-20	U	20.0	40.0	P	07/15/2025	22:43	LB136488
	Zinc	3.50	+/-20	U	15.0	40.0	P	07/15/2025	22:43	LB136488