

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group

SDG No.: Q2558

Contract: NOBI03

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168806BS							
Aluminum	mg/Kg	100	100		100	74 - 119	P
Antimony	mg/Kg	40.0	39.1		98	79 - 114	P
Arsenic	mg/Kg	40.0	35.6		89	82 - 111	P
Barium	mg/Kg	10.0	10.0		100	83 - 113	P
Beryllium	mg/Kg	10.0	10.6		106	83 - 113	P
Cadmium	mg/Kg	10.0	9.50		95	82 - 113	P
Calcium	mg/Kg	50.0	50.4	J	101	81 - 116	P
Chromium	mg/Kg	20.0	20.2		101	85 - 113	P
Cobalt	mg/Kg	10.0	9.66		97	85 - 112	P
Copper	mg/Kg	15.0	16.4		109	81 - 117	P
Iron	mg/Kg	150	170		113	81 - 118	P
Lead	mg/Kg	50.0	47.3		95	81 - 112	P
Magnesium	mg/Kg	100	105		105	78 - 115	P
Manganese	mg/Kg	10.0	10.2		102	84 - 114	P
Nickel	mg/Kg	25.0	24.3		97	83 - 113	P
Potassium	mg/Kg	500	462		92	81 - 116	P
Selenium	mg/Kg	100	97.7		98	78 - 111	P
Silver	mg/Kg	3.8	3.51		92	82 - 112	P
Sodium	mg/Kg	150	153		102	83 - 118	P
Thallium	mg/Kg	100	101		101	83 - 111	P
Vanadium	mg/Kg	15.0	14.7		98	82 - 114	P
Zinc	mg/Kg	10.0	10.3		102	82 - 113	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group

SDG No.: Q2558

Contract: NOBI03

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168821BS Mercury	mg/Kg	0.26	0.24		91	80 - 124	CV