

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136476

Review By	MOHAN	Review On	7/17/2025 1:46:12 PM
Supervise By	jaswal	Supervise On	7/17/2025 10:58:46 PM

STD. NAME	STD REF.#
ICAL Standard	MP86372,MP86373,MP86374,MP86375,MP86376,MP86377
ICV Standard	MP86378
CCV Standard	MP86380
ICSA Standard	
CRI Standard	MP86382
LCS Standard	
Chk Standard	MP86379,MP86381,MP86383,MP86385

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	07/15/25 11:46		MOHAN	OK
2	S0.2	S0.2	CAL2	07/15/25 11:49		MOHAN	OK
3	S2.5	S2.5	CAL3	07/15/25 11:51		MOHAN	OK
4	S5	S5	CAL4	07/15/25 11:53		MOHAN	OK
5	S7.5	S7.5	CAL5	07/15/25 11:56		MOHAN	OK
6	S10	S10	CAL6	07/15/25 12:03		MOHAN	OK
7	ICV27	ICV27	ICV	07/15/25 12:06		MOHAN	OK
8	ICB27	ICB27	ICB	07/15/25 12:08		MOHAN	OK
9	CCV01	CCV01	CCV	07/15/25 12:11		MOHAN	OK
10	CCB01	CCB01	CCB	07/15/25 12:13		MOHAN	OK
11	CRA	CRA	CRDL	07/15/25 12:21		MOHAN	OK
12	HighStd	HighStd	HIGH STD	07/15/25 12:23		MOHAN	OK
13	ChkStd	ChkStd	SAM	07/15/25 12:25		MOHAN	OK
14	PB168849BL	PB168849BL	MB	07/15/25 12:32		MOHAN	OK
15	PB168849BS	PB168849BS	LCS	07/15/25 12:35		MOHAN	OK
16	Q2561-07	AUD-25-0117	SAM	07/15/25 12:37		MOHAN	OK
17	Q2565-06	MOO-25-0194-0195	SAM	07/15/25 12:39		MOHAN	OK
18	Q2565-07	MOO-25-0191	SAM	07/15/25 12:42		MOHAN	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136476

Review By	MOHAN	Review On	7/17/2025 1:46:12 PM
Supervise By	jaswal	Supervise On	7/17/2025 10:58:46 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86372,MP86373,MP86374,MP86375,MP86376,MP86377		
ICV Standard	MP86378		
CCV Standard	MP86380		
ICSA Standard			
CRI Standard	MP86382		
LCS Standard			
Chk Standard	MP86379,MP86381,MP86383,MP86385		

19	Q2565-08	MOO-25-0196	SAM	07/15/25 12:44		MOHAN	OK
20	Q2565-09	MOO-25-0180	SAM	07/15/25 12:46		MOHAN	OK
21	CCV02	CCV02	CCV	07/15/25 12:48		MOHAN	OK
22	CCB02	CCB02	CCB	07/15/25 12:51		MOHAN	OK
23	Q2571-04	TP-18	SAM	07/15/25 12:53		MOHAN	OK
24	Q2571-08	TP-17	SAM	07/15/25 12:55		MOHAN	OK
25	Q2578-03	WC-A5-01-C	SAM	07/15/25 12:57		MOHAN	OK
26	Q2578-07	WC-A2-09-C	SAM	07/15/25 13:00		MOHAN	OK
27	Q2578-11	WC-A2-10-C	SAM	07/15/25 13:02		MOHAN	OK
28	Q2578-15	WC-A2-11-C	SAM	07/15/25 13:04		MOHAN	OK
29	Q2578-19	WC-A2-12-C	SAM	07/15/25 13:07		MOHAN	OK
30	Q2579-03	WC-A2-13-C	SAM	07/15/25 13:09		MOHAN	OK
31	Q2579-07	WC-A2-14-C	SAM	07/15/25 13:11		MOHAN	OK
32	Q2586-04	TP-16	SAM	07/15/25 13:13		MOHAN	OK
33	CCV03	CCV03	CCV	07/15/25 13:16		MOHAN	OK
34	CCB03	CCB03	CCB	07/15/25 13:18		MOHAN	OK
35	Q2586-04DUP	TP-16DUP	DUP	07/15/25 13:20		MOHAN	OK
36	Q2586-04MS	TP-16MS	MS	07/15/25 13:22		MOHAN	OK
37	Q2586-04MSD	TP-16MSD	MSD	07/15/25 13:25		MOHAN	OK
38	PB168860BL	PB168860BL	MB	07/15/25 13:27		MOHAN	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB136476

Review By	MOHAN	Review On	7/17/2025 1:46:12 PM
Supervise By	jaswal	Supervise On	7/17/2025 10:58:46 PM

STD. NAME	STD REF.#
ICAL Standard	MP86372,MP86373,MP86374,MP86375,MP86376,MP86377
ICV Standard	MP86378
CCV Standard	MP86380
ICSA Standard	
CRI Standard	MP86382
LCS Standard	
Chk Standard	MP86379,MP86381,MP86383,MP86385

39	PB168860BS	PB168860BS	LCS	07/15/25 13:32		MOHAN	OK
40	Q2555-02	OU4-TS-29-070925	SAM	07/15/25 13:34		MOHAN	OK
41	Q2555-02DUP	OU4-TS-29-070925D	DUP	07/15/25 13:39		MOHAN	OK
42	Q2555-02MS	OU4-TS-29-070925M	MS	07/15/25 13:47		MOHAN	OK
43	Q2555-02MSD	OU4-TS-29-070925M	MSD	07/15/25 13:49		MOHAN	OK
44	Q2555-04	OU4-TS-30-070925	SAM	07/15/25 13:50		MOHAN	OK
45	CCV04	CCV04	CCV	07/15/25 13:53		MOHAN	OK
46	CCB04	CCB04	CCB	07/15/25 13:58		MOHAN	OK
47	Q2558-02	OU4-TS-Denali-07092	SAM	07/15/25 14:00		MOHAN	OK
48	Q2558-04	OU4-TS-Grillo-OG-07	SAM	07/15/25 14:03		MOHAN	OK
49	PB168803TB	PB168803TB	MB	07/15/25 14:09		MOHAN	OK
50	PB168860TB	PB168860TB	MB	07/15/25 14:11		MOHAN	OK
51	Q2586-04L	TP-16L	SD	07/15/25 14:13		MOHAN	OK
52	Q2586-04A	TP-16A	PS	07/15/25 14:16		MOHAN	OK
53	Q2555-02L	OU4-TS-29-070925L	SD	07/15/25 14:18		MOHAN	OK
54	Q2555-02A	OU4-TS-29-070925A	PS	07/15/25 14:20		MOHAN	OK
55	CCV05	CCV05	CCV	07/15/25 14:22		MOHAN	OK
56	CCB05	CCB05	CCB	07/15/25 14:27		MOHAN	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB136512

Review By	Janvi	Review On	7/17/2025 1:50:29 PM
Supervise By	jaswal	Supervise On	7/18/2025 9:13:11 AM

STD. NAME	STD REF.#
ICAL Standard	MP86136,MP861437,MP86149,MP86151,MP86153,MP86155,MP86157,MP86160,MP86161
ICV Standard	MP86162
CCV Standard	MP86163
ICSA Standard	MP86164,MP86165
CRI Standard	MP86160
LCS Standard	
Chk Standard	MP86167,MP86168

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE	TUNE	07/16/25 09:47		Jaswal	OK
2	S0	S0	CAL1	07/16/25 10:55		Jaswal	OK
3	S2	S2	CAL3	07/16/25 10:58		Jaswal	OK
4	S3	S3	CAL4	07/16/25 11:05		Jaswal	OK
5	S4	S4	CAL5	07/16/25 11:08		Jaswal	OK
6	S5	S5	CAL6	07/16/25 11:11		Jaswal	OK
7	S6	S6	CAL7	07/16/25 11:14		Jaswal	OK
8	S7	S7	CAL8	07/16/25 11:17		Jaswal	OK
9	S8	S8	CAL9	07/16/25 11:20		Jaswal	OK
10	ICV01	ICV01	ICV	07/16/25 11:30		Jaswal	OK
11	LLICV01	LLICV01	LLICV	07/16/25 11:40		Jaswal	OK
12	ICB01	ICB01	ICB	07/16/25 11:47		Jaswal	OK
13	ICSA01	ICSA01	ICSA	07/16/25 12:21		Jaswal	OK
14	ICSAB01	ICSAB01	ICSAB	07/16/25 12:28		Jaswal	OK
15	CCV01	CCV01	CCV	07/16/25 12:51		Jaswal	OK
16	CCB01	CCB01	CCB	07/16/25 12:59		Jaswal	OK
17	CRI	CRI	CRDL	07/16/25 13:02		Jaswal	OK
18	Q2535-02DL	70125-LIQDL	SAM	07/16/25 13:06	Fe,Mn High	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB136512

Review By	Janvi	Review On	7/17/2025 1:50:29 PM
Supervise By	jaswal	Supervise On	7/18/2025 9:13:11 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86136,MP861437,MP86149,MP86151,MP86153,MP86155,MP86157,MP86160,MP86161		
ICV Standard	MP86162		
CCV Standard	MP86163		
ICSA Standard	MP86164,MP86165		
CRI Standard	MP86160		
LCS Standard			
Chk Standard	MP86167,MP86168		

19	Q2551-01	293	SAM	07/16/25 13:09	K,Na High	Jaswal	Dilution
20	Q2550-01DL	3326DL	SAM	07/16/25 13:12		Jaswal	OK
21	Q2550-01DUPDL	3326DUPDL	DUP	07/16/25 13:15		Jaswal	OK
22	Q2550-01LDL	3326LDL	SD	07/16/25 13:18		Jaswal	OK
23	Q2550-01MSDL	3326MSDL	MS	07/16/25 13:22		Jaswal	OK
24	Q2550-01MSDDL	3326MSDDL	MSD	07/16/25 13:25		Jaswal	OK
25	Q2550-01ADL	3326ADL	PS	07/16/25 13:28		Jaswal	OK
26	Q2535-02DL2	70125-LIQDL2	SAM	07/16/25 13:37	25X For Fe,Mn	Jaswal	Confirms
27	CCV02	CCV02	CCV	07/16/25 13:41		Jaswal	OK
28	CCB02	CCB02	CCB	07/16/25 13:47		Jaswal	OK
29	PB168792BL	PB168792BL	MB	07/16/25 13:50		Jaswal	OK
30	PB168792BS	PB168792BS	LCS	07/16/25 13:53		Jaswal	OK
31	Q2551-01DL	293DL	SAM	07/16/25 13:56	5X For K,Na	Jaswal	Confirms
32	PB168807BL	PB168807BL	MB	07/16/25 14:03		Jaswal	OK
33	PB168807BS	PB168807BS	LCS	07/16/25 14:07		Jaswal	OK
34	Q2562-01	17 BANK	SAM	07/16/25 14:10		Jaswal	OK
35	Q2562-02	SUPPLY	SAM	07/16/25 14:13	INT_STD 45Sc(2) Fail	Jaswal	Dilution
36	Q2562-02DUP	SUPPLYDUP	DUP	07/16/25 14:16	INT_STD 45Sc(2) Fail	Jaswal	Dilution
37	Q2562-02L	SUPPLYL	SD	07/16/25 14:19	INT_STD 45Sc(2) Fail	Jaswal	Dilution
38	Q2562-02MS	SUPPLYMS	MS	07/16/25 14:22	INT_STD 45Sc(2) Fail	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB136512

Review By	Janvi	Review On	7/17/2025 1:50:29 PM
Supervise By	jaswal	Supervise On	7/18/2025 9:13:11 AM
STD. NAME	STD REF.#		
ICAL Standard	MP86136,MP861437,MP86149,MP86151,MP86153,MP86155,MP86157,MP86160,MP86161		
ICV Standard	MP86162		
CCV Standard	MP86163		
ICSA Standard	MP86164,MP86165		
CRI Standard	MP86160		
LCS Standard			
Chk Standard	MP86167,MP86168		

39	CCV03	CCV03	CCV	07/16/25 14:40		Jaswal	OK
40	CCB03	CCB03	CCB	07/16/25 14:50		Jaswal	OK
41	Q2562-02MSRE	SUPPLYMS	MS	07/16/25 14:54	Not Use	Jaswal	Not Ok
42	Q2562-02MSD	SUPPLYMSD	MSD	07/16/25 15:00	INT_STD 45Sc(2) Fail	Jaswal	Dilution
43	Q2562-02A	SUPPLYA	PS	07/16/25 15:03		Jaswal	OK
44	Q2562-03	MW25	SAM	07/16/25 15:06	Na High	Jaswal	Dilution
45	Q2562-04	MW1	SAM	07/16/25 15:09		Jaswal	OK
46	Q2562-02DL	SUPPLYDL	SAM	07/16/25 15:16	5X For INT_STD 45Sc(2)	Jaswal	Confirms
47	Q2562-02DUPDL	SUPPLYDUPDL	DUP	07/16/25 15:19	5X For INT_STD 45Sc(2)	Jaswal	Confirms
48	Q2562-02LDL	SUPPLYLDL	SD	07/16/25 15:23	5X For INT_STD 45Sc(2)	Jaswal	Confirms
49	Q2562-02MSDL	SUPPLYMSDL	MS	07/16/25 15:26	5X For INT_STD 45Sc(2)	Jaswal	Confirms
50	Q2562-02MSDDL	SUPPLYMSDDL	MSD	07/16/25 15:29	5X For INT_STD 45Sc(2)	Jaswal	Confirms
51	CCV04	CCV04	CCV	07/16/25 15:45		Jaswal	OK
52	CCB04	CCB04	CCB	07/16/25 15:54		Jaswal	OK
53	Q2562-02ADL	SUPPLYADL	PS	07/16/25 15:58	Not Required	Jaswal	Not Ok
54	Q2562-03DL	MW25DL	SAM	07/16/25 16:01	10X For Na	Jaswal	Confirms
55	PB168856BL	PB168856BL	MB	07/16/25 16:14		Jaswal	OK
56	PB168856BS	PB168856BS	LCS	07/16/25 16:18		Jaswal	OK
57	Q2555-02LDL	OU4-TS-29-070925LDL	SD	07/16/25 16:27	INT_STD 89Y(1&2) Fail	Jaswal	Dilution
58	Q2555-02MSDL	OU4-TS-29-070925MSDL	MS	07/16/25 16:30	INT_STD 89Y(1&2) Fail	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB136512

Review By	Janvi	Review On	7/17/2025 1:50:29 PM
Supervise By	jaswal	Supervise On	7/18/2025 9:13:11 AM

STD. NAME	STD REF.#
ICAL Standard	MP86136,MP861437,MP86149,MP86151,MP86153,MP86155,MP86157,MP86160,MP86161
ICV Standard	MP86162
CCV Standard	MP86163
ICSA Standard	MP86164,MP86165
CRI Standard	MP86160
LCS Standard	
Chk Standard	MP86167,MP86168

59	Q2555-02MSDDL	OU4-TS-29-070925M	MSD	07/16/25 16:33	INT_STD 89Y(1&2) Fail	Jaswal	Dilution
60	Q2555-02ADL	OU4-TS-29-070925A	PS	07/16/25 16:36		Jaswal	OK
61	Q2555-02DUPDL	OU4-TS-29-070925D	DUP	07/16/25 16:42	INT_STD 89Y(1&2) Fail	Jaswal	Dilution
62	Q2555-02DL	OU4-TS-29-070925D	SAM	07/16/25 16:46	INT_STD 89Y(1&2) Fail	Jaswal	Dilution
63	CCV05	CCV05	CCV	07/16/25 17:02		Jaswal	OK
64	CCB05	CCB05	CCB	07/16/25 17:08		Jaswal	OK
65	PB168856TB	PB168856TB	MB	07/16/25 17:16		Jaswal	OK
66	Q2555-04DL	OU4-TS-30-070925D	SAM	07/16/25 17:19	INT_STD 89Y(1&2) Fail	Jaswal	Dilution
67	Q2558-02DL	OU4-TS-Denali-07092	SAM	07/16/25 17:22	INT_STD 89Y(1) Fail	Jaswal	Dilution
68	Q2558-04DL	OU4-TS-Grillo-OG-07	SAM	07/16/25 17:25		Jaswal	OK
69	Q2555-04DL2	OU4-TS-30-070925D	SAM	07/16/25 17:34	Not Required	Jaswal	Not Ok
70	CCV06	CCV06	CCV	07/16/25 17:57		Jaswal	OK
71	CCB06	CCB06	CCB	07/16/25 18:05		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB136576

Review By	Janvi	Review On	7/24/2025 2:58:34 PM
Supervise By	jaswal	Supervise On	7/24/2025 3:00:02 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86441,MP86469,MP86468,MP86451,MP86450,MP86449,MP86448,MP86447,MP86446		
ICV Standard	MP86500,MP86468		
CCV Standard	MP86470		
ICSA Standard	MP86496,MP86497		
CRI Standard	MP86468		
LCS Standard			
Chk Standard	MP86472,MP86473		

Sr#	SampleID	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE	TUNE	07/22/25 14:09		Jaswal	OK
2	S0	S0	CAL1	07/22/25 14:42		Jaswal	OK
3	S2	S2	CAL3	07/22/25 14:48		Jaswal	OK
4	S3	S3	CAL4	07/22/25 14:51		Jaswal	OK
5	S4	S4	CAL5	07/22/25 14:55		Jaswal	OK
6	S5	S5	CAL6	07/22/25 14:58		Jaswal	OK
7	S6	S6	CAL7	07/22/25 15:00		Jaswal	OK
8	S7	S7	CAL8	07/22/25 15:03		Jaswal	OK
9	S8	S8	CAL9	07/22/25 15:06		Jaswal	OK
10	ICV01	ICV01	ICV	07/22/25 16:39		Jaswal	OK
11	LLICV01	LLICV01	LLICV	07/22/25 16:42		Jaswal	OK
12	ICB01	ICB01	ICB	07/22/25 16:45		Jaswal	OK
13	ICSA01	ICSA01	ICSA	07/22/25 16:48		Jaswal	OK
14	ICSAB01	ICSAB01	ICSAB	07/22/25 16:55		Jaswal	OK
15	CCV01	CCV01	CCV	07/22/25 16:58		Jaswal	OK
16	CCB01	CCB01	CCB	07/22/25 17:09		Jaswal	OK
17	CRI	CRI	CRDL	07/22/25 17:12		Jaswal	OK
18	PB168937BL	PB168937BL	MB	07/22/25 17:15		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB136576

Review By	Janvi	Review On	7/24/2025 2:58:34 PM
Supervise By	jaswal	Supervise On	7/24/2025 3:00:02 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86441,MP86469,MP86468,MP86451,MP86450,MP86449,MP86448,MP86447,MP86446		
ICV Standard	MP86500,MP86468		
CCV Standard	MP86470		
ICSA Standard	MP86496,MP86497		
CRI Standard	MP86468		
LCS Standard			
Chk Standard	MP86472,MP86473		

19	PB168937BS	PB168937BS	LCS	07/22/25 17:19		Jaswal	OK
20	Q2638-02DL	OU4-TS-31-071725DI	SAM	07/22/25 17:21	IntStd89Y(1),209Bi(1) fail	Jaswal	Dilution
21	Q2638-04DL	OU4-TS-32-071725DI	SAM	07/22/25 17:25	IntStd89Y(1)(2)209Bi(1) fail	Jaswal	Dilution
22	Q2638-06DL	OU4-TS-33-071725DI	SAM	07/22/25 17:28	IntStd89Y(1)209Bi(1) fail	Jaswal	Dilution
23	Q2638-08DL	OU4-TS-34-071725DI	SAM	07/22/25 17:31	IntStd89Y(1)209Bi(1) fail	Jaswal	Dilution
24	Q2638-10DL	OU4-TS-35-071725DI	SAM	07/22/25 17:34	IntStd89Y(1)209Bi(1) fail	Jaswal	Dilution
25	Q2638-12DL	OU4-TS-36-071725DI	SAM	07/22/25 17:37	IntStd89Y(1)209Bi(1) fail	Jaswal	Dilution
26	Q2638-14DL	OU4-TS-37-071725DI	SAM	07/22/25 17:40	IntStd89Y(1)209Bi(1) fail	Jaswal	Dilution
27	CCV02	CCV02	CCV	07/22/25 17:48		Jaswal	OK
28	CCB02	CCB02	CCB	07/22/25 18:04		Jaswal	OK
29	Q2639-02DL	OU4-TS-38-071725DI	SAM	07/22/25 18:07	IntStd89Y(1&2),209Bi(1) fail	Jaswal	Dilution
30	Q2639-04DL	OU4-TS-39-071725DI	SAM	07/22/25 18:11	IntStd89Y(1),209Bi(1) fail	Jaswal	Dilution
31	Q2639-06DL	OU4-TS-40-071725DI	SAM	07/22/25 18:14	IntStd89Y(1&2),209Bi(1) fail	Jaswal	Dilution
32	Q2639-08DL	OU4-TS-41-071725DI	SAM	07/22/25 18:17	IntStd89Y(1&2),209Bi(1) fail	Jaswal	Dilution
33	Q2639-10DL	OU4-TS-42-071725DI	SAM	07/22/25 18:20	IntStd89Y(1),209Bi(1) fail	Jaswal	Dilution
34	Q2639-12DL	OU4-TS-43-071725DI	SAM	07/22/25 18:23	IntStd89Y(1),209Bi(1) fail	Jaswal	Dilution
35	Q2639-14DL	OU4-TS-44-071725DI	SAM	07/22/25 18:26	IntStd89Y(1&2),209Bi(1) fail	Jaswal	Dilution
36	Q2639-14DUPDL	OU4-TS-44-071725DI	DUP	07/22/25 18:29	IntStd89Y(1&2),209Bi(1) fail	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB136576

Review By	Janvi	Review On	7/24/2025 2:58:34 PM
Supervise By	jaswal	Supervise On	7/24/2025 3:00:02 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86441,MP86469,MP86468,MP86451,MP86450,MP86449,MP86448,MP86447,MP86446		
ICV Standard	MP86500,MP86468		
CCV Standard	MP86470		
ICSA Standard	MP86496,MP86497		
CRI Standard	MP86468		
LCS Standard			
Chk Standard	MP86472,MP86473		

37	Q2639-14LDL	OU4-TS-44-071725LDL	SD	07/22/25 18:33	IntStd89Y(1&2),209Bi(1) fail	Jaswal	Dilution
38	Q2639-14MSDL	OU4-TS-44-071725MSDL	MS	07/22/25 18:36	IntStd89Y(1&2),209Bi(1) fail	Jaswal	Dilution
39	CCV03	CCV03	CCV	07/22/25 19:00		Jaswal	OK
40	CCB03	CCB03	CCB	07/22/25 19:13		Jaswal	OK
41	Q2639-14MSDDL	OU4-TS-44-071725MSDDL	MSD	07/22/25 19:18	IntStd89Y(1&2),209Bi(1) fail	Jaswal	Dilution
42	Q2639-14ADL	OU4-TS-44-071725ADL	PS	07/22/25 19:21	IntStd89Y(1&2),209Bi(1) fail	Jaswal	Dilution
43	PB168937TB	PB168937TB	MB	07/22/25 19:24		Jaswal	OK
44	Q2555-02DL2	OU4-TS-29-070925DL2	SAM	07/22/25 19:27	Not Use	Jaswal	Not Ok
45	Q2555-02DUPDL2	OU4-TS-29-070925DUPDL2	DUP	07/22/25 19:31	Not Use	Jaswal	Not Ok
46	Q2555-02LDL2	OU4-TS-29-070925LDL2	SD	07/22/25 19:34	Not Use	Jaswal	Not Ok
47	Q2555-02MSDL2	OU4-TS-29-070925MSDL2	MS	07/22/25 19:37	Not Use	Jaswal	Not Ok
48	Q2555-02MSDDL2	OU4-TS-29-070925MSDDL2	MSD	07/22/25 19:40	Not Use	Jaswal	Not Ok
49	Q2555-02ADL2	OU4-TS-29-070925ADL2	PS	07/22/25 19:43	Not Use	Jaswal	Not Ok
50	Q2555-04DL2	OU4-TS-30-070925DL2	SAM	07/22/25 19:46	Not Use	Jaswal	Not Ok
51	CCV04	CCV04	CCV	07/22/25 19:53		Jaswal	OK
52	CCB04	CCB04	CCB	07/22/25 20:03		Jaswal	OK
53	Q2558-02DL2	OU4-TS-Denali-070925DL2	SAM	07/22/25 20:06	10X For INT_STD 89Y(1)	Jaswal	Confirms
54	Q2555-02DL2RE	OU4-TS-29-070925DL2RE	SAM	07/22/25 20:10	Not Use	Jaswal	Not Ok
55	Q2555-02DUPDL2RE	OU4-TS-29-070925DUPDL2RE	DUP	07/22/25 20:13	Not Use	Jaswal	Not Ok

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB136576

Review By	Janvi	Review On	7/24/2025 2:58:34 PM
Supervise By	jaswal	Supervise On	7/24/2025 3:00:02 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86441,MP86469,MP86468,MP86451,MP86450,MP86449,MP86448,MP86447,MP86446		
ICV Standard	MP86500,MP86468		
CCV Standard	MP86470		
ICSA Standard	MP86496,MP86497		
CRI Standard	MP86468		
LCS Standard			
Chk Standard	MP86472,MP86473		

56	Q2555-02DL3RE	OU4-TS-29-070925DI	SAM	07/22/25 20:25	Not Use	Jaswal	Not Ok
57	Q2555-02DUPDL3RE	OU4-TS-29-070925DI	DUP	07/22/25 20:28	Not Use	Jaswal	Not Ok
58	Q2555-02LDL3RE	OU4-TS-29-070925DI	SD	07/22/25 20:32	Not Use	Jaswal	Not Ok
59	Q2555-02MSDL3RE	OU4-TS-29-070925DI	MS	07/22/25 20:35	Not Use	Jaswal	Not Ok
60	Q2555-02MSDDL3RE	OU4-TS-29-070925DI	MSD	07/22/25 20:38	Not Use	Jaswal	Not Ok
61	Q2555-02ADL3RE	OU4-TS-29-070925DI	PS	07/22/25 20:41	Not Use	Jaswal	Not Ok
62	CCV05	CCV05	CCV	07/22/25 21:02		Jaswal	OK
63	CCB05	CCB05	CCB	07/22/25 21:09		Jaswal	OK
64	Q2555-02DL3	OU4-TS-29-070925DI	SAM	07/22/25 21:12	Not Use	Jaswal	Not Ok
65	Q2555-02DUPDL3	OU4-TS-29-070925DI	DUP	07/22/25 21:16	Not Use	Jaswal	Not Ok
66	Q2555-02LDL3	OU4-TS-29-070925DI	SD	07/22/25 21:19	Not Use	Jaswal	Not Ok
67	Q2555-02MSDL3	OU4-TS-29-070925DI	MS	07/22/25 21:22	Not Use	Jaswal	Not Ok
68	Q2555-02MSDDL3	OU4-TS-29-070925DI	MSD	07/22/25 21:25	Not Use	Jaswal	Not Ok
69	Q2555-02ADL3	OU4-TS-29-070925DI	PS	07/22/25 21:28	Not Use	Jaswal	Not Ok
70	Q2639-02DL2	OU4-TS-38-071725DI	SAM	07/22/25 21:32	Not Use	Jaswal	Not Ok
71	Q2639-04DL2	OU4-TS-39-071725DI	SAM	07/22/25 21:35	Not Use	Jaswal	Not Ok
72	Q2639-06DL2	OU4-TS-40-071725DI	SAM	07/22/25 21:38	Not Use	Jaswal	Not Ok
73	Q2639-08DL2	OU4-TS-41-071725DI	SAM	07/22/25 21:41	Not Use	Jaswal	Not Ok
74	CCV06	CCV06	CCV	07/22/25 22:04		Jaswal	OK
75	CCB06	CCB06	CCB	07/22/25 22:11		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB136576

Review By	Janvi	Review On	7/24/2025 2:58:34 PM
Supervise By	jaswal	Supervise On	7/24/2025 3:00:02 PM

STD. NAME	STD REF.#
ICAL Standard	MP86441,MP86469,MP86468,MP86451,MP86450,MP86449,MP86448,MP86447,MP86446
ICV Standard	MP86500,MP86468
CCV Standard	MP86470
ICSA Standard	MP86496,MP86497
CRI Standard	MP86468
LCS Standard	
Chk Standard	MP86472,MP86473

76	Q2638-02DL2	OU4-TS-31-071725DI	SAM	07/22/25 22:15	25X For INT_STD 89Y(1),209Bi(1)	Jaswal	Confirms
77	Q2638-04DL2	OU4-TS-32-071725DI	SAM	07/22/25 22:18	25X For 89Y(1&2),209Bi(1) , Still 89Y(1&2)Fail	Jaswal	Dilution
78	Q2638-06DL2	OU4-TS-33-071725DI	SAM	07/22/25 22:21	25X For 89Y(1),209Bi(1) , Still 89Y(1)Fail	Jaswal	Dilution
79	Q2638-14DL2	OU4-TS-37-071725DI	SAM	07/22/25 22:24	25X For 89Y(1),209Bi(1) , Still 89Y(1)Fail	Jaswal	Dilution
80	CCV07	CCV07	CCV	07/22/25 22:28		Jaswal	OK
81	CCB07	CCB07	CCB	07/22/25 22:30		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB136601

Review By	jaswal	Review On	7/24/2025 2:38:05 PM
Supervise By	Janvi	Supervise On	7/24/2025 2:40:05 PM
STD. NAME	STD REF.#		
ICAL Standard	MP86441,MP86469,MP86468,MP86451,MP86450,MP86449,MP86448,MP86447,MP86446		
ICV Standard	MP86500,MP86468		
CCV Standard	MP86470		
ICSA Standard	MP86496,MP86497		
CRI Standard	MP86468		
LCS Standard			
Chk Standard	MP86472,MP86473		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE	TUNE	07/24/25 09:46		Jaswal	OK
2	S0	S0	CAL1	07/24/25 10:42		Jaswal	OK
3	S2	S2	CAL3	07/24/25 10:49		Jaswal	OK
4	S3	S3	CAL4	07/24/25 10:52		Jaswal	OK
5	S4	S4	CAL5	07/24/25 10:55		Jaswal	OK
6	S5	S5	CAL6	07/24/25 10:58		Jaswal	OK
7	S6	S6	CAL7	07/24/25 11:01		Jaswal	OK
8	S7	S7	CAL8	07/24/25 11:04		Jaswal	OK
9	S8	S8	CAL9	07/24/25 11:07		Jaswal	OK
10	ICV01	ICV01	ICV	07/24/25 11:25		Jaswal	OK
11	LLICV01	LLICV01	LLICV	07/24/25 11:34		Jaswal	OK
12	ICB01	ICB01	ICB	07/24/25 11:38		Jaswal	OK
13	ICSA01	ICSA01	ICSA	07/24/25 11:41		Jaswal	OK
14	ICSAB01	ICSAB01	ICSAB	07/24/25 11:49		Jaswal	OK
15	CCV01	CCV01	CCV	07/24/25 11:52		Jaswal	OK
16	CCB01	CCB01	CCB	07/24/25 11:57		Jaswal	OK
17	CRI	CRI	CRDL	07/24/25 12:01		Jaswal	OK
18	Q2555-02DL2	OU4-TS-29-070925DI	SAM	07/24/25 12:04	100X for intstd89Y(1)(2)	Jaswal	Confirms

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB136601

Review By	jaswal	Review On	7/24/2025 2:38:05 PM
Supervise By	Janvi	Supervise On	7/24/2025 2:40:05 PM

STD. NAME	STD REF.#
ICAL Standard	MP86441,MP86469,MP86468,MP86451,MP86450,MP86449,MP86448,MP86447,MP86446
ICV Standard	MP86500,MP86468
CCV Standard	MP86470
ICSA Standard	MP86496,MP86497
CRI Standard	MP86468
LCS Standard	
Chk Standard	MP86472,MP86473

19	Q2555-02DUPDL2	OU4-TS-29-070925D	DUP	07/24/25 12:07	100X for intstd89Y(1)(2)	Jaswal	Confirms
20	Q2555-02LDL2	OU4-TS-29-070925LD	SD	07/24/25 12:11	100X for intstd89Y(1)(2)	Jaswal	Confirms
21	Q2555-02MSDL2	OU4-TS-29-070925M	MS	07/24/25 12:14	100X for intstd89Y(1)(2)	Jaswal	Confirms
22	Q2555-02MSDDL2	OU4-TS-29-070925M	MSD	07/24/25 12:17	100X for intstd89Y(1)(2)	Jaswal	Confirms
23	Q2555-02ADL2	OU4-TS-29-070925A	PS	07/24/25 12:20		Jaswal	Not Ok
24	Q2555-04DL2	OU4-TS-30-070925D	SAM	07/24/25 12:24	50X for intstd89Y(1)(2)	Jaswal	Confirms
25	Q2638-02DL	OU4-TS-31-071725D	SAM	07/24/25 12:31		Jaswal	Confirms
26	Q2638-03DL	OU4-TS-32-071725D	SAM	07/24/25 12:34		Jaswal	Confirms
27	CCV02	CCV02	CCV	07/24/25 12:39		Jaswal	OK
28	CCB02	CCB02	CCB	07/24/25 12:45		Jaswal	OK