

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Nobis Group **Level:** LOW **SDG No.:** Q2558
Contract: NOBI03 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q2555-02 **Client ID:** OU4-TS-29-070925DUP
Percent Solids for Sample: NA **Duplicate ID** Q2555-02DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Antimony	ug/L	20	10.0	UD	10.0	UD			P
Arsenic	ug/L	20	127	D	148	D	15		P
Barium	ug/L	20	120	D	118	D	2		P
Beryllium	ug/L	20	16.6	D	16.9	D	2		P
Cadmium	ug/L	20	9.60	D	9.50	D	1		P
Chromium	ug/L	20	53.3	D	50.5	D	5		P
Copper	ug/L	20	442	D	420	D	5		P
Lead	ug/L	20	81.4	D	78.0	D	4		P
Nickel	ug/L	20	217	D	207	D	5		P
Mercury	ug/L	20	0.20	U	0.20	U			CV
Selenium	ug/L	20	500	UD	500	UD			P
Silver	ug/L	20	0.45	JD	0.45	JD	0		P
Thallium	ug/L	20	0.40	JD	0.35	JD	13		P
Vanadium	ug/L	20	129	D	124	D	4		P
Zinc	ug/L	20	1670	D	1610	D	4		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Nobis Group **Level:** LOW **SDG No.:** Q2558
Contract: NOBI03 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q2555-02MS **Client ID:** OU4-TS-29-070925MSD
Percent Solids for Sample: NA **Duplicate ID** Q2555-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Antimony	ug/L	20	447	D	445	D	0		P
Arsenic	ug/L	20	562	D	626	D	11		P
Barium	ug/L	20	2040	D	2040	D	0		P
Beryllium	ug/L	20	551	D	553	D	0		P
Cadmium	ug/L	20	467	D	471	D	1		P
Chromium	ug/L	20	489	D	497	D	2		P
Copper	ug/L	20	5540	D	5690	D	3		P
Lead	ug/L	20	2600	D	2610	D	0		P
Nickel	ug/L	20	689	D	699	D	1		P
Mercury	ug/L	20	3.95		4.10		4		CV
Selenium	ug/L	20	418	JD	465	JD	11		P
Silver	ug/L	20	89.5	D	89.0	D	1		P
Thallium	ug/L	20	409	D	415	D	1		P
Vanadium	ug/L	20	567	D	576	D	2		P
Zinc	ug/L	20	5980	D	6080	D	2		P