

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group

SDG No.: Q2558

Contract: NOBI03

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168856BS							
Antimony	ug/L	500	523		105	85 - 117	P
Arsenic	ug/L	500	503		101	84 - 116	P
Barium	ug/L	2500	2570		103	86 - 114	P
Beryllium	ug/L	500	523		105	83 - 121	P
Cadmium	ug/L	500	465		93	87 - 115	P
Chromium	ug/L	500	513		103	85 - 116	P
Copper	ug/L	5000	5050		101	85 - 118	P
Lead	ug/L	2500	2510		100	88 - 115	P
Nickel	ug/L	500	512		102	85 - 117	P
Selenium	ug/L	500	494		99	80 - 120	P
Silver	ug/L	500	521		104	85 - 116	P
Thallium	ug/L	500	501		100	82 - 116	P
Vanadium	ug/L	500	506		101	86 - 115	P
Zinc	ug/L	5000	5040		101	83 - 119	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group

SDG No.: Q2558

Contract: NOBI03

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168860BS Mercury	ug/L	4.0	4.15		104	82 - 119	CV